

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2894
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL04 **Date Analyzed:** 08/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113117.D **Time Analyzed:** 09:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.747	4.648	4.848	552.430	500.000	10.5
Aroclor-1016-2	4.765	4.666	4.866	552.440	500.000	10.5
Aroclor-1016-3	4.822	4.724	4.924	557.870	500.000	11.6
Aroclor-1016-4	4.942	4.843	5.043	510.680	500.000	2.1
Aroclor-1016-5	5.198	5.099	5.299	539.090	500.000	7.8
Aroclor-1260-1	6.235	6.136	6.336	537.520	500.000	7.5
Aroclor-1260-2	6.424	6.326	6.526	537.480	500.000	7.5
Aroclor-1260-3	6.790	6.691	6.891	542.040	500.000	8.4
Aroclor-1260-4	7.049	6.951	7.151	538.930	500.000	7.8
Aroclor-1260-5	7.291	7.194	7.394	542.160	500.000	8.4
Decachlorobiphenyl	8.680	8.582	8.782	51.280	50.000	2.6
Tetrachloro-m-xylene	3.662	3.563	3.763	54.750	50.000	9.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2894
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL04 **Date Analyzed:** 08/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113117.D **Time Analyzed:** 09:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.728	4.629	4.829	510.640	500.000	2.1
Aroclor-1016-2	4.746	4.647	4.847	506.950	500.000	1.4
Aroclor-1016-3	4.920	4.822	5.022	494.770	500.000	-1.0
Aroclor-1016-4	4.963	4.864	5.064	483.490	500.000	-3.3
Aroclor-1016-5	5.175	5.076	5.276	515.140	500.000	3.0
Aroclor-1260-1	6.202	6.104	6.304	487.740	500.000	-2.5
Aroclor-1260-2	6.391	6.292	6.492	485.740	500.000	-2.9
Aroclor-1260-3	6.542	6.443	6.643	461.000	500.000	-7.8
Aroclor-1260-4	7.011	6.912	7.112	477.600	500.000	-4.5
Aroclor-1260-5	7.253	7.155	7.355	470.110	500.000	-6.0
Decachlorobiphenyl	8.626	8.528	8.728	51.090	50.000	2.2
Tetrachloro-m-xylene	3.654	3.555	3.755	51.540	50.000	3.1