

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2894
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL05 **Date Analyzed:** 08/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113121.D **Time Analyzed:** 11:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.747	4.648	4.848	539.080	500.000	7.8
Aroclor-1016-2	4.765	4.666	4.866	539.140	500.000	7.8
Aroclor-1016-3	4.822	4.724	4.924	544.630	500.000	8.9
Aroclor-1016-4	4.941	4.843	5.043	497.690	500.000	-0.5
Aroclor-1016-5	5.198	5.099	5.299	530.520	500.000	6.1
Aroclor-1260-1	6.234	6.136	6.336	522.410	500.000	4.5
Aroclor-1260-2	6.424	6.326	6.526	524.960	500.000	5.0
Aroclor-1260-3	6.789	6.691	6.891	530.520	500.000	6.1
Aroclor-1260-4	7.049	6.951	7.151	528.340	500.000	5.7
Aroclor-1260-5	7.293	7.194	7.394	531.270	500.000	6.3
Decachlorobiphenyl	8.681	8.582	8.782	50.780	50.000	1.6
Tetrachloro-m-xylene	3.662	3.563	3.763	53.320	50.000	6.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2894
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL05 **Date Analyzed:** 08/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113121.D **Time Analyzed:** 11:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.727	4.629	4.829	495.730	500.000	-0.9
Aroclor-1016-2	4.745	4.647	4.847	503.750	500.000	0.8
Aroclor-1016-3	4.920	4.822	5.022	486.710	500.000	-2.7
Aroclor-1016-4	4.963	4.864	5.064	477.370	500.000	-4.5
Aroclor-1016-5	5.175	5.076	5.276	505.560	500.000	1.1
Aroclor-1260-1	6.202	6.104	6.304	483.430	500.000	-3.3
Aroclor-1260-2	6.391	6.292	6.492	484.400	500.000	-3.1
Aroclor-1260-3	6.542	6.443	6.643	451.820	500.000	-9.6
Aroclor-1260-4	7.012	6.912	7.112	474.760	500.000	-5.0
Aroclor-1260-5	7.254	7.155	7.355	466.870	500.000	-6.6
Decachlorobiphenyl	8.626	8.528	8.728	50.470	50.000	0.9
Tetrachloro-m-xylene	3.654	3.555	3.755	50.740	50.000	1.5