

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO081825

Review By	yogesh	Review On	8/18/2025 10:23:59 AM
Supervise By	mohammad	Supervise On	8/20/2025 3:04:50 AM
SubDirectory	PO081825	HP Acquire Method	HP Processing Method PO081825

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP2416,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO113001.D	18 Aug 2025 09:16		YP/AJ	Ok
2	I.BLK	I.BLK	PO113002.D	18 Aug 2025 09:34		YP/AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PO113003.D	18 Aug 2025 09:53		YP/AJ	Ok
4	AR1660ICC750	AR1660ICC750	PO113004.D	18 Aug 2025 10:11		YP/AJ	Ok
5	AR1660ICC500	AR1660ICC500	PO113005.D	18 Aug 2025 10:29		YP/AJ	Ok
6	AR1660ICC250	AR1660ICC250	PO113006.D	18 Aug 2025 10:48		YP/AJ	Ok
7	AR1660ICC050	AR1660ICC050	PO113007.D	18 Aug 2025 11:06		YP/AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PO113008.D	18 Aug 2025 11:25		YP/AJ	Ok
9	AR1232ICC500	AR1232ICC500	PO113009.D	18 Aug 2025 11:43		YP/AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PO113010.D	18 Aug 2025 12:01		YP/AJ	Ok
11	AR1242ICC750	AR1242ICC750	PO113011.D	18 Aug 2025 12:20		YP/AJ	Ok
12	AR1242ICC500	AR1242ICC500	PO113012.D	18 Aug 2025 12:38		YP/AJ	Ok
13	AR1242ICC250	AR1242ICC250	PO113013.D	18 Aug 2025 12:57		YP/AJ	Ok
14	AR1242ICC050	AR1242ICC050	PO113014.D	18 Aug 2025 13:15		YP/AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PO113015.D	18 Aug 2025 13:34		YP/AJ	Ok
16	AR1248ICC750	AR1248ICC750	PO113016.D	18 Aug 2025 13:52		YP/AJ	Ok
17	AR1248ICC500	AR1248ICC500	PO113017.D	18 Aug 2025 14:09		YP/AJ	Ok
18	AR1248ICC250	AR1248ICC250	PO113018.D	18 Aug 2025 14:28		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO081825

Review By	yogesh	Review On	8/18/2025 10:23:59 AM
Supervise By	mohammad	Supervise On	8/20/2025 3:04:50 AM
SubDirectory	PO081825	HP Acquire Method	HP Processing Method PO081825
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24815,PP2416,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854 PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		

19	AR1248ICC050	AR1248ICC050	PO113019.D	18 Aug 2025 14:46		YP/AJ	Ok
20	AR1254ICC1000	AR1254ICC1000	PO113020.D	18 Aug 2025 15:05		YP/AJ	Ok
21	AR1254ICC750	AR1254ICC750	PO113021.D	18 Aug 2025 15:23		YP/AJ	Ok
22	AR1254ICC500	AR1254ICC500	PO113022.D	18 Aug 2025 15:41		YP/AJ	Ok
23	AR1254ICC250	AR1254ICC250	PO113023.D	18 Aug 2025 16:00		YP/AJ	Ok
24	AR1254ICC050	AR1254ICC050	PO113024.D	18 Aug 2025 16:18		YP/AJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PO113025.D	18 Aug 2025 16:54		YP/AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PO113026.D	18 Aug 2025 17:12		YP/AJ	Ok
27	AR1268ICC750	AR1268ICC750	PO113027.D	18 Aug 2025 17:30		YP/AJ	Ok
28	AR1268ICC500	AR1268ICC500	PO113028.D	18 Aug 2025 17:48		YP/AJ	Ok
29	AR1268ICC250	AR1268ICC250	PO113029.D	18 Aug 2025 18:06		YP/AJ	Ok
30	AR1268ICC050	AR1268ICC050	PO113030.D	18 Aug 2025 18:24		YP/AJ	Ok,M
31	PO081825ICV500	ICVPO081825	PO113031.D	18 Aug 2025 18:43		YP/AJ	Ok
32	AR1242ICV500	ICVPO081825AR1242	PO113032.D	18 Aug 2025 19:37		YP/AJ	Ok
33	AR1248ICV500	ICVPO081825AR1248	PO113033.D	18 Aug 2025 19:54		YP/AJ	Ok
34	AR1254ICV500	ICVPO081825AR1254	PO113034.D	18 Aug 2025 20:31		YP/AJ	Ok
35	AR1268ICV500	ICVPO081825AR1268	PO113035.D	18 Aug 2025 21:07		YP/AJ	Ok,M
36	DDT ANALOG	DDT ANALOG	PO113036.D	18 Aug 2025 21:44		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO082025

Review By	yogesh	Review On	8/20/2025 11:41:48 AM
Supervise By	mohammad	Supervise On	8/22/2025 2:04:33 AM
SubDirectory	PO082025	HP Acquire Method	HP Processing Method PO081825

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP2416,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO113073.D	20 Aug 2025 09:16		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO113074.D	20 Aug 2025 09:33		YP/AJ	Ok,M
3	AR1242CCC500	AR1242CCC500	PO113075.D	20 Aug 2025 09:52		YP/AJ	Ok
4	AR1248CCC500	AR1248CCC500	PO113076.D	20 Aug 2025 10:10		YP/AJ	Ok
5	AR1254CCC500	AR1254CCC500	PO113077.D	20 Aug 2025 10:28		YP/AJ	Ok
6	I.BLK	I.BLK	PO113078.D	20 Aug 2025 10:47		YP/AJ	Ok
7	DDT ANALOG	DDT ANALOG	PO113079.D	20 Aug 2025 11:05		YP/AJ	Ok
8	PB169322BL	PB169322BL	PO113080.D	20 Aug 2025 12:01	Run with cleanup sample	YP/AJ	Not Ok
9	PB169322BS	PB169322BS	PO113081.D	20 Aug 2025 12:19		YP/AJ	Ok
10	Q2894-01	1A1B1C	PO113082.D	20 Aug 2025 12:36		YP/AJ	Ok,M
11	Q2894-02	2A2B2C	PO113083.D	20 Aug 2025 12:55		YP/AJ	Ok,M
12	Q2894-03	3A3B3C	PO113084.D	20 Aug 2025 13:13		YP/AJ	Ok
13	Q2894-04	4A4B4C	PO113085.D	20 Aug 2025 13:32	need cleanup	YP/AJ	Not Ok
14	Q2894-05	5A5B5C	PO113086.D	20 Aug 2025 13:50	DCB high in both column	YP/AJ	Dilution
15	Q2901-01	HVAC DUCT SEALANT	PO113087.D	20 Aug 2025 14:09	AR1254 Hit, Need Dilution,	YP/AJ	Dilution
16	Q2901-02	EXTERIOR WINDOW C	PO113088.D	20 Aug 2025 14:27	AR1260 Hit	YP/AJ	Dilution
17	AR1660CCC500	AR1660CCC500	PO113089.D	20 Aug 2025 15:39		YP/AJ	Ok,M
18	AR1242CCC500	AR1242CCC500	PO113090.D	20 Aug 2025 15:57		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO082025

Review By	yogesh	Review On	8/20/2025 11:41:48 AM
Supervise By	mohammad	Supervise On	8/22/2025 2:04:33 AM
SubDirectory	PO082025	HP Acquire Method	HP Processing Method PO081825
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP2416,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PO113091.D	20 Aug 2025 16:14		YP/AJ	Ok
20	AR1254CCC500	AR1254CCC500	PO113092.D	20 Aug 2025 16:32		YP/AJ	Ok
21	I.BLK	I.BLK	PO113093.D	20 Aug 2025 16:50		YP/AJ	Ok
22	PB169322BL	PB169322BL	PO113094.D	20 Aug 2025 17:09		YP/AJ	Ok
23	PB169326BL	PB169326BL	PO113095.D	20 Aug 2025 17:27		YP/AJ	Ok
24	PB169326BS	PB169326BS	PO113096.D	20 Aug 2025 17:45		YP/AJ	Ok
25	Q2894-04	4A4B4C	PO113097.D	20 Aug 2025 18:03		YP/AJ	Ok
26	Q2907-01	JEI0237-1-1	PO113098.D	20 Aug 2025 18:21		YP/AJ	Ok,M
27	Q2907-02	JEI0237-1-2	PO113099.D	20 Aug 2025 18:40	AR1260 Hit	YP/AJ	Ok,M
28	Q2907-03	BC271336-1-1	PO113100.D	20 Aug 2025 18:57	DCB high in 2nd column	YP/AJ	Ok,M
29	Q2907-04	BC271336-1-2	PO113101.D	20 Aug 2025 19:16		YP/AJ	Ok,M
30	Q2908-01	GHA571Y-1-1	PO113102.D	20 Aug 2025 19:33		YP/AJ	Ok,M
31	Q2908-02	GHA571Y-1-2	PO113103.D	20 Aug 2025 19:51		YP/AJ	Ok,M
32	AR1660CCC500	AR1660CCC500	PO113104.D	20 Aug 2025 21:05		YP/AJ	Ok
33	AR1242CCC500	AR1242CCC500	PO113105.D	20 Aug 2025 21:42		YP/AJ	Ok
34	AR1248CCC500	AR1248CCC500	PO113106.D	20 Aug 2025 22:00		YP/AJ	Ok
35	AR1254CCC500	AR1254CCC500	PO113107.D	20 Aug 2025 22:19		YP/AJ	Ok
36	I.BLK	I.BLK	PO113108.D	20 Aug 2025 22:37		YP/AJ	Ok
37	Q2901-01DL	HVAC DUCT SEALANT	PO113109.D	20 Aug 2025 22:55	AR1254 Hit	YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO082025

Review By	yogesh	Review On	8/20/2025 11:41:48 AM
Supervise By	mohammad	Supervise On	8/22/2025 2:04:33 AM
SubDirectory	PO082025	HP Acquire Method	HP Processing Method PO081825
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP2416,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	Q2901-02DL	EXTERIOR WINDOW	PO113110.D	20 Aug 2025 23:14	AR1260 Hit	YP/AJ	Ok,M
39	AR1660CCC500	AR1660CCC500	PO113111.D	21 Aug 2025 00:27		YP/AJ	Ok,M
40	AR1242CCC500	AR1242CCC500	PO113112.D	21 Aug 2025 01:04		YP/AJ	Ok
41	AR1248CCC500	AR1248CCC500	PO113113.D	21 Aug 2025 01:22		YP/AJ	Ok
42	AR1254CCC500	AR1254CCC500	PO113114.D	21 Aug 2025 01:41		YP/AJ	Ok
43	I.BLK	I.BLK	PO113115.D	21 Aug 2025 01:59		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO082125

Review By	yogesh	Review On	8/21/2025 10:42:36 AM
Supervise By		Supervise On	
SubDirectory	PO082125	HP Acquire Method	HP Processing Method PO081825

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852
Internal Standard/PEM	
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO113116.D	21 Aug 2025 08:47		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO113117.D	21 Aug 2025 09:05		YP/AJ	Ok
3	I.BLK	I.BLK	PO113118.D	21 Aug 2025 09:46		YP/AJ	Ok
4	DDT ANALOG	DDT ANALOG	PO113119.D	21 Aug 2025 10:04		YP/AJ	Ok
5	Q2894-05	5A5B5C	PO113120.D	21 Aug 2025 10:22		YP/AJ	Ok
6	AR1660CCC500	AR1660CCC500	PO113121.D	21 Aug 2025 11:17		YP/AJ	Ok
7	AR1242CCC500	AR1242CCC500	PO113122.D	21 Aug 2025 11:39		YP/AJ	Ok
8	AR1248CCC500	AR1248CCC500	PO113123.D	21 Aug 2025 11:57		YP/AJ	Ok
9	AR1254CCC500	AR1254CCC500	PO113124.D	21 Aug 2025 12:16		YP/AJ	Ok
10	I.BLK	I.BLK	PO113125.D	21 Aug 2025 12:34		YP/AJ	Ok
11	Q2896-08	8	PO113126.D	21 Aug 2025 12:53		YP/AJ	Ok
12	Q2896-09	9	PO113127.D	21 Aug 2025 13:11		YP/AJ	Ok,NS
13	Q2896-10	10	PO113128.D	21 Aug 2025 13:30		YP/AJ	Ok
14	Q2914-01	K262-9A	PO113129.D	21 Aug 2025 13:48	DCB high in both column	YP/AJ	ReRun
15	Q2914-04	K262-10A	PO113130.D	21 Aug 2025 14:06	DCB high in 2nd column	YP/AJ	Ok
16	Q2914-07	K262-11A	PO113131.D	21 Aug 2025 14:25	DCB high in both column	YP/AJ	ReRun
17	Q2914-10	K262-12A	PO113132.D	21 Aug 2025 14:43	DCB high in both column	YP/AJ	ReRun
18	Q2914-13	K262-13A	PO113133.D	21 Aug 2025 15:02	DCB high in both column	YP/AJ	ReRun

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO082125

Review By	yogesh	Review On	8/21/2025 10:42:36 AM
Supervise By		Supervise On	
SubDirectory	PO082125	HP Acquire Method	HP Processing Method PO081825
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,PP24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q2914-13MS	K262-13AMS	PO113134.D	21 Aug 2025 15:20	No clear peak	YP/AJ	Not Ok
20	Q2914-13MSD	K262-13AMSD	PO113135.D	21 Aug 2025 15:37	No clear peak	YP/AJ	Not Ok
21	AR1660CCC500	AR1660CCC500	PO113136.D	21 Aug 2025 17:09		YP/AJ	Ok
22	AR1242CCC500	AR1242CCC500	PO113137.D	21 Aug 2025 17:45		YP/AJ	Ok
23	AR1248CCC500	AR1248CCC500	PO113138.D	21 Aug 2025 18:04		YP/AJ	Ok
24	AR1254CCC500	AR1254CCC500	PO113139.D	21 Aug 2025 18:22		YP/AJ	Ok
25	I.BLK	I.BLK	PO113140.D	21 Aug 2025 18:41		YP/AJ	Ok
26	PB169337BL	PB169337BL	PO113141.D	21 Aug 2025 18:59		YP/AJ	Ok
27	PB169337BS	PB169337BS	PO113142.D	21 Aug 2025 19:17		YP/AJ	Ok
28	Q2914-16	K262-14A	PO113143.D	21 Aug 2025 19:36	DCB high in both column	YP/AJ	ReRun
29	Q2914-19	K262-15A	PO113144.D	21 Aug 2025 19:54	DCB high in both column	YP/AJ	ReRun
30	Q2922-02	K262-16A	PO113145.D	21 Aug 2025 20:13	AR1254 Hit	YP/AJ	Ok,NS
31	Q2922-05	K262-17A	PO113146.D	21 Aug 2025 20:31	AR1254 Hit	YP/AJ	Ok,NS
32	Q2922-08	K262-18A	PO113147.D	21 Aug 2025 20:49	AR1254 Hit	YP/AJ	Ok,NS
33	AR1660CCC500	AR1660CCC500	PO113148.D	21 Aug 2025 22:03		YP/AJ	Ok,NS
34	AR1242CCC500	AR1242CCC500	PO113149.D	21 Aug 2025 22:58		YP/AJ	Ok
35	AR1248CCC500	AR1248CCC500	PO113150.D	21 Aug 2025 23:17		YP/AJ	Ok
36	AR1254CCC500	AR1254CCC500	PO113151.D	21 Aug 2025 23:35		YP/AJ	Ok
37	I.BLK	I.BLK	PO113152.D	21 Aug 2025 23:53		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO082125

Review By	yogesh	Review On	8/21/2025 10:42:36 AM
Supervise By		Supervise On	
SubDirectory	PO082125	HP Acquire Method	HP Processing Method PO081825
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24823,PP24824,PP24825,PP24826,PP24827,PP24828,PP24829,P P24830,PP24831,PP24832,PP24833,PP24834,PP24835,PP24836,PP24837,PP24838,PP24839,PP24840,PP24841,PP24842,PP24843,PP24844,PP 24845,PP24846,PP24847,PP24848,PP24849,PP24850,PP24851,PP24852,PP24853,PP24854		
CCC	PP24824,PP24828,PP24832,PP24836,PP24840,PP24844,PP24852		
Internal Standard/PEM			
ICV/I.BLK	PP24815,PP24816,PP24817,PP24818,PP24819,PP24820,PP24821,PP24822,PP24855,PP24856,PP24857,PP24858,PP24859,PP24860,PP24861,PP24862		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	Q2914-01RE	K262-9ARE	PO113153.D	22 Aug 2025 00:12	DCB high in both column	YP/AJ	Confirms
39	Q2914-04RE	K262-10ARE	PO113154.D	22 Aug 2025 00:30	not needed	YP/AJ	Not Ok
40	Q2914-07RE	K262-11ARE	PO113155.D	22 Aug 2025 00:49	DCB high in both column	YP/AJ	Confirms
41	Q2914-10RE	K262-12ARE	PO113156.D	22 Aug 2025 01:07	DCB high in both column	YP/AJ	Confirms
42	AR1660CCC500	AR1660CCC500	PO113157.D	22 Aug 2025 02:21		YP/AJ	Ok,NS
43	AR1242CCC500	AR1242CCC500	PO113158.D	22 Aug 2025 03:16		YP/AJ	Ok,NS
44	AR1248CCC500	AR1248CCC500	PO113159.D	22 Aug 2025 03:33		YP/AJ	Ok
45	AR1254CCC500	AR1254CCC500	PO113160.D	22 Aug 2025 03:52		YP/AJ	Ok,NS
46	I.BLK	I.BLK	PO113161.D	22 Aug 2025 04:10		YP/AJ	Ok

M : Manual Integration