



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2894

Analytical Method: 8082A

Client: ATC Group Services LLC

Datafile : PO113081.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB169322BS (Column 1)	AR1016	166.6	178	ug/kg	107				71	120	
	AR1260	166.6	171	ug/kg	103				65	130	
PB169322BS (Column 2)	AR1016	166.6	166	ug/kg	100				71	120	
	AR1260	166.6	162	ug/kg	97				65	130	