

284 Sheffield Street, Mountainside, New Jersey 07092, Phone: 908 789 8900,

Fax: 908 789 8922

Report of Analysis

Client: PSEG Date Collected:

Project: Harding (Morristown) Distribution Sub MA00006789 Date Received:

Client Sample ID: PB169296BL SDG No.: Q2897
Lab Sample ID: PB169296BL Matrix: Solid

Analytical Method: NJEPH % Solid: 100

Sample Wt/Vol: 30.03 Units: g Final Vol: 2000 uL

Soil Aliquot Vol: uL Test: EPH_NF

Prep Method:

Prep Date : Date Analyzed : Prep Batch ID

08/19/25 08:05 08/19/25 13:00 PB169296

CAS Number Parameter		Conc.	Qualifier	Dilution	MDL	LOQ / CRQL	Units(Dry Weight)	
TARGETS								
Aliphatic C28-C40	Aliphatic C28-C40	1.18	U	1	1.18	2.00	mg/kg FE	055379.D
Aliphatic C9-C28	Aliphatic C9-C28	0.91	U	1	0.91	3.99	mg/kg FE	055379.D
Total AliphaticEPl	H Total AliphaticEPH	2.09	U		2.09	5.99	mg/kg	
Total EPH	Total EPH	2.09	U		2.09	5.99	mg/kg	

^{*} As samples are not fractionated, all aliphatic and aromatic carbon compounds in the C9-C40 carbon range are calculated against the aliphatic calibration curve, and reported as Aliphatic EPH. Therefore, the aliphatic C9-C40 concentration for the sample is reported as the Total EPH.

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

Datafile

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution



284 Sheffield Street, Mountainside, New Jersey 07092, Phone: 908 789 8900,

Fax: 908 789 8922

Report of Analysis

Client: PSEG Date Collected:

Project: Harding (Morristown) Distribution Sub MA00006789 Date Received:

Client Sample ID: PB169296BL SDG No.: Q2897
Lab Sample ID: PB169296BL Matrix: Solid

Analytical Method: NJEPH % Solid: 100

Sample Wt/Vol: 30.03 Units: g Final Vol: 2000 uL

Soil Aliquot Vol: uL Test: EPH_NF

Prep Method:

Prep Date : Date Analyzed : Prep Batch ID

CAS Number Parameter		Conc.	Qualifier	Dilution	MDL	LOQ / CRQL	Units(Dry Weight)	
TARGETS								
Aliphatic C28-C40	Aliphatic C28-C40	1.18	U	1	1.18	2.00	mg/kg FE	055379.D
Aliphatic C9-C28	Aliphatic C9-C28	0.91	U	1	0.91	3.99	mg/kg FE	055379.D
Total AliphaticEPl	H Total AliphaticEPH	2.09	U		2.09	5.99	mg/kg	
Total EPH	Total EPH	2.09	U		2.09	5.99	mg/kg	

^{*} As samples are not fractionated, all aliphatic and aromatic carbon compounds in the C9-C40 carbon range are calculated against the aliphatic calibration curve, and reported as Aliphatic EPH. Therefore, the aliphatic C9-C40 concentration for the sample is reported as the Total EPH.

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

Datafile

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution



uL



Report of Analysis

Client: **PSEG** Date Collected:

Project: Harding (Morristown) Distribution Sub MA00006789 Date Received:

Client Sample ID: PB169296BL SDG No.: Q2897 Lab Sample ID: PB169296BL Matrix: Solid

Analytical Method: % Solid: 100 NJEPH

Sample Wt/Vol: 30.03 Units: Final Vol: 2000 g Test: EPH_NF

Soil Aliquot Vol: uL

Prep Method:

File ID: Dilution: Prep Date: Prep Batch ID Date Analyzed: 1 FE055379.D 08/19/25 08/19/25 PB169296

CAS Number	Parameter		Conc. Q	ualifier	MDL	LOQ / CRQL	Units
TARGETS							
Aliphatic C9-C	C28	Aliphatic C9-C28	0.000	U	0.91	3.99	mg/kg
Aliphatic C28-	-C40	Aliphatic C28-C40	1.18	U	1.18	2.00	mg/kg
SURROGATES	8						
3383-33-2		1-chlorooctadecane (SURR)	48.1		40 - 140	96%	SPK: 50
84-15-1		ortho-Terphenyl (SURR)	46.1		40 - 140	92%	SPK: 50



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Quantitation Report For Aliphatic EPH Range.

Lab Sample ID: PB169296BL Acq On: 19 Aug 2025 13:00

Client Sample ID: PB169296BL Operator: YP\AJ

Data file: FE055379.D Misc:

Instrument: FID_E ALS Vial: 11

Dilution Factor: 1 Sample Multiplier: 1.00

R.T.		Response	Conc	highest_standard	Units
3.317	6.951	0	0	300	ug/ml
6.952	10.402	0	0	200	ug/ml
10.403	13.781	0	0	300	ug/ml
13.782	17.453	0	0	400	ug/ml
17.454	22.471	0	0	600	ug/ml
3.317	22.471	0	0		ug/ml
12.085	12.085	6577287	46.06		ug/ml
13.521	13.521	5175453	48.13		ug/ml
3.317	17.453	0	0	1200	ug/ml
	6.952 10.403 13.782 17.454 3.317 12.085 13.521	6.952 10.402 10.403 13.781 13.782 17.453 17.454 22.471 3.317 22.471 12.085 12.085 13.521 13.521	6.952 10.402 0 10.403 13.781 0 13.782 17.453 0 17.454 22.471 0 3.317 22.471 0 12.085 12.085 6577287 13.521 5175453	6.952 10.402 0 0 10.403 13.781 0 0 13.782 17.453 0 0 17.454 22.471 0 0 3.317 22.471 0 0 12.085 12.085 6577287 46.06 13.521 13.521 5175453 48.13	6.952 10.402 0 0 200 10.403 13.781 0 0 300 13.782 17.453 0 0 400 17.454 22.471 0 0 600 3.317 22.471 0 0 0 12.085 12.085 6577287 46.06 46.06 13.521 13.521 5175453 48.13