

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AEC002</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2900</u>
Instrument ID:	<u>BNA_F</u>	Calibration Date/Time:	<u>08/20/2025 22:05</u>
Lab File ID:	<u>BF143498.D</u>	Init. Calib. Date(s):	<u>08/20/2025 08/20/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>10:55 14:20</u>
GC Column:	<u>DB-UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.145	1.150		0.4	
Benzaldehyde	1.145	1.081		-5.6	
Phenol-d6	1.414	1.395		-1.3	
Phenol	1.629	1.649		1.2	20.0
bis(2-Chloroethyl)ether	1.217	1.211		-0.5	
2-Chlorophenol	1.265	1.270		0.4	
2-Methylphenol	1.015	1.021		0.6	
2,2-oxybis(1-Chloropropane)	1.947	1.949		0.1	
Acetophenone	0.432	0.424		-1.9	
3+4-Methylphenols	1.212	1.222		0.8	
n-Nitroso-di-n-propylamine	0.864	0.852	0.050	-1.4	
Nitrobenzene-d5	0.343	0.349		1.7	
Hexachloroethane	0.485	0.481		-0.8	
Nitrobenzene	0.353	0.362		2.5	
Isophorone	0.628	0.635		1.1	
2-Nitrophenol	0.173	0.181		4.6	20.0
2,4-Dimethylphenol	0.269	0.270		0.4	
bis(2-Chloroethoxy)methane	0.386	0.385		-0.3	
2,4-Dichlorophenol	0.288	0.290		0.7	20.0
Naphthalene	0.960	0.933		-2.8	
4-Chloroaniline	0.341	0.343		0.6	
Hexachlorobutadiene	0.193	0.192		-0.5	20.0
Caprolactam	0.069	0.078		13.0	
4-Chloro-3-methylphenol	0.288	0.293		1.7	20.0
2-Methylnaphthalene	0.641	0.630		-1.7	
Hexachlorocyclopentadiene	0.184	0.163	0.050	-11.4	
2,4,6-Trichlorophenol	0.352	0.356		1.1	20.0
2-Fluorobiphenyl	1.210	1.135		-6.2	
2,4,5-Trichlorophenol	0.385	0.377		-2.1	
1,1-Biphenyl	1.420	1.373		-3.3	
2-Chloronaphthalene	1.114	1.101		-1.2	
2-Nitroaniline	0.326	0.341		4.6	
Dimethylphthalate	1.205	1.228		1.9	
Acenaphthylene	1.595	1.574		-1.3	
2,6-Dinitrotoluene	0.250	0.262		4.8	
3-Nitroaniline	0.270	0.286		5.9	
Acenaphthene	1.083	1.061		-2.0	20.0
2,4-Dinitrophenol	0.115	0.099	0.050	-13.9	
4-Nitrophenol	0.161	0.174	0.050	8.1	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2900</u>
Instrument ID:	<u>BNA_F</u>	Calibration Date/Time:	<u>08/20/2025 22:05</u>
Lab File ID:	<u>BF143498.D</u>	Init. Calib. Date(s):	<u>08/20/2025 08/20/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>10:55 14:20</u>
GC Column:	<u>DB-UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.544	1.518		-1.7	
2,4-Dinitrotoluene	0.333	0.354		6.3	
Diethylphthalate	1.080	1.104		2.2	
4-Chlorophenyl-phenylether	0.594	0.578		-2.7	
Fluorene	1.188	1.156		-2.7	
4-Nitroaniline	0.248	0.268		8.1	
4,6-Dinitro-2-methylphenol	0.113	0.110		-2.7	
n-Nitrosodiphenylamine	0.644	0.627		-2.6	20.0
2,4,6-Tribromophenol	0.186	0.184		-1.1	
4-Bromophenyl-phenylether	0.239	0.237		-0.8	
Hexachlorobenzene	0.257	0.247		-3.9	
Atrazine	0.166	0.185		11.4	
Pentachlorophenol	0.118	0.119		0.8	20.0
Phenanthrene	1.071	1.039		-3.0	
Anthracene	1.085	1.059		-2.4	
Carbazole	0.889	0.887		-0.2	
Di-n-butylphthalate	0.829	0.911		9.9	
Fluoranthene	0.961	0.962		0.1	20.0
Pyrene	1.645	1.665		1.2	
Terphenyl-d14	1.146	1.151		0.4	
Butylbenzylphthalate	0.449	0.480		6.9	
3,3-Dichlorobenzidine	0.369	0.369		0.0	
Benzo (a) anthracene	1.288	1.254		-2.6	
Chrysene	1.157	1.176		1.6	
Bis(2-ethylhexyl)phthalate	0.685	0.667		-2.6	
Di-n-octyl phthalate	1.268	1.215		-4.2	20.0
Benzo (b) fluoranthene	1.112	1.107		-0.4	
Benzo (k) fluoranthene	1.064	1.049		-1.4	
Benzo (a) pyrene	1.032	1.033		0.1	20.0
Indeno (1,2,3-cd) pyrene	1.437	1.484		3.3	
Dibenzo (a,h) anthracene	1.151	1.194		3.7	
Benzo (g,h,i) perylene	1.145	1.178		2.9	
1,2,4,5-Tetrachlorobenzene	0.572	0.553		-3.3	
1,4-Dioxane	0.531	0.507		-4.5	20.0
2,3,4,6-Tetrachlorophenol	0.286	0.296		3.5	

All other compounds must meet a minimum RRF of 0.010.