

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF082025

Review By	Rahul	Review On	8/21/2025 8:38:24 AM
Supervise By	Jagrut	Supervise On	8/21/2025 11:02:57 AM
SubDirectory	BF082025	HP Acquire Method	BNA_F
		HP Processing Method	bf082025

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13171,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF143476.D	20 Aug 2025 10:26		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF143477.D	20 Aug 2025 10:55		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF143478.D	20 Aug 2025 11:25		RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF143479.D	20 Aug 2025 11:54		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF143480.D	20 Aug 2025 12:23	compounds #32,41,54 kept on LR	RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF143481.D	20 Aug 2025 12:53	Benzidine failed in the Calibration Method.	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BF143482.D	20 Aug 2025 13:22		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BF143483.D	20 Aug 2025 13:51		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF143484.D	20 Aug 2025 14:20	Compound #9 removed from 80PPM	RC/JU	Ok
10	SSTDICV040	ICVBF082025	BF143485.D	20 Aug 2025 15:16		RC/JU	Ok
11	PB169299BL	PB169299BL	BF143486.D	20 Aug 2025 15:46		RC/JU	Ok
12	PB169299BS	PB169299BS	BF143487.D	20 Aug 2025 16:15	Recovery is low for compound #3,3-Dichlorobenzidine	RC/JU	Not Ok
13	PB169313BS	PB169313BS	BF143488.D	20 Aug 2025 16:44		RC/JU	Ok,M
14	PB169313BSD	PB169313BSD	BF143489.D	20 Aug 2025 17:13		RC/JU	Ok,M
15	Q2892-12	FRONT-1-FLOOR-F	BF143490.D	20 Aug 2025 17:43		RC/JU	Ok
16	Q2892-10	3-FLOOR-SOUTH-E	BF143491.D	20 Aug 2025 18:12		RC/JU	Ok
17	Q2892-10MS	3-FLOOR-SOUTH-EMS	BF143492.D	20 Aug 2025 18:41		RC/JU	Ok,M

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF082025

Review By	Rahul	Review On	8/21/2025 8:38:24 AM		
Supervise By	Jagrut	Supervise On	8/21/2025 11:02:57 AM		
SubDirectory	BF082025	HP Acquire Method	BNA_F	HP Processing Method	bf082025
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864				
CCC	SP6860				
Internal Standard/PEM	S13171,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q2892-10MSD	3-FLOOR-SOUTH-EMS	BF143493.D	20 Aug 2025 19:10		RC/JU	Ok,M
19	Q2900-06	MW-17C-081825	BF143494.D	20 Aug 2025 19:39		RC/JU	Ok
20	Q2900-07	MW-11B-081825	BF143495.D	20 Aug 2025 20:08		RC/JU	Ok
21	Q2900-04	MW-17B-081825	BF143496.D	20 Aug 2025 20:37		RC/JU	Ok
22	DFTPP	DFTPP	BF143497.D	20 Aug 2025 21:36		RC/JU	Ok
23	SSTDCCC040	SSTDCCC040	BF143498.D	20 Aug 2025 22:05		RC/JU	Ok
24	PB169313BL	PB169313BL	BF143499.D	20 Aug 2025 22:34		RC/JU	Ok
25	PB169223BS	PB169223BS	BF143500.D	20 Aug 2025 23:03		RC/JU	Ok,M
26	PB169223BSD	PB169223BSD	BF143501.D	20 Aug 2025 23:32		RC/JU	Ok,M
27	Q2815-12	TW-22M-S	BF143502.D	21 Aug 2025 00:00		RC/JU	Ok
28	Q2900-01	MN-9C-081825	BF143503.D	21 Aug 2025 00:29		RC/JU	Ok
29	Q2819-20	82H-E	BF143504.D	21 Aug 2025 00:58		RC/JU	Ok
30	Q2819-14	17M-W	BF143505.D	21 Aug 2025 01:27		RC/JU	Ok
31	Q2819-15	17M-N	BF143506.D	21 Aug 2025 01:56		RC/JU	Ok
32	Q2819-18	38M-W	BF143507.D	21 Aug 2025 02:25		RC/JU	Ok
33	Q2819-19	38M-E	BF143508.D	21 Aug 2025 02:54		RC/JU	Ok
34	Q2892-06	BACK-1-FLOOR-CENT	BF143509.D	21 Aug 2025 03:23		RC/JU	Ok
35	Q2892-08	3-FLOOR-NORTH-D	BF143510.D	21 Aug 2025 03:52		RC/JU	Ok
36	Q2900-02	MN-12C-081825	BF143511.D	21 Aug 2025 04:21	Internal standard fail and need 20X then decide further dilution.	RC/JU	Dilution

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF082025

Review By	Rahul	Review On	8/21/2025 8:38:24 AM		
Supervise By	Jagrut	Supervise On	8/21/2025 11:02:57 AM		
SubDirectory	BF082025	HP Acquire Method	BNA_F	HP Processing Method	bf082025
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864				
CCC	SP6860				
Internal Standard/PEM	S13171,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

37	Q2900-03	DUP-01-081825	BF143512.D	21 Aug 2025 04:49	Internal standard fail and need 20X then decide further dilution.	RC/JU	Dilution
38	Q2900-05	MW-11C-081825	BF143513.D	21 Aug 2025 05:18	Internal standard fail and need 20X then decide further dilution.	RC/JU	Dilution
39	Q2815-11	TW-22M-W	BF143514.D	21 Aug 2025 05:47		RC/JU	Ok
40	Q2815-08	TW-88H-N	BF143515.D	21 Aug 2025 06:15		RC/JU	Ok
41	SSTDCCC040	SSTDCCC040EC	BF143516.D	21 Aug 2025 06:44		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF082125

Review By	Rahul	Review On	8/22/2025 8:18:48 AM
Supervise By	mohammad	Supervise On	8/22/2025 8:27:40 AM
SubDirectory	BF082125	HP Acquire Method	BNA_F
		HP Processing Method	bf082025

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864
CCC	SP6860
Internal Standard/PEM	S13171,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF143517.D	21 Aug 2025 08:20		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF143518.D	21 Aug 2025 09:21		RC/JU	Ok
3	PB169299BL	PB169299BL	BF143519.D	21 Aug 2025 09:50		RC/JU	Ok
4	Q2820-24	22M-S	BF143520.D	21 Aug 2025 10:20		RC/JU	Ok
5	Q2820-12	10PC-S	BF143521.D	21 Aug 2025 10:49		RC/JU	Ok
6	Q2900-02DL	MN-12C-081825DL	BF143522.D	21 Aug 2025 11:18	Need 5X further Dilution	RC/JU	Dilution
7	Q2900-03DL	DUP-01-081825DL	BF143523.D	21 Aug 2025 11:47	Need 5X further Dilution	RC/JU	Dilution
8	Q2900-05DL	MW-11C-081825DL	BF143524.D	21 Aug 2025 12:17	Need 2X further Dilution	RC/JU	Dilution
9	Q2819-07	11M-N	BF143525.D	21 Aug 2025 12:46		RC/JU	Ok,M
10	Q2819-08	11M-E	BF143526.D	21 Aug 2025 13:15		RC/JU	Ok,M
11	Q2900-02DL2	MN-12C-081825DL2	BF143527.D	21 Aug 2025 13:44		RC/JU	Ok
12	Q2900-03DL2	DUP-01-081825DL2	BF143528.D	21 Aug 2025 14:13		RC/JU	Ok
13	Q2900-05DL2	MW-11C-081825DL2	BF143529.D	21 Aug 2025 14:42		RC/JU	Ok
14	Q2819-10RE	84SB-SRE	BF143530.D	21 Aug 2025 15:11	Surrogate Fail	RC/JU	Confirms
15	Q2819-11	84SB-W	BF143531.D	21 Aug 2025 15:40		RC/JU	Ok,M
16	Q2819-06	11M-S	BF143532.D	21 Aug 2025 16:09		RC/JU	Ok
17	Q2819-01	22BP-N	BF143533.D	21 Aug 2025 16:38		RC/JU	Ok,M
18	Q2814-07	TW-82H-W	BF143534.D	21 Aug 2025 17:07		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF082125

Review By	Rahul	Review On	8/22/2025 8:18:48 AM		
Supervise By	mohammad	Supervise On	8/22/2025 8:27:40 AM		
SubDirectory	BF082125	HP Acquire Method	BNA_F	HP Processing Method	bf082025
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864				
CCC	SP6860				
Internal Standard/PEM	S13171,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q2815-15	TW-17M-E	BF143535.D	21 Aug 2025 17:36		RC/JU	Ok
20	Q2814-19	TW-518R-E	BF143536.D	21 Aug 2025 18:05		RC/JU	Ok
21	Q2819-13	17M-E	BF143537.D	21 Aug 2025 18:34		RC/JU	Ok
22	Q2819-04RE	22BP-SRE	BF143538.D	21 Aug 2025 19:03	Surrogate fail.	RC/JU	Confirms
23	SSTDCCC040	SSTDCCC040EC	BF143539.D	21 Aug 2025 19:32		RC/JU	Ok
24	DFTPP	DFTPP	BF143540.D	21 Aug 2025 20:02		RC/JU	Ok
25	SSTDCCC040	SSTDCCC040	BF143541.D	21 Aug 2025 21:00	CCC fail high for compound#9,69,70,74 & CCC fail low for compound # 78,79,87,91,92.	RC/JU	Ok
26	PB169330BL	PB169330BL	BF143542.D	21 Aug 2025 21:58	CCC fail low for compound # 78,79,87,91,92.	RC/JU	Not Ok
27	PB169330BS	PB169330BS	BF143543.D	21 Aug 2025 22:27	CCC fail low for compound # 78,79,87,91,92.	RC/JU	Not Ok
28	Q2903-01	MW-6A-081925	BF143544.D	21 Aug 2025 22:56	CCC fail low for compound # 78,79,87,91,92.	RC/JU	ReRun
29	Q2903-02	MW-1C-081925	BF143545.D	21 Aug 2025 23:25	CCC fail low for compound # 78,79,87,91,92.	RC/JU	ReRun
30	Q2903-03	MW-6B-081925	BF143546.D	21 Aug 2025 23:54	CCC fail low for compound # 78,79,87,91,92.	RC/JU	ReRun
31	Q2903-04	MW-1B-081925	BF143547.D	22 Aug 2025 00:23	CCC fail low for compound # 78,79,87,91,92.	RC/JU	ReRun
32	Q2903-05MS	MW-1B-081925MS	BF143548.D	22 Aug 2025 00:52	CCC fail low for compound # 78,79,87,91,92.	RC/JU	Not Ok
33	Q2903-06MSD	MW-1B-081925MSD	BF143549.D	22 Aug 2025 01:21	CCC fail low for compound # 78,79,87,91,92.	RC/JU	Not Ok
34	Q2903-07	MW-12B-081925	BF143550.D	22 Aug 2025 01:51	CCC fail low for compound # 78,79,87,91,92.	RC/JU	ReRun

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF082125

Review By	Rahul	Review On	8/22/2025 8:18:48 AM		
Supervise By	mohammad	Supervise On	8/22/2025 8:27:40 AM		
SubDirectory	BF082125	HP Acquire Method	BNA_F	HP Processing Method	bf082025
STD. NAME		STD REF.#			
Tune/Reschk		SP6757			
Initial Calibration Stds		SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864			
CCC		SP6860			
Internal Standard/PEM		S13171,10ul/1000ul sample			
ICV/I.BLK		SP6770			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

35	Q2903-08	MW-18B-081925	BF143551.D	22 Aug 2025 02:19	CCC fail low for compound # 78,79,87,91,92.	RC/JU	ReRun
36	Q2903-09	DUP-02-081925	BF143552.D	22 Aug 2025 02:49	CCC fail low for compound # 78,79,87,91,92.	RC/JU	ReRun
37	Q2903-10	MW-17A-081925	BF143553.D	22 Aug 2025 03:18	CCC fail low for compound # 78,79,87,91,92.	RC/JU	ReRun
38	Q2903-11	MW-18A-081925	BF143554.D	22 Aug 2025 03:46	CCC fail low for compound # 78,87,91,92.Need 2X Dilution	RC/JU	Dilution
39	Q2903-12	MW-16B-081925	BF143555.D	22 Aug 2025 04:16	CCC fail low for compound # 78,79,87,91,92.	RC/JU	ReRun
40	Q2915-01	BRIDGE-WEST	BF143556.D	22 Aug 2025 04:45	CCC fail low for compound # 78,79,87,91,92.	RC/JU	ReRun
41	Q2915-01MS	BRIDGE-WESTMS	BF143557.D	22 Aug 2025 05:14	CCC fail low for compound # 78,79,87,91,92.	RC/JU	Not Ok
42	Q2915-01MSD	BRIDGE-WESTMSD	BF143558.D	22 Aug 2025 05:43	CCC fail low for compound # 78,79,87,91,92.	RC/JU	Not Ok
43	Q2915-03	BRIDGE-EAST	BF143559.D	22 Aug 2025 06:11	CCC fail low for compound # 78,79,87,91,92.	RC/JU	ReRun
44	Q2892-04	1-FLOOR-DRILL-B-BAL	BF143560.D	22 Aug 2025 06:40	CCC fail low for compound # 78,79,87,91,92.	RC/JU	ReRun
45	Q2892-02	1-FLOOR-NORTH-EAST	BF143561.D	22 Aug 2025 07:09	CCC fail low for compound # 78,79,87,91,92.	RC/JU	Not Ok

M : Manual Integration