

LAB CHRONICLE

OrderID:	Q2900	OrderDate:	8/18/2025 4:05:40 PM
Client:	AECOM	Project:	National Grid Equity - Brooklyn NY
Contact:	Peter S	Location:	J11,VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2900-01	MN-9C-081825	Water			08/18/25			08/18/25
			SVOC-TCL BNA -20	8270E		08/19/25	08/21/25	
Q2900-02	MN-12C-081825	Water			08/18/25			08/18/25
			SVOC-TCL BNA -20	8270E		08/19/25	08/21/25	
Q2900-02DL	MN-12C-081825DL	Water			08/18/25			08/18/25
			SVOC-TCL BNA -20	8270E		08/19/25	08/21/25	
Q2900-02DL 2	MN-12C-081825DL2	Water			08/18/25			08/18/25
			SVOC-TCL BNA -20	8270E		08/19/25	08/21/25	
Q2900-03	DUP-01-081825	Water			08/18/25			08/18/25
			SVOC-TCL BNA -20	8270E		08/19/25	08/21/25	
Q2900-03DL	DUP-01-081825DL	Water			08/18/25			08/18/25
			SVOC-TCL BNA -20	8270E		08/19/25	08/21/25	
Q2900-03DL 2	DUP-01-081825DL2	Water			08/18/25			08/18/25
			SVOC-TCL BNA -20	8270E		08/19/25	08/21/25	
Q2900-04	MW-17B-081825	Water			08/18/25			08/18/25
			SVOC-TCL BNA -20	8270E		08/19/25	08/20/25	
Q2900-05	MW-11C-081825	Water			08/18/25			08/18/25
			SVOC-TCL BNA -20	8270E		08/19/25	08/21/25	
Q2900-05DL	MW-11C-081825DL	Water			08/18/25			08/18/25
			SVOC-TCL BNA -20	8270E		08/19/25	08/21/25	
Q2900-05DL 2	MW-11C-081825DL2	Water			08/18/25			08/18/25

LAB CHRONICLE

Q2900-06	MW-17C-081825	Water	SVOC-TCL BNA -20	8270E		08/19/25	08/21/25	
					08/18/25			08/18/25
Q2900-07	MW-11B-081825	Water	SVOC-TCL BNA -20	8270E		08/19/25	08/20/25	
					08/18/25			08/18/25
			SVOC-TCL BNA -20	8270E		08/19/25	08/20/25	

Hit Summary Sheet SW-846

SDG No.: Q2900
Client: AECOM

Sample ID	Client ID		Parameter	Concentration	C	MDL	RDL	Units
Client ID :	MN-9C-081825							
Q2900-01	MN-9C-081825	WATER	2-Pentanone, 4-hydroxy-4-methyl *	4.400	AB	0	0	ug/L
Q2900-01	MN-9C-081825	WATER	Butane, 2-methoxy-2-methyl- *	110.000	J	0	0	ug/L
Total Tics :				114.40				
Total Concentration:				114.40				
Client ID :	MN-12C-081825							
Q2900-02	MN-12C-081825	WATER	Naphthalene	3,800.000	E	0.55	5.5	ug/L
Q2900-02	MN-12C-081825	WATER	2-Methylnaphthalene	800.000	E	0.62	5.5	ug/L
Q2900-02	MN-12C-081825	WATER	1,1-Biphenyl	27.700		0.58	5.5	ug/L
Q2900-02	MN-12C-081825	WATER	Acenaphthylene	150.000	E	0.82	5.5	ug/L
Q2900-02	MN-12C-081825	WATER	Acenaphthene	30.200		0.6	5.5	ug/L
Q2900-02	MN-12C-081825	WATER	Dibenzofuran	3.800	J	0.67	5.5	ug/L
Q2900-02	MN-12C-081825	WATER	Fluorene	27.100		0.69	5.5	ug/L
Q2900-02	MN-12C-081825	WATER	Phenanthrene	11.400		0.55	5.5	ug/L
Q2900-02	MN-12C-081825	WATER	Carbazole	13.700		0.79	5.5	ug/L
Total Svoc :				4,863.90				
Q2900-02	MN-12C-081825	WATER	.alpha.-Methylstyrene *	7.800	J	0	0	ug/L
Q2900-02	MN-12C-081825	WATER	1,3,5-Cycloheptatriene *	180.000	J	0	0	ug/L
Q2900-02	MN-12C-081825	WATER	1,3-Cyclopentadiene, 5-(1-methyl) *	110.000	J	0	0	ug/L
Q2900-02	MN-12C-081825	WATER	1,3-Methanopentalene, 1,2,3,5-tet *	8.700	J	0	0	ug/L
Q2900-02	MN-12C-081825	WATER	Benzene, (2-methyl-1-propenyl)- *	7.700	J	0	0	ug/L
Q2900-02	MN-12C-081825	WATER	Benzene, 1,2,4-trimethyl- *	22.000	J	0	0	ug/L
Q2900-02	MN-12C-081825	WATER	Benzene, 1-ethenyl-2-methyl- *	68.000	J	0	0	ug/L
Q2900-02	MN-12C-081825	WATER	Benzene, 1-ethyl-2-methyl- *	40.800	J	0	0	ug/L
Q2900-02	MN-12C-081825	WATER	Benzene, 1-ethyl-3-methyl- *	7.300	J	0	0	ug/L
Q2900-02	MN-12C-081825	WATER	Benzene, 1-ethyl-4-methyl- *	21.600	J	0	0	ug/L
Q2900-02	MN-12C-081825	WATER	Bicyclo[4.2.0]octa-1,3,5-triene *	120.000	J	0	0	ug/L
Q2900-02	MN-12C-081825	WATER	Butane, 2-methoxy-2-methyl- *	19.600	J	0	0	ug/L
Q2900-02	MN-12C-081825	WATER	Naphthalene, 1,6-dimethyl- *	26.400	J	0	0	ug/L
Q2900-02	MN-12C-081825	WATER	Naphthalene, 2,3-dimethyl- *	13.500	J	0	0	ug/L
Q2900-02	MN-12C-081825	WATER	Naphthalene, 2,7-dimethyl- *	22.200	J	0	0	ug/L
Q2900-02	MN-12C-081825	WATER	Indane *	64.800	J	0	0	ug/L
Q2900-02	MN-12C-081825	WATER	Indene *	200.000	J	0	0	ug/L
Q2900-02	MN-12C-081825	WATER	unknown7.045 *	8.400	J	0	0	ug/L
Q2900-02	MN-12C-081825	WATER	1-Methylnaphthalene *	670.000	J	0.73	5.5	ug/L
Total Tics :				1,618.80				
Total Concentration:				6,482.70				

Hit Summary Sheet SW-846

SDG No.: Q2900
Client: AECOM

Sample ID	Client ID		Parameter	Concentration	C	MDL	RDL	Units
Client ID : MN-12C-081825DL								
Q2900-02DL	MN-12C-081825DL	WATER	Naphthalene	4,600.000	ED	11	110	ug/L
Q2900-02DL	MN-12C-081825DL	WATER	2-Methylnaphthalene	550.000	D	12.3	110	ug/L
Q2900-02DL	MN-12C-081825DL	WATER	Acenaphthylene	200.000	D	16.5	110	ug/L
Total Svoc :				5,350.00				
Total Concentration:				5,350.00				
Client ID : MN-12C-081825DL2								
Q2900-02DL2	MN-12C-081825DL2	WATER	Naphthalene	8,000.000	D	54.9	550	ug/L
Q2900-02DL2	MN-12C-081825DL2	WATER	2-Methylnaphthalene	690.000	D	61.5	550	ug/L
Q2900-02DL2	MN-12C-081825DL2	WATER	Acenaphthylene	240.000	JD	82.4	550	ug/L
Total Svoc :				8,930.00				
Total Concentration:				8,930.00				
Client ID : DUP-01-081825								
Q2900-03	DUP-01-081825	WATER	Naphthalene	4,300.000	E	0.54	5.4	ug/L
Q2900-03	DUP-01-081825	WATER	2-Methylnaphthalene	900.000	E	0.6	5.4	ug/L
Q2900-03	DUP-01-081825	WATER	1,1-Biphenyl	30.800		0.57	5.4	ug/L
Q2900-03	DUP-01-081825	WATER	Acenaphthylene	170.000	E	0.81	5.4	ug/L
Q2900-03	DUP-01-081825	WATER	Acenaphthene	28.900		0.59	5.4	ug/L
Q2900-03	DUP-01-081825	WATER	Dibenzofuran	4.400	J	0.66	5.4	ug/L
Q2900-03	DUP-01-081825	WATER	Fluorene	25.900		0.68	5.4	ug/L
Q2900-03	DUP-01-081825	WATER	Phenanthrene	13.900		0.54	5.4	ug/L
Q2900-03	DUP-01-081825	WATER	Carbazole	14.700		0.77	5.4	ug/L
Total Svoc :				5,488.60				
Q2900-03	DUP-01-081825	WATER	.alpha.-Methylstyrene	*	8.300	J	0	ug/L
Q2900-03	DUP-01-081825	WATER	Naphthalene, 1,3-dimethyl-	*	14.900	J	0	ug/L
Q2900-03	DUP-01-081825	WATER	Naphthalene, 1,5-dimethyl-	*	29.100	J	0	ug/L
Q2900-03	DUP-01-081825	WATER	Naphthalene, 1,6-dimethyl-	*	24.900	J	0	ug/L
Q2900-03	DUP-01-081825	WATER	Naphthalene, 1,7-dimethyl-	*	12.700	J	0	ug/L
Q2900-03	DUP-01-081825	WATER	Naphthalene, 2-ethenyl-	*	13.300	J	0	ug/L
Q2900-03	DUP-01-081825	WATER	1,3,5-Cycloheptatriene	*	160.000	J	0	ug/L
Q2900-03	DUP-01-081825	WATER	1,3-Cyclopentadiene, 5-(1-methyl-	*	91.200	J	0	ug/L
Q2900-03	DUP-01-081825	WATER	Benzene, 1,2,3-trimethyl-	*	65.800	J	0	ug/L
Q2900-03	DUP-01-081825	WATER	Benzene, 1,2,4-trimethyl-	*	21.500	J	0	ug/L
Q2900-03	DUP-01-081825	WATER	Benzene, 1-ethyl-2-methyl-	*	35.900	J	0	ug/L
Q2900-03	DUP-01-081825	WATER	Benzene, 1-ethyl-4-methyl-	*	20.000	J	0	ug/L
Q2900-03	DUP-01-081825	WATER	Bicyclo[4.2.0]octa-1,3,5-triene	*	110.000	J	0	ug/L
Q2900-03	DUP-01-081825	WATER	Butane, 2-methoxy-2-methyl-	*	17.400	J	0	ug/L
Q2900-03	DUP-01-081825	WATER	Tetracyclo[3.3.1.0(2,8).0(4,6)]-no	*	7.900	J	0	ug/L
Q2900-03	DUP-01-081825	WATER	trans-Cinnamyl bromide	*	7.300	J	0	ug/L
Q2900-03	DUP-01-081825	WATER	Indane	*	54.000	J	0	ug/L

Hit Summary Sheet SW-846

SDG No.: Q2900
Client: AECOM

Sample ID	Client ID		Parameter	Concentration	C	MDL		RDL	Units
Q2900-03	DUP-01-081825	WATER	Indene	*	180.000	J	0	0	ug/L
Q2900-03	DUP-01-081825	WATER	1-Methylnaphthalene	*	770.000	J	0.71	5.4	ug/L
Total Tics :				1,644.20					
Total Concentration:				7,132.80					
Client ID : DUP-01-081825DL									
Q2900-03DL	DUP-01-081825DL	WATER	Naphthalene		4,300.000	ED	10.8	110	ug/L
Q2900-03DL	DUP-01-081825DL	WATER	2-Methylnaphthalene		520.000	D	12	110	ug/L
Q2900-03DL	DUP-01-081825DL	WATER	Acenaphthylene		190.000	D	16.1	110	ug/L
Total Svoc :				5,010.00					
Total Concentration:				5,010.00					
Client ID : DUP-01-081825DL2									
Q2900-03DL2	DUP-01-081825DL2	WATER	Naphthalene		8,300.000	D	53.8	540	ug/L
Q2900-03DL2	DUP-01-081825DL2	WATER	2-Methylnaphthalene		760.000	D	60.2	540	ug/L
Q2900-03DL2	DUP-01-081825DL2	WATER	Acenaphthylene		270.000	JD	80.6	540	ug/L
Total Svoc :				9,330.00					
Total Concentration:				9,330.00					
Client ID : MW-17B-081825									
Q2900-04	MW-17B-081825	WATER	2-Pentanone, 4-hydroxy-4-methyl	*	5.200	AB	0	0	ug/L
Q2900-04	MW-17B-081825	WATER	Butane, 2-methoxy-2-methyl-	*	120.000	J	0	0	ug/L
Total Tics :				125.20					
Total Concentration:				125.20					
Client ID : MW-11C-081825									
Q2900-05	MW-11C-081825	WATER	Naphthalene		2,200.000	E	0.53	5.3	ug/L
Q2900-05	MW-11C-081825	WATER	2-Methylnaphthalene		160.000	E	0.59	5.3	ug/L
Q2900-05	MW-11C-081825	WATER	1,1-Biphenyl		24.000		0.56	5.3	ug/L
Q2900-05	MW-11C-081825	WATER	Acenaphthylene		110.000	E	0.79	5.3	ug/L
Q2900-05	MW-11C-081825	WATER	Acenaphthene		69.200		0.58	5.3	ug/L
Q2900-05	MW-11C-081825	WATER	Dibenzofuran		5.000	J	0.64	5.3	ug/L
Q2900-05	MW-11C-081825	WATER	Fluorene		17.900		0.66	5.3	ug/L
Q2900-05	MW-11C-081825	WATER	Phenanthrene		23.600		0.53	5.3	ug/L
Q2900-05	MW-11C-081825	WATER	Anthracene		3.300	J	0.64	5.3	ug/L
Q2900-05	MW-11C-081825	WATER	Carbazole		13.100		0.76	5.3	ug/L
Total Svoc :				2,626.10					
Q2900-05	MW-11C-081825	WATER	1,3,5-Cycloheptatriene	*	400.000	J	0	0	ug/L
Q2900-05	MW-11C-081825	WATER	1,3-Cyclopentadiene, 5-(1-methyl-	*	360.000	J	0	0	ug/L
Q2900-05	MW-11C-081825	WATER	1H-Indene, 1-methyl-	*	4.200	J	0	0	ug/L
Q2900-05	MW-11C-081825	WATER	Benzene, (1-methylethyl)-	*	39.600	J	0	0	ug/L
Q2900-05	MW-11C-081825	WATER	Benzene, 1,2,3-trimethyl-	*	130.000	J	0	0	ug/L
Q2900-05	MW-11C-081825	WATER	Benzene, 1,3-dimethyl-	*	480.000	J	0	0	ug/L

Hit Summary Sheet SW-846

SDG No.: Q2900
Client: AECOM

Sample ID	Client ID		Parameter	Concentration	C	MDL	RDL	Units
Q2900-05	MW-11C-081825	WATER	Benzene, 1-ethyl-2-methyl-	*	29.600	J 0	0	ug/L
Q2900-05	MW-11C-081825	WATER	Benzene, 1-ethyl-4-methyl-	*	95.700	J 0	0	ug/L
Q2900-05	MW-11C-081825	WATER	Benzene, 2-propenyl-	*	25.000	J 0	0	ug/L
Q2900-05	MW-11C-081825	WATER	Benzene, 1-methyl-1,2-propadieny	*	3.500	J 0	0	ug/L
Q2900-05	MW-11C-081825	WATER	Butane, 2-methoxy-2-methyl-	*	130.000	J 0	0	ug/L
Q2900-05	MW-11C-081825	WATER	Indane	*	230.000	J 0	0	ug/L
Q2900-05	MW-11C-081825	WATER	Indene	*	890.000	J 0	0	ug/L
Q2900-05	MW-11C-081825	WATER	Naphthalene, 1,3-dimethyl-	*	17.900	J 0	0	ug/L
Q2900-05	MW-11C-081825	WATER	Naphthalene, 1,6-dimethyl-	*	25.900	J 0	0	ug/L
Q2900-05	MW-11C-081825	WATER	Naphthalene, 1-ethyl-	*	12.500	J 0	0	ug/L
Q2900-05	MW-11C-081825	WATER	Naphthalene, 2,3-dimethyl-	*	36.400	J 0	0	ug/L
Q2900-05	MW-11C-081825	WATER	Tetracyclo[3.3.1.0(2,8).0(4,6)]-no	*	37.900	J 0	0	ug/L
Q2900-05	MW-11C-081825	WATER	1-Methylnaphthalene	*	570.000	J 0.69	5.3	ug/L
Total Tics :					3,518.20			
Total Concentration:					6,144.30			
Client ID : MW-11C-081825DL								
Q2900-05DL	MW-11C-081825DL	WATER	Naphthalene		2,800.000	ED 10.5	110	ug/L
Q2900-05DL	MW-11C-081825DL	WATER	2-Methylnaphthalene		76.600	JD 11.8	110	ug/L
Q2900-05DL	MW-11C-081825DL	WATER	Acenaphthylene		120.000	D 15.8	110	ug/L
Q2900-05DL	MW-11C-081825DL	WATER	Acenaphthene		72.400	JD 11.6	110	ug/L
Total Svoc :					3,069.00			
Total Concentration:					3,069.00			
Client ID : MW-11C-081825DL2								
Q2900-05DL2	MW-11C-081825DL2	WATER	Naphthalene		5,300.000	D 52.6	530	ug/L
Total Svoc :					5,300.00			
Total Concentration:					5,300.00			
Client ID : MW-17C-081825								
Q2900-06	MW-17C-081825	WATER	2-Pentanone, 4-hydroxy-4-methyl *		4.900	AB 0	0	ug/L
Q2900-06	MW-17C-081825	WATER	Butane, 2-methoxy-2-methyl-	*	120.000	J 0	0	ug/L
Total Tics :					124.90			
Total Concentration:					124.90			
Client ID : MW-11B-081825								
Q2900-07	MW-11B-081825	WATER	Naphthalene		8.500	0.52	5.2	ug/L
Total Svoc :					8.50			
Q2900-07	MW-11B-081825	WATER	2,6-Dimethylphenyl isocyanate *		3.700	J 0	0	ug/L
Q2900-07	MW-11B-081825	WATER	2-Indolinone, 1-methyl-	*	5.600	J 0	0	ug/L
Q2900-07	MW-11B-081825	WATER	2-Pentanone, 4-hydroxy-4-methyl *		4.200	AB 0	0	ug/L
Q2900-07	MW-11B-081825	WATER	Benzene, (1-methylethyl)-	*	6.200	J 0	0	ug/L
Q2900-07	MW-11B-081825	WATER	Benzene, 1-ethyl-2-methyl-	*	3.600	J 0	0	ug/L

Hit Summary Sheet SW-846

SDG No.: Q2900
Client: AECOM

Sample ID	Client ID		Parameter	Concentration	C	MDL	RDL	Units
Q2900-07	MW-11B-081825	WATER	Benzene, 2-ethenyl-1,4-dimethyl-	*	7.500	J 0	0	ug/L
Q2900-07	MW-11B-081825	WATER	Benzophenone	*	5.300	J 0	0	ug/L
Q2900-07	MW-11B-081825	WATER	Butane, 2-methoxy-2-methyl-	*	110.000	J 0	0	ug/L
Q2900-07	MW-11B-081825	WATER	Indane	*	210.000	J 0	0	ug/L
Q2900-07	MW-11B-081825	WATER	1-Methylnaphthalene	*	4.600	J 0.69	5.2	ug/L
Total Tics :					360.70			
Total Concentration:					369.20			



SAMPLE DATA

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MN-9C-081825		SDG No.:	Q2900	
Lab Sample ID:	Q2900-01		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	980	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143503.D	1	08/19/25 09:29	08/21/25 00:29	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
100-52-7	Benzaldehyde	10.2	U	4.00	10.2	ug/L
108-95-2	Phenol	5.10	U	0.93	5.10	ug/L
111-44-4	bis(2-Chloroethyl)ether	5.10	U	0.83	5.10	ug/L
95-57-8	2-Chlorophenol	5.10	U	0.59	5.10	ug/L
95-48-7	2-Methylphenol	5.10	U	1.10	5.10	ug/L
108-60-1	2,2-oxybis(1-Chloropropane)	5.10	U	1.30	5.10	ug/L
98-86-2	Acetophenone	5.10	U	0.76	5.10	ug/L
65794-96-9	3+4-Methylphenols	10.2	U	1.10	10.2	ug/L
621-64-7	n-Nitroso-di-n-propylamine	2.60	U	1.40	2.60	ug/L
67-72-1	Hexachloroethane	5.10	U	0.66	5.10	ug/L
98-95-3	Nitrobenzene	5.10	U	0.78	5.10	ug/L
78-59-1	Isophorone	5.10	U	0.77	5.10	ug/L
88-75-5	2-Nitrophenol	5.10	U	1.80	5.10	ug/L
105-67-9	2,4-Dimethylphenol	5.10	U	1.90	5.10	ug/L
111-91-1	bis(2-Chloroethoxy)methane	5.10	U	0.69	5.10	ug/L
120-83-2	2,4-Dichlorophenol	5.10	U	0.53	5.10	ug/L
91-20-3	Naphthalene	5.10	U	0.51	5.10	ug/L
106-47-8	4-Chloroaniline	5.10	U	0.86	5.10	ug/L
87-68-3	Hexachlorobutadiene	5.10	U	0.55	5.10	ug/L
105-60-2	Caprolactam	10.2	U	1.20	10.2	ug/L
59-50-7	4-Chloro-3-methylphenol	5.10	U	0.60	5.10	ug/L
91-57-6	2-Methylnaphthalene	5.10	U	0.57	5.10	ug/L
77-47-4	Hexachlorocyclopentadiene	10.2	U	3.70	10.2	ug/L
88-06-2	2,4,6-Trichlorophenol	5.10	U	0.52	5.10	ug/L
95-95-4	2,4,5-Trichlorophenol	5.10	U	0.63	5.10	ug/L
92-52-4	1,1-Biphenyl	5.10	U	0.54	5.10	ug/L
91-58-7	2-Chloronaphthalene	5.10	U	0.62	5.10	ug/L
88-74-4	2-Nitroaniline	5.10	U	1.30	5.10	ug/L
131-11-3	Dimethylphthalate	5.10	U	0.62	5.10	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MN-9C-081825		SDG No.:	Q2900	
Lab Sample ID:	Q2900-01		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	980	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143503.D	1	08/19/25 09:29	08/21/25 00:29	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
208-96-8	Acenaphthylene	5.10	U	0.77	5.10	ug/L
606-20-2	2,6-Dinitrotoluene	5.10	U	0.94	5.10	ug/L
99-09-2	3-Nitroaniline	5.10	U	1.10	5.10	ug/L
83-32-9	Acenaphthene	5.10	U	0.56	5.10	ug/L
51-28-5	2,4-Dinitrophenol	10.2	U	6.10	10.2	ug/L
100-02-7	4-Nitrophenol	10.2	U	2.40	10.2	ug/L
132-64-9	Dibenzofuran	5.10	U	0.62	5.10	ug/L
121-14-2	2,4-Dinitrotoluene	5.10	U	1.20	5.10	ug/L
84-66-2	Diethylphthalate	5.10	U	0.70	5.10	ug/L
7005-72-3	4-Chlorophenyl-phenylether	5.10	U	0.69	5.10	ug/L
86-73-7	Fluorene	5.10	U	0.64	5.10	ug/L
100-01-6	4-Nitroaniline	5.10	U	1.50	5.10	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	10.2	U	2.90	10.2	ug/L
86-30-6	n-Nitrosodiphenylamine	5.10	U	0.59	5.10	ug/L
101-55-3	4-Bromophenyl-phenylether	5.10	U	0.41	5.10	ug/L
118-74-1	Hexachlorobenzene	5.10	U	0.53	5.10	ug/L
1912-24-9	Atrazine	5.10	U	1.00	5.10	ug/L
87-86-5	Pentachlorophenol	10.2	U	1.60	10.2	ug/L
85-01-8	Phenanthrene	5.10	U	0.51	5.10	ug/L
120-12-7	Anthracene	5.10	U	0.62	5.10	ug/L
86-74-8	Carbazole	5.10	U	0.73	5.10	ug/L
84-74-2	Di-n-butylphthalate	5.10	U	1.20	5.10	ug/L
206-44-0	Fluoranthene	5.10	U	0.84	5.10	ug/L
129-00-0	Pyrene	5.10	U	0.51	5.10	ug/L
85-68-7	Butylbenzylphthalate	5.10	UQ	2.00	5.10	ug/L
91-94-1	3,3-Dichlorobenzidine	10.2	UQ	0.95	10.2	ug/L
56-55-3	Benzo(a)anthracene	5.10	U	0.46	5.10	ug/L
218-01-9	Chrysene	5.10	U	0.45	5.10	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	5.10	U	1.60	5.10	ug/L
117-84-0	Di-n-octyl phthalate	10.2	U	2.40	10.2	ug/L
205-99-2	Benzo(b)fluoranthene	5.10	U	0.50	5.10	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MN-9C-081825		SDG No.:	Q2900	
Lab Sample ID:	Q2900-01		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	980	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143503.D	1	08/19/25 09:29	08/21/25 00:29	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
207-08-9	Benzo(k)fluoranthene	5.10	U	0.49	5.10	ug/L
50-32-8	Benzo(a)pyrene	5.10	U	0.56	5.10	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	5.10	U	0.60	5.10	ug/L
53-70-3	Dibenzo(a,h)anthracene	5.10	U	0.68	5.10	ug/L
191-24-2	Benzo(g,h,i)perylene	5.10	U	0.70	5.10	ug/L
95-94-3	1,2,4,5-Tetrachlorobenzene	5.10	U	0.53	5.10	ug/L
123-91-1	1,4-Dioxane	5.10	U	1.00	5.10	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	5.10	U	0.73	5.10	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	65.0		23 - 138	43%	SPK: 150
13127-88-3	Phenol-d6	43.0		10 - 134	29%	SPK: 150
4165-60-0	Nitrobenzene-d5	93.7		67 - 132	94%	SPK: 100
321-60-8	2-Fluorobiphenyl	87.7		52 - 132	88%	SPK: 100
118-79-6	2,4,6-Tribromophenol	140		44 - 137	93%	SPK: 150
1718-51-0	Terphenyl-d14	86.7		42 - 152	87%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	128000	6.928			
1146-65-2	Naphthalene-d8	499000	8.204			
15067-26-2	Acenaphthene-d10	270000	9.963			
1517-22-2	Phenanthrene-d10	422000	11.451			
1719-03-5	Chrysene-d12	236000	14.092			
1520-96-3	Perylene-d12	262000	15.592			
TENTATIVE IDENTIFIED COMPOUNDS						
000994-05-8	Butane, 2-methoxy-2-methyl-	110	J		2.27	ug/L
000123-42-2	2-Pentanone, 4-hydroxy-4-methyl-	4.40	AB		5.16	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MN-9C-081825		SDG No.:	Q2900	
Lab Sample ID:	Q2900-01		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	980	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143503.D	1	08/19/25 09:29	08/21/25 00:29	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MN-12C-081825		SDG No.:	Q2900	
Lab Sample ID:	Q2900-02		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	910	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143511.D	1	08/19/25 09:29	08/21/25 04:21	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
100-52-7	Benzaldehyde	11.0	U	4.30	11.0	ug/L
108-95-2	Phenol	5.50	U	1.00	5.50	ug/L
111-44-4	bis(2-Chloroethyl)ether	5.50	U	0.89	5.50	ug/L
95-57-8	2-Chlorophenol	5.50	U	0.64	5.50	ug/L
95-48-7	2-Methylphenol	5.50	U	1.20	5.50	ug/L
108-60-1	2,2-oxybis(1-Chloropropane)	5.50	U	1.40	5.50	ug/L
98-86-2	Acetophenone	5.50	U	0.81	5.50	ug/L
65794-96-9	3+4-Methylphenols	11.0	U	1.20	11.0	ug/L
621-64-7	n-Nitroso-di-n-propylamine	2.70	U	1.50	2.70	ug/L
67-72-1	Hexachloroethane	5.50	U	0.71	5.50	ug/L
98-95-3	Nitrobenzene	5.50	U	0.84	5.50	ug/L
78-59-1	Isophorone	5.50	U	0.82	5.50	ug/L
88-75-5	2-Nitrophenol	5.50	U	1.90	5.50	ug/L
105-67-9	2,4-Dimethylphenol	5.50	U	2.00	5.50	ug/L
111-91-1	bis(2-Chloroethoxy)methane	5.50	U	0.75	5.50	ug/L
120-83-2	2,4-Dichlorophenol	5.50	U	0.57	5.50	ug/L
91-20-3	Naphthalene	3800	E	0.55	5.50	ug/L
106-47-8	4-Chloroaniline	5.50	U	0.92	5.50	ug/L
87-68-3	Hexachlorobutadiene	5.50	U	0.59	5.50	ug/L
105-60-2	Caprolactam	11.0	U	1.20	11.0	ug/L
59-50-7	4-Chloro-3-methylphenol	5.50	U	0.65	5.50	ug/L
91-57-6	2-Methylnaphthalene	800	E	0.62	5.50	ug/L
77-47-4	Hexachlorocyclopentadiene	11.0	U	4.00	11.0	ug/L
88-06-2	2,4,6-Trichlorophenol	5.50	U	0.56	5.50	ug/L
95-95-4	2,4,5-Trichlorophenol	5.50	U	0.68	5.50	ug/L
92-52-4	1,1-Biphenyl	27.7		0.58	5.50	ug/L
91-58-7	2-Chloronaphthalene	5.50	U	0.67	5.50	ug/L
88-74-4	2-Nitroaniline	5.50	U	1.40	5.50	ug/L
131-11-3	Dimethylphthalate	5.50	U	0.67	5.50	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MN-12C-081825		SDG No.:	Q2900	
Lab Sample ID:	Q2900-02		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	910	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143511.D	1	08/19/25 09:29	08/21/25 04:21	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
208-96-8	Acenaphthylene	150	E	0.82	5.50	ug/L
606-20-2	2,6-Dinitrotoluene	5.50	U	1.00	5.50	ug/L
99-09-2	3-Nitroaniline	5.50	U	1.20	5.50	ug/L
83-32-9	Acenaphthene	30.2		0.60	5.50	ug/L
51-28-5	2,4-Dinitrophenol	11.0	U	6.60	11.0	ug/L
100-02-7	4-Nitrophenol	11.0	U	2.60	11.0	ug/L
132-64-9	Dibenzofuran	3.80	J	0.67	5.50	ug/L
121-14-2	2,4-Dinitrotoluene	5.50	U	1.30	5.50	ug/L
84-66-2	Diethylphthalate	5.50	U	0.76	5.50	ug/L
7005-72-3	4-Chlorophenyl-phenylether	5.50	U	0.75	5.50	ug/L
86-73-7	Fluorene	27.1		0.69	5.50	ug/L
100-01-6	4-Nitroaniline	5.50	U	1.60	5.50	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	11.0	U	3.20	11.0	ug/L
86-30-6	n-Nitrosodiphenylamine	5.50	U	0.64	5.50	ug/L
101-55-3	4-Bromophenyl-phenylether	5.50	U	0.44	5.50	ug/L
118-74-1	Hexachlorobenzene	5.50	U	0.57	5.50	ug/L
1912-24-9	Atrazine	5.50	U	1.10	5.50	ug/L
87-86-5	Pentachlorophenol	11.0	U	1.70	11.0	ug/L
85-01-8	Phenanthrene	11.4		0.55	5.50	ug/L
120-12-7	Anthracene	5.50	U	0.67	5.50	ug/L
86-74-8	Carbazole	13.7		0.79	5.50	ug/L
84-74-2	Di-n-butylphthalate	5.50	U	1.30	5.50	ug/L
206-44-0	Fluoranthene	5.50	U	0.90	5.50	ug/L
129-00-0	Pyrene	5.50	U	0.55	5.50	ug/L
85-68-7	Butylbenzylphthalate	5.50	UQ	2.10	5.50	ug/L
91-94-1	3,3-Dichlorobenzidine	11.0	UQ	1.00	11.0	ug/L
56-55-3	Benzo(a)anthracene	5.50	U	0.49	5.50	ug/L
218-01-9	Chrysene	5.50	U	0.48	5.50	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	5.50	U	1.80	5.50	ug/L
117-84-0	Di-n-octyl phthalate	11.0	U	2.60	11.0	ug/L
205-99-2	Benzo(b)fluoranthene	5.50	U	0.54	5.50	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MN-12C-081825		SDG No.:	Q2900	
Lab Sample ID:	Q2900-02		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	910	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143511.D	1	08/19/25 09:29	08/21/25 04:21	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
207-08-9	Benzo(k)fluoranthene	5.50	U	0.53	5.50	ug/L
50-32-8	Benzo(a)pyrene	5.50	U	0.60	5.50	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	5.50	U	0.65	5.50	ug/L
53-70-3	Dibenzo(a,h)anthracene	5.50	U	0.74	5.50	ug/L
191-24-2	Benzo(g,h,i)perylene	5.50	U	0.76	5.50	ug/L
95-94-3	1,2,4,5-Tetrachlorobenzene	5.50	U	0.57	5.50	ug/L
123-91-1	1,4-Dioxane	5.50	U	1.10	5.50	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	5.50	U	0.79	5.50	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	33.1	*	23 - 138	22%	SPK: 150
13127-88-3	Phenol-d6	42.2		10 - 134	28%	SPK: 150
4165-60-0	Nitrobenzene-d5	246	*	67 - 132	246%	SPK: 100
321-60-8	2-Fluorobiphenyl	82.2		52 - 132	82%	SPK: 100
118-79-6	2,4,6-Tribromophenol	133		44 - 137	89%	SPK: 150
1718-51-0	Terphenyl-d14	87.4		42 - 152	87%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	122000		6.934		
1146-65-2	Naphthalene-d8	165000		8.228		
15067-26-2	Acenaphthene-d10	245000		9.963		
1517-22-2	Phenanthrene-d10	368000		11.451		
1719-03-5	Chrysene-d12	219000		14.092		
1520-96-3	Perylene-d12	260000		15.592		
TENTATIVE IDENTIFIED COMPOUNDS						
000994-05-8	Butane, 2-methoxy-2-methyl-	19.6	J		2.26	ug/L
000544-25-2	1,3,5-Cycloheptatriene	180	J		4.13	ug/L
002175-91-9	1,3-Cyclopentadiene, 5-(1-methylet	110	J		5.46	ug/L
000694-87-1	Bicyclo[4.2.0]octa-1,3,5-triene	120	J		5.82	ug/L
000611-14-3	Benzene, 1-ethyl-2-methyl-	40.8	J		6.47	ug/L
000622-96-8	Benzene, 1-ethyl-4-methyl-	21.6	J		6.50	ug/L
000620-14-4	Benzene, 1-ethyl-3-methyl-	7.30	J		6.62	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MN-12C-081825		SDG No.:	Q2900	
Lab Sample ID:	Q2900-02		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	910	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143511.D	1	08/19/25 09:29	08/21/25 04:21	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
000098-83-9	.alpha.-Methylstyrene	7.80	J		6.65	ug/L
000611-15-4	Benzene, 1-ethenyl-2-methyl-	68.0	J		6.78	ug/L
128600-88-4	1,3-Methanopentalene, 1,2,3,5-tetr	8.70	J		6.81	ug/L
000095-63-6	Benzene, 1,2,4-trimethyl-	22.0	J		7.00	ug/L
	unknown7.045	8.40	J		7.04	ug/L
000496-11-7	Indane	64.8	J		7.13	ug/L
000095-13-6	Indene	200	J		7.25	ug/L
000768-49-0	Benzene, (2-methyl-1-propenyl)-	7.70	J		7.54	ug/L
90-12-0	1-Methylnaphthalene	670	J		9.04	ug/L
000582-16-1	Naphthalene, 2,7-dimethyl-	22.2	J		9.55	ug/L
000575-43-9	Naphthalene, 1,6-dimethyl-	26.4	J		9.62	ug/L
000581-40-8	Naphthalene, 2,3-dimethyl-	13.5	J		9.74	ug/L

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MN-12C-081825DL		SDG No.:	Q2900	
Lab Sample ID:	Q2900-02DL		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	910	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143522.D	20	08/19/25 09:29	08/21/25 11:18	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
100-52-7	Benzaldehyde	220	UD	85.9	220	ug/L
108-95-2	Phenol	110	UD	20.0	110	ug/L
111-44-4	bis(2-Chloroethyl)ether	110	UD	17.8	110	ug/L
95-57-8	2-Chlorophenol	110	UD	12.7	110	ug/L
95-48-7	2-Methylphenol	110	UD	24.6	110	ug/L
108-60-1	2,2-oxybis(1-Chloropropane)	110	UD	28.1	110	ug/L
98-86-2	Acetophenone	110	UD	16.3	110	ug/L
65794-96-9	3+4-Methylphenols	220	UD	24.2	220	ug/L
621-64-7	n-Nitroso-di-n-propylamine	54.9	UD	31.0	54.9	ug/L
67-72-1	Hexachloroethane	110	UD	14.3	110	ug/L
98-95-3	Nitrobenzene	110	UD	16.7	110	ug/L
78-59-1	Isophorone	110	UD	16.5	110	ug/L
88-75-5	2-Nitrophenol	110	UD	38.7	110	ug/L
105-67-9	2,4-Dimethylphenol	110	UD	40.7	110	ug/L
111-91-1	bis(2-Chloroethoxy)methane	110	UD	14.9	110	ug/L
120-83-2	2,4-Dichlorophenol	110	UD	11.4	110	ug/L
91-20-3	Naphthalene	4600	ED	11.0	110	ug/L
106-47-8	4-Chloroaniline	110	UD	18.5	110	ug/L
87-68-3	Hexachlorobutadiene	110	UD	11.9	110	ug/L
105-60-2	Caprolactam	220	UD	24.8	220	ug/L
59-50-7	4-Chloro-3-methylphenol	110	UD	13.0	110	ug/L
91-57-6	2-Methylnaphthalene	550	D	12.3	110	ug/L
77-47-4	Hexachlorocyclopentadiene	220	UD	79.8	220	ug/L
88-06-2	2,4,6-Trichlorophenol	110	UD	11.2	110	ug/L
95-95-4	2,4,5-Trichlorophenol	110	UD	13.6	110	ug/L
92-52-4	1,1-Biphenyl	110	UD	11.6	110	ug/L
91-58-7	2-Chloronaphthalene	110	UD	13.4	110	ug/L
88-74-4	2-Nitroaniline	110	UD	27.7	110	ug/L
131-11-3	Dimethylphthalate	110	UD	13.4	110	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MN-12C-081825DL		SDG No.:	Q2900	
Lab Sample ID:	Q2900-02DL		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	910	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143522.D	20	08/19/25 09:29	08/21/25 11:18	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
208-96-8	Acenaphthylene	200	D	16.5	110	ug/L
606-20-2	2,6-Dinitrotoluene	110	UD	20.2	110	ug/L
99-09-2	3-Nitroaniline	110	UD	23.1	110	ug/L
83-32-9	Acenaphthene	110	UD	12.1	110	ug/L
51-28-5	2,4-Dinitrophenol	220	UD	130	220	ug/L
100-02-7	4-Nitrophenol	220	UD	52.3	220	ug/L
132-64-9	Dibenzofuran	110	UD	13.4	110	ug/L
121-14-2	2,4-Dinitrotoluene	110	UD	26.8	110	ug/L
84-66-2	Diethylphthalate	110	UD	15.2	110	ug/L
7005-72-3	4-Chlorophenyl-phenylether	110	UD	14.9	110	ug/L
86-73-7	Fluorene	110	UD	13.8	110	ug/L
100-01-6	4-Nitroaniline	110	UD	33.0	110	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	220	UD	63.3	220	ug/L
86-30-6	n-Nitrosodiphenylamine	110	UD	12.7	110	ug/L
101-55-3	4-Bromophenyl-phenylether	110	UD	8.80	110	ug/L
118-74-1	Hexachlorobenzene	110	UD	11.4	110	ug/L
1912-24-9	Atrazine	110	UD	22.2	110	ug/L
87-86-5	Pentachlorophenol	220	UD	34.7	220	ug/L
85-01-8	Phenanthrene	110	UD	11.0	110	ug/L
120-12-7	Anthracene	110	UD	13.4	110	ug/L
86-74-8	Carbazole	110	UD	15.8	110	ug/L
84-74-2	Di-n-butylphthalate	110	UD	26.8	110	ug/L
206-44-0	Fluoranthene	110	UD	18.0	110	ug/L
129-00-0	Pyrene	110	UD	11.0	110	ug/L
85-68-7	Butylbenzylphthalate	110	UDQ	42.4	110	ug/L
91-94-1	3,3-Dichlorobenzidine	220	UDQ	20.4	220	ug/L
56-55-3	Benzo(a)anthracene	110	UD	9.90	110	ug/L
218-01-9	Chrysene	110	UD	9.70	110	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	110	UD	35.2	110	ug/L
117-84-0	Di-n-octyl phthalate	220	UD	51.4	220	ug/L
205-99-2	Benzo(b)fluoranthene	110	UD	10.8	110	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MN-12C-081825DL		SDG No.:	Q2900	
Lab Sample ID:	Q2900-02DL		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	910	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143522.D	20	08/19/25 09:29	08/21/25 11:18	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
207-08-9	Benzo(k)fluoranthene	110	UD	10.5	110	ug/L
50-32-8	Benzo(a)pyrene	110	UD	12.1	110	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	110	UD	13.0	110	ug/L
53-70-3	Dibenzo(a,h)anthracene	110	UD	14.7	110	ug/L
191-24-2	Benzo(g,h,i)perylene	110	UD	15.2	110	ug/L
95-94-3	1,2,4,5-Tetrachlorobenzene	110	UD	11.4	110	ug/L
123-91-1	1,4-Dioxane	110	UD	22.0	110	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	110	UD	15.8	110	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	51.2		23 - 138	34%	SPK: 150
13127-88-3	Phenol-d6	37.2		10 - 134	25%	SPK: 150
4165-60-0	Nitrobenzene-d5	84.4		67 - 132	84%	SPK: 100
321-60-8	2-Fluorobiphenyl	97.9		52 - 132	98%	SPK: 100
118-79-6	2,4,6-Tribromophenol	98.2		44 - 137	65%	SPK: 150
1718-51-0	Terphenyl-d14	80.9		42 - 152	81%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	110000	6.928			
1146-65-2	Naphthalene-d8	414000	8.21			
15067-26-2	Acenaphthene-d10	236000	9.963			
1517-22-2	Phenanthrene-d10	348000	11.451			
1719-03-5	Chrysene-d12	223000	14.092			
1520-96-3	Perylene-d12	252000	15.592			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MN-12C-081825DL2		SDG No.:	Q2900	
Lab Sample ID:	Q2900-02DL2		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	910	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143527.D	100	08/19/25 09:29	08/21/25 13:44	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
100-52-7	Benzaldehyde	1100	UD	430	1100	ug/L
108-95-2	Phenol	550	UD	100	550	ug/L
111-44-4	bis(2-Chloroethyl)ether	550	UD	89.0	550	ug/L
95-57-8	2-Chlorophenol	550	UD	63.7	550	ug/L
95-48-7	2-Methylphenol	550	UD	120	550	ug/L
108-60-1	2,2-oxybis(1-Chloropropane)	550	UD	140	550	ug/L
98-86-2	Acetophenone	550	UD	81.3	550	ug/L
65794-96-9	3+4-Methylphenols	1100	UD	120	1100	ug/L
621-64-7	n-Nitroso-di-n-propylamine	270	UD	150	270	ug/L
67-72-1	Hexachloroethane	550	UD	71.4	550	ug/L
98-95-3	Nitrobenzene	550	UD	83.5	550	ug/L
78-59-1	Isophorone	550	UD	82.4	550	ug/L
88-75-5	2-Nitrophenol	550	UD	190	550	ug/L
105-67-9	2,4-Dimethylphenol	550	UD	200	550	ug/L
111-91-1	bis(2-Chloroethoxy)methane	550	UD	74.7	550	ug/L
120-83-2	2,4-Dichlorophenol	550	UD	57.1	550	ug/L
91-20-3	Naphthalene	8000	D	54.9	550	ug/L
106-47-8	4-Chloroaniline	550	UD	92.3	550	ug/L
87-68-3	Hexachlorobutadiene	550	UD	59.3	550	ug/L
105-60-2	Caprolactam	1100	UD	120	1100	ug/L
59-50-7	4-Chloro-3-methylphenol	550	UD	64.8	550	ug/L
91-57-6	2-Methylnaphthalene	690	D	61.5	550	ug/L
77-47-4	Hexachlorocyclopentadiene	1100	UD	400	1100	ug/L
88-06-2	2,4,6-Trichlorophenol	550	UD	56.0	550	ug/L
95-95-4	2,4,5-Trichlorophenol	550	UD	68.1	550	ug/L
92-52-4	1,1-Biphenyl	550	UD	58.2	550	ug/L
91-58-7	2-Chloronaphthalene	550	UD	67.0	550	ug/L
88-74-4	2-Nitroaniline	550	UD	140	550	ug/L
131-11-3	Dimethylphthalate	550	UD	67.0	550	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MN-12C-081825DL2		SDG No.:	Q2900	
Lab Sample ID:	Q2900-02DL2		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	910	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143527.D	100	08/19/25 09:29	08/21/25 13:44	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
208-96-8	Acenaphthylene	240	JD	82.4	550	ug/L
606-20-2	2,6-Dinitrotoluene	550	UD	100	550	ug/L
99-09-2	3-Nitroaniline	550	UD	120	550	ug/L
83-32-9	Acenaphthene	550	UD	60.4	550	ug/L
51-28-5	2,4-Dinitrophenol	1100	UD	660	1100	ug/L
100-02-7	4-Nitrophenol	1100	UD	260	1100	ug/L
132-64-9	Dibenzofuran	550	UD	67.0	550	ug/L
121-14-2	2,4-Dinitrotoluene	550	UD	130	550	ug/L
84-66-2	Diethylphthalate	550	UD	75.8	550	ug/L
7005-72-3	4-Chlorophenyl-phenylether	550	UD	74.7	550	ug/L
86-73-7	Fluorene	550	UD	69.2	550	ug/L
100-01-6	4-Nitroaniline	550	UD	160	550	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	1100	UD	320	1100	ug/L
86-30-6	n-Nitrosodiphenylamine	550	UD	63.7	550	ug/L
101-55-3	4-Bromophenyl-phenylether	550	UD	44.0	550	ug/L
118-74-1	Hexachlorobenzene	550	UD	57.1	550	ug/L
1912-24-9	Atrazine	550	UD	110	550	ug/L
87-86-5	Pentachlorophenol	1100	UD	170	1100	ug/L
85-01-8	Phenanthrene	550	UD	54.9	550	ug/L
120-12-7	Anthracene	550	UD	67.0	550	ug/L
86-74-8	Carbazole	550	UD	79.1	550	ug/L
84-74-2	Di-n-butylphthalate	550	UD	130	550	ug/L
206-44-0	Fluoranthene	550	UD	90.1	550	ug/L
129-00-0	Pyrene	550	UD	54.9	550	ug/L
85-68-7	Butylbenzylphthalate	550	UDQ	210	550	ug/L
91-94-1	3,3-Dichlorobenzidine	1100	UDQ	100	1100	ug/L
56-55-3	Benzo(a)anthracene	550	UD	49.5	550	ug/L
218-01-9	Chrysene	550	UD	48.4	550	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	550	UD	180	550	ug/L
117-84-0	Di-n-octyl phthalate	1100	UD	260	1100	ug/L
205-99-2	Benzo(b)fluoranthene	550	UD	53.8	550	ug/L

Report of Analysis

Client:	AECOM	Date Collected:	08/18/25
Project:	National Grid Equity - Brooklyn NY	Date Received:	08/18/25
Client Sample ID:	MN-12C-081825DL2	SDG No.:	Q2900
Lab Sample ID:	Q2900-02DL2	Matrix:	Water
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	910 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-TCL BNA -20
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143527.D	100	08/19/25 09:29	08/21/25 13:44	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
207-08-9	Benzo(k)fluoranthene	550	UD	52.7	550	ug/L
50-32-8	Benzo(a)pyrene	550	UD	60.4	550	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	550	UD	64.8	550	ug/L
53-70-3	Dibenzo(a,h)anthracene	550	UD	73.6	550	ug/L
191-24-2	Benzo(g,h,i)perylene	550	UD	75.8	550	ug/L
95-94-3	1,2,4,5-Tetrachlorobenzene	550	UD	57.1	550	ug/L
123-91-1	1,4-Dioxane	550	UD	110	550	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	550	UD	79.1	550	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	46.4		23 - 138	31%	SPK: 150
13127-88-3	Phenol-d6	42.2		10 - 134	28%	SPK: 150
4165-60-0	Nitrobenzene-d5	91.9		67 - 132	92%	SPK: 100
321-60-8	2-Fluorobiphenyl	133	*	52 - 132	133%	SPK: 100
118-79-6	2,4,6-Tribromophenol	74.4		44 - 137	50%	SPK: 150
1718-51-0	Terphenyl-d14	96.7		42 - 152	97%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	101000	6.928			
1146-65-2	Naphthalene-d8	382000	8.21			
15067-26-2	Acenaphthene-d10	199000	9.963			
1517-22-2	Phenanthrene-d10	282000	11.451			
1719-03-5	Chrysene-d12	192000	14.098			
1520-96-3	Perylene-d12	235000	15.598			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	DUP-01-081825		SDG No.:	Q2900	
Lab Sample ID:	Q2900-03		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	930	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143512.D	1	08/19/25 09:29	08/21/25 04:49	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
100-52-7	Benzaldehyde	10.8	U	4.20	10.8	ug/L
108-95-2	Phenol	5.40	U	0.98	5.40	ug/L
111-44-4	bis(2-Chloroethyl)ether	5.40	U	0.87	5.40	ug/L
95-57-8	2-Chlorophenol	5.40	U	0.62	5.40	ug/L
95-48-7	2-Methylphenol	5.40	U	1.20	5.40	ug/L
108-60-1	2,2-oxybis(1-Chloropropane)	5.40	U	1.40	5.40	ug/L
98-86-2	Acetophenone	5.40	U	0.80	5.40	ug/L
65794-96-9	3+4-Methylphenols	10.8	U	1.20	10.8	ug/L
621-64-7	n-Nitroso-di-n-propylamine	2.70	U	1.50	2.70	ug/L
67-72-1	Hexachloroethane	5.40	U	0.70	5.40	ug/L
98-95-3	Nitrobenzene	5.40	U	0.82	5.40	ug/L
78-59-1	Isophorone	5.40	U	0.81	5.40	ug/L
88-75-5	2-Nitrophenol	5.40	U	1.90	5.40	ug/L
105-67-9	2,4-Dimethylphenol	5.40	U	2.00	5.40	ug/L
111-91-1	bis(2-Chloroethoxy)methane	5.40	U	0.73	5.40	ug/L
120-83-2	2,4-Dichlorophenol	5.40	U	0.56	5.40	ug/L
91-20-3	Naphthalene	4300	E	0.54	5.40	ug/L
106-47-8	4-Chloroaniline	5.40	U	0.90	5.40	ug/L
87-68-3	Hexachlorobutadiene	5.40	U	0.58	5.40	ug/L
105-60-2	Caprolactam	10.8	U	1.20	10.8	ug/L
59-50-7	4-Chloro-3-methylphenol	5.40	U	0.63	5.40	ug/L
91-57-6	2-Methylnaphthalene	900	E	0.60	5.40	ug/L
77-47-4	Hexachlorocyclopentadiene	10.8	U	3.90	10.8	ug/L
88-06-2	2,4,6-Trichlorophenol	5.40	U	0.55	5.40	ug/L
95-95-4	2,4,5-Trichlorophenol	5.40	U	0.67	5.40	ug/L
92-52-4	1,1-Biphenyl	30.8		0.57	5.40	ug/L
91-58-7	2-Chloronaphthalene	5.40	U	0.66	5.40	ug/L
88-74-4	2-Nitroaniline	5.40	U	1.40	5.40	ug/L
131-11-3	Dimethylphthalate	5.40	U	0.66	5.40	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	DUP-01-081825		SDG No.:	Q2900	
Lab Sample ID:	Q2900-03		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	930	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143512.D	1	08/19/25 09:29	08/21/25 04:49	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
208-96-8	Acenaphthylene	170	E	0.81	5.40	ug/L
606-20-2	2,6-Dinitrotoluene	5.40	U	0.99	5.40	ug/L
99-09-2	3-Nitroaniline	5.40	U	1.10	5.40	ug/L
83-32-9	Acenaphthene	28.9		0.59	5.40	ug/L
51-28-5	2,4-Dinitrophenol	10.8	U	6.40	10.8	ug/L
100-02-7	4-Nitrophenol	10.8	U	2.60	10.8	ug/L
132-64-9	Dibenzofuran	4.40	J	0.66	5.40	ug/L
121-14-2	2,4-Dinitrotoluene	5.40	U	1.30	5.40	ug/L
84-66-2	Diethylphthalate	5.40	U	0.74	5.40	ug/L
7005-72-3	4-Chlorophenyl-phenylether	5.40	U	0.73	5.40	ug/L
86-73-7	Fluorene	25.9		0.68	5.40	ug/L
100-01-6	4-Nitroaniline	5.40	U	1.60	5.40	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	10.8	U	3.10	10.8	ug/L
86-30-6	n-Nitrosodiphenylamine	5.40	U	0.62	5.40	ug/L
101-55-3	4-Bromophenyl-phenylether	5.40	U	0.43	5.40	ug/L
118-74-1	Hexachlorobenzene	5.40	U	0.56	5.40	ug/L
1912-24-9	Atrazine	5.40	U	1.10	5.40	ug/L
87-86-5	Pentachlorophenol	10.8	U	1.70	10.8	ug/L
85-01-8	Phenanthrene	13.9		0.54	5.40	ug/L
120-12-7	Anthracene	5.40	U	0.66	5.40	ug/L
86-74-8	Carbazole	14.7		0.77	5.40	ug/L
84-74-2	Di-n-butylphthalate	5.40	U	1.30	5.40	ug/L
206-44-0	Fluoranthene	5.40	U	0.88	5.40	ug/L
129-00-0	Pyrene	5.40	U	0.54	5.40	ug/L
85-68-7	Butylbenzylphthalate	5.40	UQ	2.10	5.40	ug/L
91-94-1	3,3-Dichlorobenzidine	10.8	UQ	1.00	10.8	ug/L
56-55-3	Benzo(a)anthracene	5.40	U	0.48	5.40	ug/L
218-01-9	Chrysene	5.40	U	0.47	5.40	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	5.40	U	1.70	5.40	ug/L
117-84-0	Di-n-octyl phthalate	10.8	U	2.50	10.8	ug/L
205-99-2	Benzo(b)fluoranthene	5.40	U	0.53	5.40	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	DUP-01-081825		SDG No.:	Q2900	
Lab Sample ID:	Q2900-03		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	930	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143512.D	1	08/19/25 09:29	08/21/25 04:49	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
207-08-9	Benzo(k)fluoranthene	5.40	U	0.52	5.40	ug/L
50-32-8	Benzo(a)pyrene	5.40	U	0.59	5.40	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	5.40	U	0.63	5.40	ug/L
53-70-3	Dibenzo(a,h)anthracene	5.40	U	0.72	5.40	ug/L
191-24-2	Benzo(g,h,i)perylene	5.40	U	0.74	5.40	ug/L
95-94-3	1,2,4,5-Tetrachlorobenzene	5.40	U	0.56	5.40	ug/L
123-91-1	1,4-Dioxane	5.40	U	1.10	5.40	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	5.40	U	0.77	5.40	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	29.6	*	23 - 138	20%	SPK: 150
13127-88-3	Phenol-d6	39.9		10 - 134	27%	SPK: 150
4165-60-0	Nitrobenzene-d5	258	*	67 - 132	258%	SPK: 100
321-60-8	2-Fluorobiphenyl	79.8		52 - 132	80%	SPK: 100
118-79-6	2,4,6-Tribromophenol	132		44 - 137	88%	SPK: 150
1718-51-0	Terphenyl-d14	82.7		42 - 152	83%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	128000		6.94		
1146-65-2	Naphthalene-d8	162000		8.228		
15067-26-2	Acenaphthene-d10	252000		9.963		
1517-22-2	Phenanthrene-d10	378000		11.451		
1719-03-5	Chrysene-d12	229000		14.092		
1520-96-3	Perylene-d12	281000		15.592		
TENTATIVE IDENTIFIED COMPOUNDS						
000994-05-8	Butane, 2-methoxy-2-methyl-	17.4	J		2.27	ug/L
000544-25-2	1,3,5-Cycloheptatriene	160	J		4.14	ug/L
002175-91-9	1,3-Cyclopentadiene, 5-(1-methylet	91.2	J		5.46	ug/L
000694-87-1	Bicyclo[4.2.0]octa-1,3,5-triene	110	J		5.83	ug/L
000611-14-3	Benzene, 1-ethyl-2-methyl-	35.9	J		6.47	ug/L
000622-96-8	Benzene, 1-ethyl-4-methyl-	20.0	J		6.50	ug/L
000098-83-9	.alpha.-Methylstyrene	8.30	J		6.64	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	DUP-01-081825		SDG No.:	Q2900	
Lab Sample ID:	Q2900-03		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	930	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143512.D	1	08/19/25 09:29	08/21/25 04:49	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
000526-73-8	Benzene, 1,2,3-trimethyl-	65.8	J		6.78	ug/L
1000191-13-7	Tetracyclo[3.3.1.0(2,8).0(4,6)]-no	7.90	J		6.81	ug/L
000095-63-6	Benzene, 1,2,4-trimethyl-	21.5	J		7.00	ug/L
026146-77-0	trans-Cinnamyl bromide	7.30	J		7.04	ug/L
000496-11-7	Indane	54.0	J		7.13	ug/L
000095-13-6	Indene	180	J		7.26	ug/L
90-12-0	1-Methylnaphthalene	770	J		9.05	ug/L
000575-43-9	Naphthalene, 1,6-dimethyl-	24.9	J		9.55	ug/L
000571-61-9	Naphthalene, 1,5-dimethyl-	29.1	J		9.62	ug/L
000575-37-1	Naphthalene, 1,7-dimethyl-	12.7	J		9.65	ug/L
000827-54-3	Naphthalene, 2-ethenyl-	13.3	J		9.69	ug/L
000575-41-7	Naphthalene, 1,3-dimethyl-	14.9	J		9.74	ug/L

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	DUP-01-081825DL		SDG No.:	Q2900	
Lab Sample ID:	Q2900-03DL		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	930	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143523.D	20	08/19/25 09:29	08/21/25 11:47	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
100-52-7	Benzaldehyde	220	UD	84.1	220	ug/L
108-95-2	Phenol	110	UD	19.6	110	ug/L
111-44-4	bis(2-Chloroethyl)ether	110	UD	17.4	110	ug/L
95-57-8	2-Chlorophenol	110	UD	12.5	110	ug/L
95-48-7	2-Methylphenol	110	UD	24.1	110	ug/L
108-60-1	2,2-oxybis(1-Chloropropane)	110	UD	27.5	110	ug/L
98-86-2	Acetophenone	110	UD	15.9	110	ug/L
65794-96-9	3+4-Methylphenols	220	UD	23.7	220	ug/L
621-64-7	n-Nitroso-di-n-propylamine	53.8	UD	30.3	53.8	ug/L
67-72-1	Hexachloroethane	110	UD	14.0	110	ug/L
98-95-3	Nitrobenzene	110	UD	16.3	110	ug/L
78-59-1	Isophorone	110	UD	16.1	110	ug/L
88-75-5	2-Nitrophenol	110	UD	37.8	110	ug/L
105-67-9	2,4-Dimethylphenol	110	UD	39.8	110	ug/L
111-91-1	bis(2-Chloroethoxy)methane	110	UD	14.6	110	ug/L
120-83-2	2,4-Dichlorophenol	110	UD	11.2	110	ug/L
91-20-3	Naphthalene	4300	ED	10.8	110	ug/L
106-47-8	4-Chloroaniline	110	UD	18.1	110	ug/L
87-68-3	Hexachlorobutadiene	110	UD	11.6	110	ug/L
105-60-2	Caprolactam	220	UD	24.3	220	ug/L
59-50-7	4-Chloro-3-methylphenol	110	UD	12.7	110	ug/L
91-57-6	2-Methylnaphthalene	520	D	12.0	110	ug/L
77-47-4	Hexachlorocyclopentadiene	220	UD	78.1	220	ug/L
88-06-2	2,4,6-Trichlorophenol	110	UD	11.0	110	ug/L
95-95-4	2,4,5-Trichlorophenol	110	UD	13.3	110	ug/L
92-52-4	1,1-Biphenyl	110	UD	11.4	110	ug/L
91-58-7	2-Chloronaphthalene	110	UD	13.1	110	ug/L
88-74-4	2-Nitroaniline	110	UD	27.1	110	ug/L
131-11-3	Dimethylphthalate	110	UD	13.1	110	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	DUP-01-081825DL		SDG No.:	Q2900	
Lab Sample ID:	Q2900-03DL		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	930	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143523.D	20	08/19/25 09:29	08/21/25 11:47	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
208-96-8	Acenaphthylene	190	D	16.1	110	ug/L
606-20-2	2,6-Dinitrotoluene	110	UD	19.8	110	ug/L
99-09-2	3-Nitroaniline	110	UD	22.6	110	ug/L
83-32-9	Acenaphthene	110	UD	11.8	110	ug/L
51-28-5	2,4-Dinitrophenol	220	UD	130	220	ug/L
100-02-7	4-Nitrophenol	220	UD	51.2	220	ug/L
132-64-9	Dibenzofuran	110	UD	13.1	110	ug/L
121-14-2	2,4-Dinitrotoluene	110	UD	26.2	110	ug/L
84-66-2	Diethylphthalate	110	UD	14.8	110	ug/L
7005-72-3	4-Chlorophenyl-phenylether	110	UD	14.6	110	ug/L
86-73-7	Fluorene	110	UD	13.5	110	ug/L
100-01-6	4-Nitroaniline	110	UD	32.3	110	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	220	UD	61.9	220	ug/L
86-30-6	n-Nitrosodiphenylamine	110	UD	12.5	110	ug/L
101-55-3	4-Bromophenyl-phenylether	110	UD	8.60	110	ug/L
118-74-1	Hexachlorobenzene	110	UD	11.2	110	ug/L
1912-24-9	Atrazine	110	UD	21.7	110	ug/L
87-86-5	Pentachlorophenol	220	UD	34.0	220	ug/L
85-01-8	Phenanthrene	110	UD	10.8	110	ug/L
120-12-7	Anthracene	110	UD	13.1	110	ug/L
86-74-8	Carbazole	110	UD	15.5	110	ug/L
84-74-2	Di-n-butylphthalate	110	UD	26.2	110	ug/L
206-44-0	Fluoranthene	110	UD	17.6	110	ug/L
129-00-0	Pyrene	110	UD	10.8	110	ug/L
85-68-7	Butylbenzylphthalate	110	UDQ	41.5	110	ug/L
91-94-1	3,3-Dichlorobenzidine	220	UDQ	20.0	220	ug/L
56-55-3	Benzo(a)anthracene	110	UD	9.70	110	ug/L
218-01-9	Chrysene	110	UD	9.50	110	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	110	UD	34.4	110	ug/L
117-84-0	Di-n-octyl phthalate	220	UD	50.3	220	ug/L
205-99-2	Benzo(b)fluoranthene	110	UD	10.5	110	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	DUP-01-081825DL		SDG No.:	Q2900	
Lab Sample ID:	Q2900-03DL		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	930	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143523.D	20	08/19/25 09:29	08/21/25 11:47	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
207-08-9	Benzo(k)fluoranthene	110	UD	10.3	110	ug/L
50-32-8	Benzo(a)pyrene	110	UD	11.8	110	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	110	UD	12.7	110	ug/L
53-70-3	Dibenzo(a,h)anthracene	110	UD	14.4	110	ug/L
191-24-2	Benzo(g,h,i)perylene	110	UD	14.8	110	ug/L
95-94-3	1,2,4,5-Tetrachlorobenzene	110	UD	11.2	110	ug/L
123-91-1	1,4-Dioxane	110	UD	21.5	110	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	110	UD	15.5	110	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	40.7		23 - 138	27%	SPK: 150
13127-88-3	Phenol-d6	28.2		10 - 134	19%	SPK: 150
4165-60-0	Nitrobenzene-d5	68.1		67 - 132	68%	SPK: 100
321-60-8	2-Fluorobiphenyl	76.8		52 - 132	77%	SPK: 100
118-79-6	2,4,6-Tribromophenol	76.5		44 - 137	51%	SPK: 150
1718-51-0	Terphenyl-d14	66.7		42 - 152	67%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	120000	6.928			
1146-65-2	Naphthalene-d8	450000	8.21			
15067-26-2	Acenaphthene-d10	258000	9.963			
1517-22-2	Phenanthrene-d10	382000	11.451			
1719-03-5	Chrysene-d12	254000	14.092			
1520-96-3	Perylene-d12	291000	15.592			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	DUP-01-081825DL2		SDG No.:	Q2900	
Lab Sample ID:	Q2900-03DL2		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	930	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143528.D	100	08/19/25 09:29	08/21/25 14:13	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
100-52-7	Benzaldehyde	1100	UD	420	1100	ug/L
108-95-2	Phenol	540	UD	97.8	540	ug/L
111-44-4	bis(2-Chloroethyl)ether	540	UD	87.1	540	ug/L
95-57-8	2-Chlorophenol	540	UD	62.4	540	ug/L
95-48-7	2-Methylphenol	540	UD	120	540	ug/L
108-60-1	2,2-oxybis(1-Chloropropane)	540	UD	140	540	ug/L
98-86-2	Acetophenone	540	UD	79.6	540	ug/L
65794-96-9	3+4-Methylphenols	1100	UD	120	1100	ug/L
621-64-7	n-Nitroso-di-n-propylamine	270	UD	150	270	ug/L
67-72-1	Hexachloroethane	540	UD	69.9	540	ug/L
98-95-3	Nitrobenzene	540	UD	81.7	540	ug/L
78-59-1	Isophorone	540	UD	80.6	540	ug/L
88-75-5	2-Nitrophenol	540	UD	190	540	ug/L
105-67-9	2,4-Dimethylphenol	540	UD	200	540	ug/L
111-91-1	bis(2-Chloroethoxy)methane	540	UD	73.1	540	ug/L
120-83-2	2,4-Dichlorophenol	540	UD	55.9	540	ug/L
91-20-3	Naphthalene	8300	D	53.8	540	ug/L
106-47-8	4-Chloroaniline	540	UD	90.3	540	ug/L
87-68-3	Hexachlorobutadiene	540	UD	58.1	540	ug/L
105-60-2	Caprolactam	1100	UD	120	1100	ug/L
59-50-7	4-Chloro-3-methylphenol	540	UD	63.4	540	ug/L
91-57-6	2-Methylnaphthalene	760	D	60.2	540	ug/L
77-47-4	Hexachlorocyclopentadiene	1100	UD	390	1100	ug/L
88-06-2	2,4,6-Trichlorophenol	540	UD	54.8	540	ug/L
95-95-4	2,4,5-Trichlorophenol	540	UD	66.7	540	ug/L
92-52-4	1,1-Biphenyl	540	UD	57.0	540	ug/L
91-58-7	2-Chloronaphthalene	540	UD	65.6	540	ug/L
88-74-4	2-Nitroaniline	540	UD	140	540	ug/L
131-11-3	Dimethylphthalate	540	UD	65.6	540	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	DUP-01-081825DL2		SDG No.:	Q2900	
Lab Sample ID:	Q2900-03DL2		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	930	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143528.D	100	08/19/25 09:29	08/21/25 14:13	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
208-96-8	Acenaphthylene	270	JD	80.6	540	ug/L
606-20-2	2,6-Dinitrotoluene	540	UD	98.9	540	ug/L
99-09-2	3-Nitroaniline	540	UD	110	540	ug/L
83-32-9	Acenaphthene	540	UD	59.1	540	ug/L
51-28-5	2,4-Dinitrophenol	1100	UD	640	1100	ug/L
100-02-7	4-Nitrophenol	1100	UD	260	1100	ug/L
132-64-9	Dibenzofuran	540	UD	65.6	540	ug/L
121-14-2	2,4-Dinitrotoluene	540	UD	130	540	ug/L
84-66-2	Diethylphthalate	540	UD	74.2	540	ug/L
7005-72-3	4-Chlorophenyl-phenylether	540	UD	73.1	540	ug/L
86-73-7	Fluorene	540	UD	67.7	540	ug/L
100-01-6	4-Nitroaniline	540	UD	160	540	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	1100	UD	310	1100	ug/L
86-30-6	n-Nitrosodiphenylamine	540	UD	62.4	540	ug/L
101-55-3	4-Bromophenyl-phenylether	540	UD	43.0	540	ug/L
118-74-1	Hexachlorobenzene	540	UD	55.9	540	ug/L
1912-24-9	Atrazine	540	UD	110	540	ug/L
87-86-5	Pentachlorophenol	1100	UD	170	1100	ug/L
85-01-8	Phenanthrene	540	UD	53.8	540	ug/L
120-12-7	Anthracene	540	UD	65.6	540	ug/L
86-74-8	Carbazole	540	UD	77.4	540	ug/L
84-74-2	Di-n-butylphthalate	540	UD	130	540	ug/L
206-44-0	Fluoranthene	540	UD	88.2	540	ug/L
129-00-0	Pyrene	540	UD	53.8	540	ug/L
85-68-7	Butylbenzylphthalate	540	UDQ	210	540	ug/L
91-94-1	3,3-Dichlorobenzidine	1100	UDQ	100	1100	ug/L
56-55-3	Benzo(a)anthracene	540	UD	48.4	540	ug/L
218-01-9	Chrysene	540	UD	47.3	540	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	540	UD	170	540	ug/L
117-84-0	Di-n-octyl phthalate	1100	UD	250	1100	ug/L
205-99-2	Benzo(b)fluoranthene	540	UD	52.7	540	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	DUP-01-081825DL2		SDG No.:	Q2900	
Lab Sample ID:	Q2900-03DL2		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	930	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143528.D	100	08/19/25 09:29	08/21/25 14:13	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
207-08-9	Benzo(k)fluoranthene	540	UD	51.6	540	ug/L
50-32-8	Benzo(a)pyrene	540	UD	59.1	540	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	540	UD	63.4	540	ug/L
53-70-3	Dibenzo(a,h)anthracene	540	UD	72.0	540	ug/L
191-24-2	Benzo(g,h,i)perylene	540	UD	74.2	540	ug/L
95-94-3	1,2,4,5-Tetrachlorobenzene	540	UD	55.9	540	ug/L
123-91-1	1,4-Dioxane	540	UD	110	540	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	540	UD	77.4	540	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	37.8		23 - 138	25%	SPK: 150
13127-88-3	Phenol-d6	31.3		10 - 134	21%	SPK: 150
4165-60-0	Nitrobenzene-d5	82.9		67 - 132	83%	SPK: 100
321-60-8	2-Fluorobiphenyl	115		52 - 132	115%	SPK: 100
118-79-6	2,4,6-Tribromophenol	60.6	*	44 - 137	40%	SPK: 150
1718-51-0	Terphenyl-d14	89.0		42 - 152	89%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	106000	6.928			
1146-65-2	Naphthalene-d8	407000	8.21			
15067-26-2	Acenaphthene-d10	216000	9.963			
1517-22-2	Phenanthrene-d10	311000	11.451			
1719-03-5	Chrysene-d12	211000	14.092			
1520-96-3	Perylene-d12	252000	15.592			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MW-17B-081825		SDG No.:	Q2900	
Lab Sample ID:	Q2900-04		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	940	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143496.D	1	08/19/25 09:29	08/20/25 20:37	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
100-52-7	Benzaldehyde	10.6	U	4.20	10.6	ug/L
108-95-2	Phenol	5.30	U	0.97	5.30	ug/L
111-44-4	bis(2-Chloroethyl)ether	5.30	U	0.86	5.30	ug/L
95-57-8	2-Chlorophenol	5.30	U	0.62	5.30	ug/L
95-48-7	2-Methylphenol	5.30	U	1.20	5.30	ug/L
108-60-1	2,2-oxybis(1-Chloropropane)	5.30	U	1.40	5.30	ug/L
98-86-2	Acetophenone	5.30	U	0.79	5.30	ug/L
65794-96-9	3+4-Methylphenols	10.6	U	1.20	10.6	ug/L
621-64-7	n-Nitroso-di-n-propylamine	2.70	U	1.50	2.70	ug/L
67-72-1	Hexachloroethane	5.30	U	0.69	5.30	ug/L
98-95-3	Nitrobenzene	5.30	U	0.81	5.30	ug/L
78-59-1	Isophorone	5.30	U	0.80	5.30	ug/L
88-75-5	2-Nitrophenol	5.30	U	1.90	5.30	ug/L
105-67-9	2,4-Dimethylphenol	5.30	U	2.00	5.30	ug/L
111-91-1	bis(2-Chloroethoxy)methane	5.30	U	0.72	5.30	ug/L
120-83-2	2,4-Dichlorophenol	5.30	U	0.55	5.30	ug/L
91-20-3	Naphthalene	5.30	U	0.53	5.30	ug/L
106-47-8	4-Chloroaniline	5.30	U	0.89	5.30	ug/L
87-68-3	Hexachlorobutadiene	5.30	U	0.57	5.30	ug/L
105-60-2	Caprolactam	10.6	U	1.20	10.6	ug/L
59-50-7	4-Chloro-3-methylphenol	5.30	U	0.63	5.30	ug/L
91-57-6	2-Methylnaphthalene	5.30	U	0.60	5.30	ug/L
77-47-4	Hexachlorocyclopentadiene	10.6	U	3.90	10.6	ug/L
88-06-2	2,4,6-Trichlorophenol	5.30	U	0.54	5.30	ug/L
95-95-4	2,4,5-Trichlorophenol	5.30	U	0.66	5.30	ug/L
92-52-4	1,1-Biphenyl	5.30	U	0.56	5.30	ug/L
91-58-7	2-Chloronaphthalene	5.30	U	0.65	5.30	ug/L
88-74-4	2-Nitroaniline	5.30	U	1.30	5.30	ug/L
131-11-3	Dimethylphthalate	5.30	U	0.65	5.30	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MW-17B-081825		SDG No.:	Q2900	
Lab Sample ID:	Q2900-04		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	940	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143496.D	1	08/19/25 09:29	08/20/25 20:37	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
208-96-8	Acenaphthylene	5.30	U	0.80	5.30	ug/L
606-20-2	2,6-Dinitrotoluene	5.30	U	0.98	5.30	ug/L
99-09-2	3-Nitroaniline	5.30	U	1.10	5.30	ug/L
83-32-9	Acenaphthene	5.30	U	0.59	5.30	ug/L
51-28-5	2,4-Dinitrophenol	10.6	U	6.40	10.6	ug/L
100-02-7	4-Nitrophenol	10.6	U	2.50	10.6	ug/L
132-64-9	Dibenzofuran	5.30	U	0.65	5.30	ug/L
121-14-2	2,4-Dinitrotoluene	5.30	U	1.30	5.30	ug/L
84-66-2	Diethylphthalate	5.30	U	0.73	5.30	ug/L
7005-72-3	4-Chlorophenyl-phenylether	5.30	U	0.72	5.30	ug/L
86-73-7	Fluorene	5.30	U	0.67	5.30	ug/L
100-01-6	4-Nitroaniline	5.30	U	1.60	5.30	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	10.6	U	3.10	10.6	ug/L
86-30-6	n-Nitrosodiphenylamine	5.30	U	0.62	5.30	ug/L
101-55-3	4-Bromophenyl-phenylether	5.30	U	0.43	5.30	ug/L
118-74-1	Hexachlorobenzene	5.30	U	0.55	5.30	ug/L
1912-24-9	Atrazine	5.30	U	1.10	5.30	ug/L
87-86-5	Pentachlorophenol	10.6	U	1.70	10.6	ug/L
85-01-8	Phenanthrene	5.30	U	0.53	5.30	ug/L
120-12-7	Anthracene	5.30	U	0.65	5.30	ug/L
86-74-8	Carbazole	5.30	U	0.77	5.30	ug/L
84-74-2	Di-n-butylphthalate	5.30	U	1.30	5.30	ug/L
206-44-0	Fluoranthene	5.30	U	0.87	5.30	ug/L
129-00-0	Pyrene	5.30	U	0.53	5.30	ug/L
85-68-7	Butylbenzylphthalate	5.30	UQ	2.10	5.30	ug/L
91-94-1	3,3-Dichlorobenzidine	10.6	UQ	0.99	10.6	ug/L
56-55-3	Benzo(a)anthracene	5.30	U	0.48	5.30	ug/L
218-01-9	Chrysene	5.30	U	0.47	5.30	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	5.30	U	1.70	5.30	ug/L
117-84-0	Di-n-octyl phthalate	10.6	U	2.50	10.6	ug/L
205-99-2	Benzo(b)fluoranthene	5.30	U	0.52	5.30	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MW-17B-081825		SDG No.:	Q2900	
Lab Sample ID:	Q2900-04		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	940	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143496.D	1	08/19/25 09:29	08/20/25 20:37	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
207-08-9	Benzo(k)fluoranthene	5.30	U	0.51	5.30	ug/L
50-32-8	Benzo(a)pyrene	5.30	U	0.59	5.30	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	5.30	U	0.63	5.30	ug/L
53-70-3	Dibenzo(a,h)anthracene	5.30	U	0.71	5.30	ug/L
191-24-2	Benzo(g,h,i)perylene	5.30	U	0.73	5.30	ug/L
95-94-3	1,2,4,5-Tetrachlorobenzene	5.30	U	0.55	5.30	ug/L
123-91-1	1,4-Dioxane	5.30	U	1.10	5.30	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	5.30	U	0.77	5.30	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	63.2		23 - 138	42%	SPK: 150
13127-88-3	Phenol-d6	39.9		10 - 134	27%	SPK: 150
4165-60-0	Nitrobenzene-d5	93.5		67 - 132	94%	SPK: 100
321-60-8	2-Fluorobiphenyl	84.0		52 - 132	84%	SPK: 100
118-79-6	2,4,6-Tribromophenol	128		44 - 137	86%	SPK: 150
1718-51-0	Terphenyl-d14	91.9		42 - 152	92%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	132000	6.928			
1146-65-2	Naphthalene-d8	508000	8.204			
15067-26-2	Acenaphthene-d10	274000	9.963			
1517-22-2	Phenanthrene-d10	430000	11.451			
1719-03-5	Chrysene-d12	242000	14.092			
1520-96-3	Perylene-d12	249000	15.592			
TENTATIVE IDENTIFIED COMPOUNDS						
000994-05-8	Butane, 2-methoxy-2-methyl-	120	J		2.27	ug/L
000123-42-2	2-Pentanone, 4-hydroxy-4-methyl-	5.20	AB		5.16	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MW-17B-081825		SDG No.:	Q2900	
Lab Sample ID:	Q2900-04		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	940	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143496.D	1	08/19/25 09:29	08/20/25 20:37	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MW-11C-081825		SDG No.:	Q2900	
Lab Sample ID:	Q2900-05		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	950	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143513.D	1	08/19/25 09:29	08/21/25 05:18	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
100-52-7	Benzaldehyde	10.5	U	4.10	10.5	ug/L
108-95-2	Phenol	5.30	U	0.96	5.30	ug/L
111-44-4	bis(2-Chloroethyl)ether	5.30	U	0.85	5.30	ug/L
95-57-8	2-Chlorophenol	5.30	U	0.61	5.30	ug/L
95-48-7	2-Methylphenol	5.30	U	1.20	5.30	ug/L
108-60-1	2,2-oxybis(1-Chloropropane)	5.30	U	1.30	5.30	ug/L
98-86-2	Acetophenone	5.30	U	0.78	5.30	ug/L
65794-96-9	3+4-Methylphenols	10.5	U	1.20	10.5	ug/L
621-64-7	n-Nitroso-di-n-propylamine	2.60	U	1.50	2.60	ug/L
67-72-1	Hexachloroethane	5.30	U	0.68	5.30	ug/L
98-95-3	Nitrobenzene	5.30	U	0.80	5.30	ug/L
78-59-1	Isophorone	5.30	U	0.79	5.30	ug/L
88-75-5	2-Nitrophenol	5.30	U	1.90	5.30	ug/L
105-67-9	2,4-Dimethylphenol	5.30	U	1.90	5.30	ug/L
111-91-1	bis(2-Chloroethoxy)methane	5.30	U	0.72	5.30	ug/L
120-83-2	2,4-Dichlorophenol	5.30	U	0.55	5.30	ug/L
91-20-3	Naphthalene	2200	E	0.53	5.30	ug/L
106-47-8	4-Chloroaniline	5.30	U	0.88	5.30	ug/L
87-68-3	Hexachlorobutadiene	5.30	U	0.57	5.30	ug/L
105-60-2	Caprolactam	10.5	U	1.20	10.5	ug/L
59-50-7	4-Chloro-3-methylphenol	5.30	U	0.62	5.30	ug/L
91-57-6	2-Methylnaphthalene	160	E	0.59	5.30	ug/L
77-47-4	Hexachlorocyclopentadiene	10.5	U	3.80	10.5	ug/L
88-06-2	2,4,6-Trichlorophenol	5.30	U	0.54	5.30	ug/L
95-95-4	2,4,5-Trichlorophenol	5.30	U	0.65	5.30	ug/L
92-52-4	1,1-Biphenyl	24.0		0.56	5.30	ug/L
91-58-7	2-Chloronaphthalene	5.30	U	0.64	5.30	ug/L
88-74-4	2-Nitroaniline	5.30	U	1.30	5.30	ug/L
131-11-3	Dimethylphthalate	5.30	U	0.64	5.30	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MW-11C-081825		SDG No.:	Q2900	
Lab Sample ID:	Q2900-05		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	950	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143513.D	1	08/19/25 09:29	08/21/25 05:18	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
208-96-8	Acenaphthylene	110	E	0.79	5.30	ug/L
606-20-2	2,6-Dinitrotoluene	5.30	U	0.97	5.30	ug/L
99-09-2	3-Nitroaniline	5.30	U	1.10	5.30	ug/L
83-32-9	Acenaphthene	69.2		0.58	5.30	ug/L
51-28-5	2,4-Dinitrophenol	10.5	U	6.30	10.5	ug/L
100-02-7	4-Nitrophenol	10.5	U	2.50	10.5	ug/L
132-64-9	Dibenzofuran	5.00	J	0.64	5.30	ug/L
121-14-2	2,4-Dinitrotoluene	5.30	U	1.30	5.30	ug/L
84-66-2	Diethylphthalate	5.30	U	0.73	5.30	ug/L
7005-72-3	4-Chlorophenyl-phenylether	5.30	U	0.72	5.30	ug/L
86-73-7	Fluorene	17.9		0.66	5.30	ug/L
100-01-6	4-Nitroaniline	5.30	U	1.60	5.30	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	10.5	U	3.00	10.5	ug/L
86-30-6	n-Nitrosodiphenylamine	5.30	U	0.61	5.30	ug/L
101-55-3	4-Bromophenyl-phenylether	5.30	U	0.42	5.30	ug/L
118-74-1	Hexachlorobenzene	5.30	U	0.55	5.30	ug/L
1912-24-9	Atrazine	5.30	U	1.10	5.30	ug/L
87-86-5	Pentachlorophenol	10.5	U	1.70	10.5	ug/L
85-01-8	Phenanthrene	23.6		0.53	5.30	ug/L
120-12-7	Anthracene	3.30	J	0.64	5.30	ug/L
86-74-8	Carbazole	13.1		0.76	5.30	ug/L
84-74-2	Di-n-butylphthalate	5.30	U	1.30	5.30	ug/L
206-44-0	Fluoranthene	5.30	U	0.86	5.30	ug/L
129-00-0	Pyrene	5.30	U	0.53	5.30	ug/L
85-68-7	Butylbenzylphthalate	5.30	UQ	2.00	5.30	ug/L
91-94-1	3,3-Dichlorobenzidine	10.5	UQ	0.98	10.5	ug/L
56-55-3	Benzo(a)anthracene	5.30	U	0.47	5.30	ug/L
218-01-9	Chrysene	5.30	U	0.46	5.30	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	5.30	U	1.70	5.30	ug/L
117-84-0	Di-n-octyl phthalate	10.5	U	2.50	10.5	ug/L
205-99-2	Benzo(b)fluoranthene	5.30	U	0.52	5.30	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MW-11C-081825		SDG No.:	Q2900	
Lab Sample ID:	Q2900-05		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	950	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143513.D	1	08/19/25 09:29	08/21/25 05:18	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
207-08-9	Benzo(k)fluoranthene	5.30	U	0.51	5.30	ug/L
50-32-8	Benzo(a)pyrene	5.30	U	0.58	5.30	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	5.30	U	0.62	5.30	ug/L
53-70-3	Dibenzo(a,h)anthracene	5.30	U	0.71	5.30	ug/L
191-24-2	Benzo(g,h,i)perylene	5.30	U	0.73	5.30	ug/L
95-94-3	1,2,4,5-Tetrachlorobenzene	5.30	U	0.55	5.30	ug/L
123-91-1	1,4-Dioxane	5.30	U	1.10	5.30	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	5.30	U	0.76	5.30	ug/L

SURROGATES

367-12-4	2-Fluorophenol	48.1		23 - 138	32%	SPK: 150
13127-88-3	Phenol-d6	37.8		10 - 134	25%	SPK: 150
4165-60-0	Nitrobenzene-d5	200	*	67 - 132	200%	SPK: 100
321-60-8	2-Fluorobiphenyl	84.4		52 - 132	84%	SPK: 100
118-79-6	2,4,6-Tribromophenol	135		44 - 137	90%	SPK: 150
1718-51-0	Terphenyl-d14	83.5		42 - 152	83%	SPK: 100

INTERNAL STANDARDS

3855-82-1	1,4-Dichlorobenzene-d4	125000	6.934
1146-65-2	Naphthalene-d8	212000	8.228
15067-26-2	Acenaphthene-d10	252000	9.963
1517-22-2	Phenanthrene-d10	380000	11.451
1719-03-5	Chrysene-d12	223000	14.092
1520-96-3	Perylene-d12	274000	15.592

TENTATIVE IDENTIFIED COMPOUNDS

000994-05-8	Butane, 2-methoxy-2-methyl-	130	J	2.26	ug/L
000544-25-2	1,3,5-Cycloheptatriene	400	J	4.08	ug/L
002175-91-9	1,3-Cyclopentadiene, 5-(1-methylet	360	J	5.45	ug/L
000108-38-3	Benzene, 1,3-dimethyl-	480	J	5.80	ug/L
000098-82-8	Benzene, (1-methylethyl)-	39.6	J	6.10	ug/L
000611-14-3	Benzene, 1-ethyl-2-methyl-	29.6	J	6.46	ug/L
000622-96-8	Benzene, 1-ethyl-4-methyl-	95.7	J	6.49	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MW-11C-081825		SDG No.:	Q2900	
Lab Sample ID:	Q2900-05		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	950	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143513.D	1	08/19/25 09:29	08/21/25 05:18	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
1000191-13-7	Tetracyclo[3.3.1.0(2,8).0(4,6)]-no	37.9	J		6.80	ug/L
000526-73-8	Benzene, 1,2,3-trimethyl-	130	J		6.99	ug/L
000300-57-2	Benzene, 2-propenyl-	25.0	J		7.03	ug/L
000496-11-7	Indane	230	J		7.12	ug/L
000095-13-6	Indene	890	J		7.22	ug/L
000767-59-9	1H-Indene, 1-methyl-	4.20	J		7.97	ug/L
022433-39-2	Benzene, 1-methyl-1,2-propadienyl-	3.50	J		8.01	ug/L
90-12-0	1-Methylnaphthalene	570	J		9.04	ug/L
001127-76-0	Naphthalene, 1-ethyl-	12.5	J		9.48	ug/L
000575-43-9	Naphthalene, 1,6-dimethyl-	25.9	J		9.55	ug/L
000581-40-8	Naphthalene, 2,3-dimethyl-	36.4	J		9.62	ug/L
000575-41-7	Naphthalene, 1,3-dimethyl-	17.9	J		9.74	ug/L

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MW-11C-081825DL		SDG No.:	Q2900	
Lab Sample ID:	Q2900-05DL		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	950	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143524.D	20	08/19/25 09:29	08/21/25 12:17	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
100-52-7	Benzaldehyde	210	UD	82.3	210	ug/L
108-95-2	Phenol	110	UD	19.2	110	ug/L
111-44-4	bis(2-Chloroethyl)ether	110	UD	17.1	110	ug/L
95-57-8	2-Chlorophenol	110	UD	12.2	110	ug/L
95-48-7	2-Methylphenol	110	UD	23.6	110	ug/L
108-60-1	2,2-oxybis(1-Chloropropane)	110	UD	26.9	110	ug/L
98-86-2	Acetophenone	110	UD	15.6	110	ug/L
65794-96-9	3+4-Methylphenols	210	UD	23.2	210	ug/L
621-64-7	n-Nitroso-di-n-propylamine	52.6	UD	29.7	52.6	ug/L
67-72-1	Hexachloroethane	110	UD	13.7	110	ug/L
98-95-3	Nitrobenzene	110	UD	16.0	110	ug/L
78-59-1	Isophorone	110	UD	15.8	110	ug/L
88-75-5	2-Nitrophenol	110	UD	37.1	110	ug/L
105-67-9	2,4-Dimethylphenol	110	UD	38.9	110	ug/L
111-91-1	bis(2-Chloroethoxy)methane	110	UD	14.3	110	ug/L
120-83-2	2,4-Dichlorophenol	110	UD	10.9	110	ug/L
91-20-3	Naphthalene	2800	ED	10.5	110	ug/L
106-47-8	4-Chloroaniline	110	UD	17.7	110	ug/L
87-68-3	Hexachlorobutadiene	110	UD	11.4	110	ug/L
105-60-2	Caprolactam	210	UD	23.8	210	ug/L
59-50-7	4-Chloro-3-methylphenol	110	UD	12.4	110	ug/L
91-57-6	2-Methylnaphthalene	76.6	JD	11.8	110	ug/L
77-47-4	Hexachlorocyclopentadiene	210	UD	76.4	210	ug/L
88-06-2	2,4,6-Trichlorophenol	110	UD	10.7	110	ug/L
95-95-4	2,4,5-Trichlorophenol	110	UD	13.1	110	ug/L
92-52-4	1,1-Biphenyl	110	UD	11.2	110	ug/L
91-58-7	2-Chloronaphthalene	110	UD	12.8	110	ug/L
88-74-4	2-Nitroaniline	110	UD	26.5	110	ug/L
131-11-3	Dimethylphthalate	110	UD	12.8	110	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MW-11C-081825DL		SDG No.:	Q2900	
Lab Sample ID:	Q2900-05DL		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	950	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143524.D	20	08/19/25 09:29	08/21/25 12:17	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
208-96-8	Acenaphthylene	120	D	15.8	110	ug/L
606-20-2	2,6-Dinitrotoluene	110	UD	19.4	110	ug/L
99-09-2	3-Nitroaniline	110	UD	22.1	110	ug/L
83-32-9	Acenaphthene	72.4	JD	11.6	110	ug/L
51-28-5	2,4-Dinitrophenol	210	UD	130	210	ug/L
100-02-7	4-Nitrophenol	210	UD	50.1	210	ug/L
132-64-9	Dibenzofuran	110	UD	12.8	110	ug/L
121-14-2	2,4-Dinitrotoluene	110	UD	25.7	110	ug/L
84-66-2	Diethylphthalate	110	UD	14.5	110	ug/L
7005-72-3	4-Chlorophenyl-phenylether	110	UD	14.3	110	ug/L
86-73-7	Fluorene	110	UD	13.3	110	ug/L
100-01-6	4-Nitroaniline	110	UD	31.6	110	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	210	UD	60.6	210	ug/L
86-30-6	n-Nitrosodiphenylamine	110	UD	12.2	110	ug/L
101-55-3	4-Bromophenyl-phenylether	110	UD	8.40	110	ug/L
118-74-1	Hexachlorobenzene	110	UD	10.9	110	ug/L
1912-24-9	Atrazine	110	UD	21.3	110	ug/L
87-86-5	Pentachlorophenol	210	UD	33.3	210	ug/L
85-01-8	Phenanthrene	110	UD	10.5	110	ug/L
120-12-7	Anthracene	110	UD	12.8	110	ug/L
86-74-8	Carbazole	110	UD	15.2	110	ug/L
84-74-2	Di-n-butylphthalate	110	UD	25.7	110	ug/L
206-44-0	Fluoranthene	110	UD	17.3	110	ug/L
129-00-0	Pyrene	110	UD	10.5	110	ug/L
85-68-7	Butylbenzylphthalate	110	UDQ	40.6	110	ug/L
91-94-1	3,3-Dichlorobenzidine	210	UDQ	19.6	210	ug/L
56-55-3	Benzo(a)anthracene	110	UD	9.50	110	ug/L
218-01-9	Chrysene	110	UD	9.30	110	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	110	UD	33.7	110	ug/L
117-84-0	Di-n-octyl phthalate	210	UD	49.3	210	ug/L
205-99-2	Benzo(b)fluoranthene	110	UD	10.3	110	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MW-11C-081825DL		SDG No.:	Q2900	
Lab Sample ID:	Q2900-05DL		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	950	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143524.D	20	08/19/25 09:29	08/21/25 12:17	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
207-08-9	Benzo(k)fluoranthene	110	UD	10.1	110	ug/L
50-32-8	Benzo(a)pyrene	110	UD	11.6	110	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	110	UD	12.4	110	ug/L
53-70-3	Dibenzo(a,h)anthracene	110	UD	14.1	110	ug/L
191-24-2	Benzo(g,h,i)perylene	110	UD	14.5	110	ug/L
95-94-3	1,2,4,5-Tetrachlorobenzene	110	UD	10.9	110	ug/L
123-91-1	1,4-Dioxane	110	UD	21.1	110	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	110	UD	15.2	110	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	39.7		23 - 138	26%	SPK: 150
13127-88-3	Phenol-d6	27.6		10 - 134	18%	SPK: 150
4165-60-0	Nitrobenzene-d5	73.7		67 - 132	74%	SPK: 100
321-60-8	2-Fluorobiphenyl	88.6		52 - 132	89%	SPK: 100
118-79-6	2,4,6-Tribromophenol	77.6		44 - 137	52%	SPK: 150
1718-51-0	Terphenyl-d14	67.2		42 - 152	67%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	113000	6.922			
1146-65-2	Naphthalene-d8	431000	8.21			
15067-26-2	Acenaphthene-d10	238000	9.963			
1517-22-2	Phenanthrene-d10	357000	11.451			
1719-03-5	Chrysene-d12	226000	14.092			
1520-96-3	Perylene-d12	263000	15.592			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MW-11C-081825DL2		SDG No.:	Q2900	
Lab Sample ID:	Q2900-05DL2		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	950	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143529.D	100	08/19/25 09:29	08/21/25 14:42	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
100-52-7	Benzaldehyde	1100	UD	410	1100	ug/L
108-95-2	Phenol	530	UD	95.8	530	ug/L
111-44-4	bis(2-Chloroethyl)ether	530	UD	85.3	530	ug/L
95-57-8	2-Chlorophenol	530	UD	61.1	530	ug/L
95-48-7	2-Methylphenol	530	UD	120	530	ug/L
108-60-1	2,2-oxybis(1-Chloropropane)	530	UD	130	530	ug/L
98-86-2	Acetophenone	530	UD	77.9	530	ug/L
65794-96-9	3+4-Methylphenols	1100	UD	120	1100	ug/L
621-64-7	n-Nitroso-di-n-propylamine	260	UD	150	260	ug/L
67-72-1	Hexachloroethane	530	UD	68.4	530	ug/L
98-95-3	Nitrobenzene	530	UD	80.0	530	ug/L
78-59-1	Isophorone	530	UD	78.9	530	ug/L
88-75-5	2-Nitrophenol	530	UD	190	530	ug/L
105-67-9	2,4-Dimethylphenol	530	UD	190	530	ug/L
111-91-1	bis(2-Chloroethoxy)methane	530	UD	71.6	530	ug/L
120-83-2	2,4-Dichlorophenol	530	UD	54.7	530	ug/L
91-20-3	Naphthalene	5300	D	52.6	530	ug/L
106-47-8	4-Chloroaniline	530	UD	88.4	530	ug/L
87-68-3	Hexachlorobutadiene	530	UD	56.8	530	ug/L
105-60-2	Caprolactam	1100	UD	120	1100	ug/L
59-50-7	4-Chloro-3-methylphenol	530	UD	62.1	530	ug/L
91-57-6	2-Methylnaphthalene	530	UD	58.9	530	ug/L
77-47-4	Hexachlorocyclopentadiene	1100	UD	380	1100	ug/L
88-06-2	2,4,6-Trichlorophenol	530	UD	53.7	530	ug/L
95-95-4	2,4,5-Trichlorophenol	530	UD	65.3	530	ug/L
92-52-4	1,1-Biphenyl	530	UD	55.8	530	ug/L
91-58-7	2-Chloronaphthalene	530	UD	64.2	530	ug/L
88-74-4	2-Nitroaniline	530	UD	130	530	ug/L
131-11-3	Dimethylphthalate	530	UD	64.2	530	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MW-11C-081825DL2		SDG No.:	Q2900	
Lab Sample ID:	Q2900-05DL2		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	950	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143529.D	100	08/19/25 09:29	08/21/25 14:42	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
208-96-8	Acenaphthylene	530	UD	78.9	530	ug/L
606-20-2	2,6-Dinitrotoluene	530	UD	96.8	530	ug/L
99-09-2	3-Nitroaniline	530	UD	110	530	ug/L
83-32-9	Acenaphthene	530	UD	57.9	530	ug/L
51-28-5	2,4-Dinitrophenol	1100	UD	630	1100	ug/L
100-02-7	4-Nitrophenol	1100	UD	250	1100	ug/L
132-64-9	Dibenzofuran	530	UD	64.2	530	ug/L
121-14-2	2,4-Dinitrotoluene	530	UD	130	530	ug/L
84-66-2	Diethylphthalate	530	UD	72.6	530	ug/L
7005-72-3	4-Chlorophenyl-phenylether	530	UD	71.6	530	ug/L
86-73-7	Fluorene	530	UD	66.3	530	ug/L
100-01-6	4-Nitroaniline	530	UD	160	530	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	1100	UD	300	1100	ug/L
86-30-6	n-Nitrosodiphenylamine	530	UD	61.1	530	ug/L
101-55-3	4-Bromophenyl-phenylether	530	UD	42.1	530	ug/L
118-74-1	Hexachlorobenzene	530	UD	54.7	530	ug/L
1912-24-9	Atrazine	530	UD	110	530	ug/L
87-86-5	Pentachlorophenol	1100	UD	170	1100	ug/L
85-01-8	Phenanthrene	530	UD	52.6	530	ug/L
120-12-7	Anthracene	530	UD	64.2	530	ug/L
86-74-8	Carbazole	530	UD	75.8	530	ug/L
84-74-2	Di-n-butylphthalate	530	UD	130	530	ug/L
206-44-0	Fluoranthene	530	UD	86.3	530	ug/L
129-00-0	Pyrene	530	UD	52.6	530	ug/L
85-68-7	Butylbenzylphthalate	530	UDQ	200	530	ug/L
91-94-1	3,3-Dichlorobenzidine	1100	UDQ	97.9	1100	ug/L
56-55-3	Benzo(a)anthracene	530	UD	47.4	530	ug/L
218-01-9	Chrysene	530	UD	46.3	530	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	530	UD	170	530	ug/L
117-84-0	Di-n-octyl phthalate	1100	UD	250	1100	ug/L
205-99-2	Benzo(b)fluoranthene	530	UD	51.6	530	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MW-11C-081825DL2		SDG No.:	Q2900	
Lab Sample ID:	Q2900-05DL2		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	950	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143529.D	100	08/19/25 09:29	08/21/25 14:42	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
207-08-9	Benzo(k)fluoranthene	530	UD	50.5	530	ug/L
50-32-8	Benzo(a)pyrene	530	UD	57.9	530	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	530	UD	62.1	530	ug/L
53-70-3	Dibenzo(a,h)anthracene	530	UD	70.5	530	ug/L
191-24-2	Benzo(g,h,i)perylene	530	UD	72.6	530	ug/L
95-94-3	1,2,4,5-Tetrachlorobenzene	530	UD	54.7	530	ug/L
123-91-1	1,4-Dioxane	530	UD	110	530	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	530	UD	75.8	530	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	54.0		23 - 138	36%	SPK: 150
13127-88-3	Phenol-d6	27.1		10 - 134	18%	SPK: 150
4165-60-0	Nitrobenzene-d5	90.5		67 - 132	90%	SPK: 100
321-60-8	2-Fluorobiphenyl	134	*	52 - 132	134%	SPK: 100
118-79-6	2,4,6-Tribromophenol	57.5	*	44 - 137	38%	SPK: 150
1718-51-0	Terphenyl-d14	88.8		42 - 152	89%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	101000	6.928			
1146-65-2	Naphthalene-d8	391000	8.204			
15067-26-2	Acenaphthene-d10	202000	9.963			
1517-22-2	Phenanthrene-d10	279000	11.451			
1719-03-5	Chrysene-d12	184000	14.092			
1520-96-3	Perylene-d12	231000	15.592			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MW-17C-081825		SDG No.:	Q2900	
Lab Sample ID:	Q2900-06		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	910	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143494.D	1	08/19/25 09:29	08/20/25 19:39	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
100-52-7	Benzaldehyde	11.0	U	4.30	11.0	ug/L
108-95-2	Phenol	5.50	U	1.00	5.50	ug/L
111-44-4	bis(2-Chloroethyl)ether	5.50	U	0.89	5.50	ug/L
95-57-8	2-Chlorophenol	5.50	U	0.64	5.50	ug/L
95-48-7	2-Methylphenol	5.50	U	1.20	5.50	ug/L
108-60-1	2,2-oxybis(1-Chloropropane)	5.50	U	1.40	5.50	ug/L
98-86-2	Acetophenone	5.50	U	0.81	5.50	ug/L
65794-96-9	3+4-Methylphenols	11.0	U	1.20	11.0	ug/L
621-64-7	n-Nitroso-di-n-propylamine	2.70	U	1.50	2.70	ug/L
67-72-1	Hexachloroethane	5.50	U	0.71	5.50	ug/L
98-95-3	Nitrobenzene	5.50	U	0.84	5.50	ug/L
78-59-1	Isophorone	5.50	U	0.82	5.50	ug/L
88-75-5	2-Nitrophenol	5.50	U	1.90	5.50	ug/L
105-67-9	2,4-Dimethylphenol	5.50	U	2.00	5.50	ug/L
111-91-1	bis(2-Chloroethoxy)methane	5.50	U	0.75	5.50	ug/L
120-83-2	2,4-Dichlorophenol	5.50	U	0.57	5.50	ug/L
91-20-3	Naphthalene	5.50	U	0.55	5.50	ug/L
106-47-8	4-Chloroaniline	5.50	U	0.92	5.50	ug/L
87-68-3	Hexachlorobutadiene	5.50	U	0.59	5.50	ug/L
105-60-2	Caprolactam	11.0	U	1.20	11.0	ug/L
59-50-7	4-Chloro-3-methylphenol	5.50	U	0.65	5.50	ug/L
91-57-6	2-Methylnaphthalene	5.50	U	0.62	5.50	ug/L
77-47-4	Hexachlorocyclopentadiene	11.0	U	4.00	11.0	ug/L
88-06-2	2,4,6-Trichlorophenol	5.50	U	0.56	5.50	ug/L
95-95-4	2,4,5-Trichlorophenol	5.50	U	0.68	5.50	ug/L
92-52-4	1,1-Biphenyl	5.50	U	0.58	5.50	ug/L
91-58-7	2-Chloronaphthalene	5.50	U	0.67	5.50	ug/L
88-74-4	2-Nitroaniline	5.50	U	1.40	5.50	ug/L
131-11-3	Dimethylphthalate	5.50	U	0.67	5.50	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MW-17C-081825		SDG No.:	Q2900	
Lab Sample ID:	Q2900-06		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	910	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143494.D	1	08/19/25 09:29	08/20/25 19:39	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
208-96-8	Acenaphthylene	5.50	U	0.82	5.50	ug/L
606-20-2	2,6-Dinitrotoluene	5.50	U	1.00	5.50	ug/L
99-09-2	3-Nitroaniline	5.50	U	1.20	5.50	ug/L
83-32-9	Acenaphthene	5.50	U	0.60	5.50	ug/L
51-28-5	2,4-Dinitrophenol	11.0	U	6.60	11.0	ug/L
100-02-7	4-Nitrophenol	11.0	U	2.60	11.0	ug/L
132-64-9	Dibenzofuran	5.50	U	0.67	5.50	ug/L
121-14-2	2,4-Dinitrotoluene	5.50	U	1.30	5.50	ug/L
84-66-2	Diethylphthalate	5.50	U	0.76	5.50	ug/L
7005-72-3	4-Chlorophenyl-phenylether	5.50	U	0.75	5.50	ug/L
86-73-7	Fluorene	5.50	U	0.69	5.50	ug/L
100-01-6	4-Nitroaniline	5.50	U	1.60	5.50	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	11.0	U	3.20	11.0	ug/L
86-30-6	n-Nitrosodiphenylamine	5.50	U	0.64	5.50	ug/L
101-55-3	4-Bromophenyl-phenylether	5.50	U	0.44	5.50	ug/L
118-74-1	Hexachlorobenzene	5.50	U	0.57	5.50	ug/L
1912-24-9	Atrazine	5.50	U	1.10	5.50	ug/L
87-86-5	Pentachlorophenol	11.0	U	1.70	11.0	ug/L
85-01-8	Phenanthrene	5.50	U	0.55	5.50	ug/L
120-12-7	Anthracene	5.50	U	0.67	5.50	ug/L
86-74-8	Carbazole	5.50	U	0.79	5.50	ug/L
84-74-2	Di-n-butylphthalate	5.50	U	1.30	5.50	ug/L
206-44-0	Fluoranthene	5.50	U	0.90	5.50	ug/L
129-00-0	Pyrene	5.50	U	0.55	5.50	ug/L
85-68-7	Butylbenzylphthalate	5.50	UQ	2.10	5.50	ug/L
91-94-1	3,3-Dichlorobenzidine	11.0	UQ	1.00	11.0	ug/L
56-55-3	Benzo(a)anthracene	5.50	U	0.49	5.50	ug/L
218-01-9	Chrysene	5.50	U	0.48	5.50	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	5.50	U	1.80	5.50	ug/L
117-84-0	Di-n-octyl phthalate	11.0	U	2.60	11.0	ug/L
205-99-2	Benzo(b)fluoranthene	5.50	U	0.54	5.50	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MW-17C-081825		SDG No.:	Q2900	
Lab Sample ID:	Q2900-06		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	910	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143494.D	1	08/19/25 09:29	08/20/25 19:39	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
207-08-9	Benzo(k)fluoranthene	5.50	U	0.53	5.50	ug/L
50-32-8	Benzo(a)pyrene	5.50	U	0.60	5.50	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	5.50	U	0.65	5.50	ug/L
53-70-3	Dibenzo(a,h)anthracene	5.50	U	0.74	5.50	ug/L
191-24-2	Benzo(g,h,i)perylene	5.50	U	0.76	5.50	ug/L
95-94-3	1,2,4,5-Tetrachlorobenzene	5.50	U	0.57	5.50	ug/L
123-91-1	1,4-Dioxane	5.50	U	1.10	5.50	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	5.50	U	0.79	5.50	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	69.8		23 - 138	47%	SPK: 150
13127-88-3	Phenol-d6	51.4		10 - 134	34%	SPK: 150
4165-60-0	Nitrobenzene-d5	91.1		67 - 132	91%	SPK: 100
321-60-8	2-Fluorobiphenyl	80.7		52 - 132	81%	SPK: 100
118-79-6	2,4,6-Tribromophenol	137		44 - 137	91%	SPK: 150
1718-51-0	Terphenyl-d14	80.3		42 - 152	80%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	137000	6.928			
1146-65-2	Naphthalene-d8	522000	8.204			
15067-26-2	Acenaphthene-d10	285000	9.963			
1517-22-2	Phenanthrene-d10	459000	11.451			
1719-03-5	Chrysene-d12	263000	14.098			
1520-96-3	Perylene-d12	271000	15.592			
TENTATIVE IDENTIFIED COMPOUNDS						
000994-05-8	Butane, 2-methoxy-2-methyl-	120	J		2.27	ug/L
000123-42-2	2-Pentanone, 4-hydroxy-4-methyl-	4.90	AB		5.16	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MW-17C-081825		SDG No.:	Q2900	
Lab Sample ID:	Q2900-06		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	910	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143494.D	1	08/19/25 09:29	08/20/25 19:39	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MW-11B-081825		SDG No.:	Q2900	
Lab Sample ID:	Q2900-07		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	960	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143495.D	1	08/19/25 09:29	08/20/25 20:08	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
100-52-7	Benzaldehyde	10.4	U	4.10	10.4	ug/L
108-95-2	Phenol	5.20	U	0.95	5.20	ug/L
111-44-4	bis(2-Chloroethyl)ether	5.20	U	0.84	5.20	ug/L
95-57-8	2-Chlorophenol	5.20	U	0.60	5.20	ug/L
95-48-7	2-Methylphenol	5.20	U	1.20	5.20	ug/L
108-60-1	2,2-oxybis(1-Chloropropane)	5.20	U	1.30	5.20	ug/L
98-86-2	Acetophenone	5.20	U	0.77	5.20	ug/L
65794-96-9	3+4-Methylphenols	10.4	U	1.10	10.4	ug/L
621-64-7	n-Nitroso-di-n-propylamine	2.60	U	1.50	2.60	ug/L
67-72-1	Hexachloroethane	5.20	U	0.68	5.20	ug/L
98-95-3	Nitrobenzene	5.20	U	0.79	5.20	ug/L
78-59-1	Isophorone	5.20	U	0.78	5.20	ug/L
88-75-5	2-Nitrophenol	5.20	U	1.80	5.20	ug/L
105-67-9	2,4-Dimethylphenol	5.20	U	1.90	5.20	ug/L
111-91-1	bis(2-Chloroethoxy)methane	5.20	U	0.71	5.20	ug/L
120-83-2	2,4-Dichlorophenol	5.20	U	0.54	5.20	ug/L
91-20-3	Naphthalene	8.50		0.52	5.20	ug/L
106-47-8	4-Chloroaniline	5.20	U	0.88	5.20	ug/L
87-68-3	Hexachlorobutadiene	5.20	U	0.56	5.20	ug/L
105-60-2	Caprolactam	10.4	U	1.20	10.4	ug/L
59-50-7	4-Chloro-3-methylphenol	5.20	U	0.61	5.20	ug/L
91-57-6	2-Methylnaphthalene	5.20	U	0.58	5.20	ug/L
77-47-4	Hexachlorocyclopentadiene	10.4	U	3.80	10.4	ug/L
88-06-2	2,4,6-Trichlorophenol	5.20	U	0.53	5.20	ug/L
95-95-4	2,4,5-Trichlorophenol	5.20	U	0.65	5.20	ug/L
92-52-4	1,1-Biphenyl	5.20	U	0.55	5.20	ug/L
91-58-7	2-Chloronaphthalene	5.20	U	0.64	5.20	ug/L
88-74-4	2-Nitroaniline	5.20	U	1.30	5.20	ug/L
131-11-3	Dimethylphthalate	5.20	U	0.64	5.20	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MW-11B-081825		SDG No.:	Q2900	
Lab Sample ID:	Q2900-07		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	960	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143495.D	1	08/19/25 09:29	08/20/25 20:08	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
208-96-8	Acenaphthylene	5.20	U	0.78	5.20	ug/L
606-20-2	2,6-Dinitrotoluene	5.20	U	0.96	5.20	ug/L
99-09-2	3-Nitroaniline	5.20	U	1.10	5.20	ug/L
83-32-9	Acenaphthene	5.20	U	0.57	5.20	ug/L
51-28-5	2,4-Dinitrophenol	10.4	U	6.20	10.4	ug/L
100-02-7	4-Nitrophenol	10.4	U	2.50	10.4	ug/L
132-64-9	Dibenzofuran	5.20	U	0.64	5.20	ug/L
121-14-2	2,4-Dinitrotoluene	5.20	U	1.30	5.20	ug/L
84-66-2	Diethylphthalate	5.20	U	0.72	5.20	ug/L
7005-72-3	4-Chlorophenyl-phenylether	5.20	U	0.71	5.20	ug/L
86-73-7	Fluorene	5.20	U	0.66	5.20	ug/L
100-01-6	4-Nitroaniline	5.20	U	1.60	5.20	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	10.4	U	3.00	10.4	ug/L
86-30-6	n-Nitrosodiphenylamine	5.20	U	0.60	5.20	ug/L
101-55-3	4-Bromophenyl-phenylether	5.20	U	0.42	5.20	ug/L
118-74-1	Hexachlorobenzene	5.20	U	0.54	5.20	ug/L
1912-24-9	Atrazine	5.20	U	1.10	5.20	ug/L
87-86-5	Pentachlorophenol	10.4	U	1.60	10.4	ug/L
85-01-8	Phenanthrene	5.20	U	0.52	5.20	ug/L
120-12-7	Anthracene	5.20	U	0.64	5.20	ug/L
86-74-8	Carbazole	5.20	U	0.75	5.20	ug/L
84-74-2	Di-n-butylphthalate	5.20	U	1.30	5.20	ug/L
206-44-0	Fluoranthene	5.20	U	0.85	5.20	ug/L
129-00-0	Pyrene	5.20	U	0.52	5.20	ug/L
85-68-7	Butylbenzylphthalate	5.20	UQ	2.00	5.20	ug/L
91-94-1	3,3-Dichlorobenzidine	10.4	UQ	0.97	10.4	ug/L
56-55-3	Benzo(a)anthracene	5.20	U	0.47	5.20	ug/L
218-01-9	Chrysene	5.20	U	0.46	5.20	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	5.20	U	1.70	5.20	ug/L
117-84-0	Di-n-octyl phthalate	10.4	U	2.40	10.4	ug/L
205-99-2	Benzo(b)fluoranthene	5.20	U	0.51	5.20	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MW-11B-081825		SDG No.:	Q2900	
Lab Sample ID:	Q2900-07		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	960	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143495.D	1	08/19/25 09:29	08/20/25 20:08	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
207-08-9	Benzo(k)fluoranthene	5.20	U	0.50	5.20	ug/L
50-32-8	Benzo(a)pyrene	5.20	U	0.57	5.20	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	5.20	U	0.61	5.20	ug/L
53-70-3	Dibenzo(a,h)anthracene	5.20	U	0.70	5.20	ug/L
191-24-2	Benzo(g,h,i)perylene	5.20	U	0.72	5.20	ug/L
95-94-3	1,2,4,5-Tetrachlorobenzene	5.20	U	0.54	5.20	ug/L
123-91-1	1,4-Dioxane	5.20	U	1.00	5.20	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	5.20	U	0.75	5.20	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	60.0		23 - 138	40%	SPK: 150
13127-88-3	Phenol-d6	40.4		10 - 134	27%	SPK: 150
4165-60-0	Nitrobenzene-d5	85.6		67 - 132	86%	SPK: 100
321-60-8	2-Fluorobiphenyl	80.6		52 - 132	81%	SPK: 100
118-79-6	2,4,6-Tribromophenol	122		44 - 137	82%	SPK: 150
1718-51-0	Terphenyl-d14	80.7		42 - 152	81%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	135000	6.928			
1146-65-2	Naphthalene-d8	510000	8.204			
15067-26-2	Acenaphthene-d10	268000	9.963			
1517-22-2	Phenanthrene-d10	401000	11.451			
1719-03-5	Chrysene-d12	238000	14.098			
1520-96-3	Perylene-d12	292000	15.592			
TENTATIVE IDENTIFIED COMPOUNDS						
000994-05-8	Butane, 2-methoxy-2-methyl-	110	J		2.27	ug/L
000123-42-2	2-Pentanone, 4-hydroxy-4-methyl-	4.20	AB		5.16	ug/L
000098-82-8	Benzene, (1-methylethyl)-	6.20	J		6.10	ug/L
000611-14-3	Benzene, 1-ethyl-2-methyl-	3.60	J		6.62	ug/L
000496-11-7	Indane	210	J		7.12	ug/L
002039-89-6	Benzene, 2-ethenyl-1,4-dimethyl-	7.50	J		7.95	ug/L
90-12-0	1-Methylnaphthalene	4.60	J		9.02	ug/L

Report of Analysis

Client:	AECOM		Date Collected:	08/18/25	
Project:	National Grid Equity - Brooklyn NY		Date Received:	08/18/25	
Client Sample ID:	MW-11B-081825		SDG No.:	Q2900	
Lab Sample ID:	Q2900-07		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	960	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted :	N	Level :	LOW
Injection Volume :		GPC Factor :	1.0	GPC Cleanup :	N PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143495.D	1	08/19/25 09:29	08/20/25 20:08	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
028556-81-2	2,6-Dimethylphenyl isocyanate	3.70	J		10.2	ug/L
000061-70-1	2-Indolinone, 1-methyl-	5.60	J		10.4	ug/L
000119-61-9	Benzophenone	5.30	J		10.7	ug/L

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: Q2900

Client: AECOM

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB169313BL	PB169313BL	2-Fluorophenol	150	120	80		23	138
		Phenol-d6	150	121	81		10	134
		Nitrobenzene-d5	100	85.9	86		67	132
		2-Fluorobiphenyl	100	83.4	83		52	132
		2,4,6-Tribromophenol	150	121	81		44	137
		Terphenyl-d14	100	94.0	94		42	152
PB169313BS	PB169313BS	2-Fluorophenol	150	126	84		23	138
		Phenol-d6	150	127	84		10	134
		Nitrobenzene-d5	100	92.7	93		67	132
		2-Fluorobiphenyl	100	86.6	87		52	132
		2,4,6-Tribromophenol	150	140	93		44	137
		Terphenyl-d14	100	94.8	95		42	152
PB169313BSD	PB169313BSD	2-Fluorophenol	150	118	79		23	138
		Phenol-d6	150	120	80		10	134
		Nitrobenzene-d5	100	87.2	87		67	132
		2-Fluorobiphenyl	100	81.4	81		52	132
		2,4,6-Tribromophenol	150	127	84		44	137
		Terphenyl-d14	100	91.2	91		42	152
Q2900-01	MN-9C-081825	2-Fluorophenol	150	65.0	43		23	138
		Phenol-d6	150	43.0	29		10	134
		Nitrobenzene-d5	100	93.7	94		67	132
		2-Fluorobiphenyl	100	87.7	88		52	132
		2,4,6-Tribromophenol	150	140	93		44	137
		Terphenyl-d14	100	86.7	87		42	152
Q2900-02	MN-12C-081825	2-Fluorophenol	150	33.1	22	*	23	138
		Phenol-d6	150	42.2	28		10	134
		Nitrobenzene-d5	100	246	246	*	67	132
		2-Fluorobiphenyl	100	82.2	82		52	132
		2,4,6-Tribromophenol	150	133	89		44	137
		Terphenyl-d14	100	87.4	87		42	152
Q2900-02DL	MN-12C-081825DL	2-Fluorophenol	150	51.2	34		23	138
		Phenol-d6	150	37.2	25		10	134
		Nitrobenzene-d5	100	84.4	84		67	132
		2-Fluorobiphenyl	100	97.9	98		52	132
		2,4,6-Tribromophenol	150	98.2	65		44	137
		Terphenyl-d14	100	80.9	81		42	152
Q2900-02DL2	MN-12C-081825DL2	2-Fluorophenol	150	46.4	31		23	138
		Phenol-d6	150	42.2	28		10	134
		Nitrobenzene-d5	100	91.9	92		67	132
		2-Fluorobiphenyl	100	133	133	*	52	132
		2,4,6-Tribromophenol	150	74.4	50		44	137
		Terphenyl-d14	100	96.7	97		42	152
Q2900-03	DUP-01-081825	2-Fluorophenol	150	29.6	20	*	23	138
		Phenol-d6	150	39.9	27		10	134
		Nitrobenzene-d5	100	258	258	*	67	132
		2-Fluorobiphenyl	100	79.8	80		52	132
		2,4,6-Tribromophenol	150	132	88		44	137
		Terphenyl-d14	100	82.7	83		42	152
Q2900-03DL	DUP-01-081825DL	2-Fluorophenol	150	40.7	27		23	138
		Phenol-d6	150	28.2	19		10	134
		Nitrobenzene-d5	100	68.1	68		67	132

Surrogate Summary

SW-846

SDG No.: Q2900

Client: AECOM

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
Q2900-03DL	DUP-01-081825DL	2-Fluorobiphenyl	100	76.8	77		52	132
		2,4,6-Tribromophenol	150	76.5	51		44	137
		Terphenyl-d14	100	66.7	67		42	152
Q2900-03DL2	DUP-01-081825DL2	2-Fluorophenol	150	37.8	25		23	138
		Phenol-d6	150	31.3	21		10	134
		Nitrobenzene-d5	100	82.9	83		67	132
		2-Fluorobiphenyl	100	115	115		52	132
		2,4,6-Tribromophenol	150	60.6	40	*	44	137
		Terphenyl-d14	100	89.0	89		42	152
Q2900-04	MW-17B-081825	2-Fluorophenol	150	63.2	42		23	138
		Phenol-d6	150	39.9	27		10	134
		Nitrobenzene-d5	100	93.5	94		67	132
		2-Fluorobiphenyl	100	84.0	84		52	132
		2,4,6-Tribromophenol	150	128	86		44	137
		Terphenyl-d14	100	91.9	92		42	152
Q2900-05	MW-11C-081825	2-Fluorophenol	150	48.1	32		23	138
		Phenol-d6	150	37.8	25		10	134
		Nitrobenzene-d5	100	200	200	*	67	132
		2-Fluorobiphenyl	100	84.4	84		52	132
		2,4,6-Tribromophenol	150	135	90		44	137
		Terphenyl-d14	100	83.5	83		42	152
Q2900-05DL	MW-11C-081825DL	2-Fluorophenol	150	39.7	26		23	138
		Phenol-d6	150	27.6	18		10	134
		Nitrobenzene-d5	100	73.7	74		67	132
		2-Fluorobiphenyl	100	88.6	89		52	132
		2,4,6-Tribromophenol	150	77.6	52		44	137
		Terphenyl-d14	100	67.2	67		42	152
Q2900-05DL2	MW-11C-081825DL2	2-Fluorophenol	150	54.0	36		23	138
		Phenol-d6	150	27.1	18		10	134
		Nitrobenzene-d5	100	90.5	90		67	132
		2-Fluorobiphenyl	100	134	134	*	52	132
		2,4,6-Tribromophenol	150	57.5	38	*	44	137
		Terphenyl-d14	100	88.8	89		42	152
Q2900-06	MW-17C-081825	2-Fluorophenol	150	69.8	47		23	138
		Phenol-d6	150	51.4	34		10	134
		Nitrobenzene-d5	100	91.1	91		67	132
		2-Fluorobiphenyl	100	80.7	81		52	132
		2,4,6-Tribromophenol	150	137	91		44	137
		Terphenyl-d14	100	80.3	80		42	152
Q2900-07	MW-11B-081825	2-Fluorophenol	150	60.0	40		23	138
		Phenol-d6	150	40.4	27		10	134
		Nitrobenzene-d5	100	85.6	86		67	132
		2-Fluorobiphenyl	100	80.6	81		52	132
		2,4,6-Tribromophenol	150	122	82		44	137
		Terphenyl-d14	100	80.7	81		42	152

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: <u>Q2900</u>	Analytical Method: <u>8270E</u>
Client: <u>AECOM</u>	DataFile: <u>BF143488.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB169313BS	Benzaldehyde	50	25.5	ug/L	51				10	162		
	Phenol	50	42.6	ug/L	85				66	118		
	bis(2-Chloroethyl)ether	50	44.6	ug/L	89				62	103		
	2-Chlorophenol	50	45.3	ug/L	91				70	117		
	2-Methylphenol	50	45.7	ug/L	91				69	109		
	2,2-oxybis(1-Chloropropane)	50	44.5	ug/L	89				65	100		
	Acetophenone	50	45.9	ug/L	92				60	104		
	3+4-Methylphenols	50	45.1	ug/L	90				67	106		
	N-Nitroso-di-n-propylamine	50	46.7	ug/L	93				57	107		
	Hexachloroethane	50	45.1	ug/L	90				76	118		
	Nitrobenzene	50	45.0	ug/L	90				58	106		
	Isophorone	50	46.3	ug/L	93				61	102		
	2-Nitrophenol	50	47.1	ug/L	94				70	115		
	2,4-Dimethylphenol	50	48.5	ug/L	97				42	142		
	bis(2-Chloroethoxy)methane	50	45.7	ug/L	91				58	109		
	2,4-Dichlorophenol	50	44.3	ug/L	89				66	115		
	Naphthalene	50	43.9	ug/L	88				64	107		
	4-Chloroaniline	50	29.3	ug/L	59				10	85		
	Hexachlorobutadiene	50	44.8	ug/L	90				69	101		
	Caprolactam	50	51.4	ug/L	103				58	128		
	4-Chloro-3-methylphenol	50	45.4	ug/L	91				65	114		
	2-Methylnaphthalene	50	40.4	ug/L	81				64	107		
	Hexachlorocyclopentadiene	100	120	ug/L	120				36	160		
	2,4,6-Trichlorophenol	50	45.3	ug/L	91				61	110		
	2,4,5-Trichlorophenol	50	44.2	ug/L	88				70	106		
	1,1-Biphenyl	50	44.7	ug/L	89				72	98		
	2-Chloronaphthalene	50	43.8	ug/L	88				59	106		
	2-Nitroaniline	50	46.2	ug/L	92				73	114		
	Dimethylphthalate	50	46.7	ug/L	93				64	103		
	Acenaphthylene	50	49.5	ug/L	99				79	103		
	2,6-Dinitrotoluene	50	50.4	ug/L	101				64	110		
	3-Nitroaniline	50	34.3	ug/L	69				28	100		
	Acenaphthene	50	47.9	ug/L	96				59	113		
	2,4-Dinitrophenol	100	90.6	ug/L	91				36	166		
	4-Nitrophenol	100	99.2	ug/L	99				45	147		
	Dibenzofuran	50	44.9	ug/L	90				65	106		
	2,4-Dinitrotoluene	50	50.9	ug/L	102				60	115		
	Diethylphthalate	50	47.7	ug/L	95				63	105		
	4-Chlorophenyl-phenylether	50	45.2	ug/L	90				61	104		
	Fluorene	50	45.4	ug/L	91				64	107		
	4-Nitroaniline	50	48.5	ug/L	97				55	125		
	4,6-Dinitro-2-methylphenol	50	52.6	ug/L	105				62	132		
	N-Nitrosodiphenylamine	50	47.9	ug/L	96				61	109		
	4-Bromophenyl-phenylether	50	46.0	ug/L	92				73	103		

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: <u>Q2900</u>	Analytical Method: <u>8270E</u>
Client: <u>AECOM</u>	DataFile: <u>BF143488.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB169313BS	Hexachlorobenzene	50	45.7	ug/L	91				73	106	
	Atrazine	50	49.7	ug/L	99				76	120	
	Pentachlorophenol	100	98.0	ug/L	98				47	114	
	Phenanthrene	50	45.1	ug/L	90				62	109	
	Anthracene	50	45.8	ug/L	92				65	110	
	Carbazole	50	46.1	ug/L	92				62	106	
	Di-n-butylphthalate	50	49.8	ug/L	100				64	106	
	Fluoranthene	50	45.7	ug/L	91				64	110	
	Pyrene	50	47.0	ug/L	94				71	103	
	Butylbenzylphthalate	50	50.7	ug/L	101				61	105	
	3,3-Dichlorobenzidine	50	33.0	ug/L	66				43	108	
	Benzo(a)anthracene	50	46.5	ug/L	93				62	107	
	Chrysene	50	48.8	ug/L	98				61	108	
	bis(2-Ethylhexyl)phthalate	50	51.0	ug/L	102				59	110	
	Di-n-octyl phthalate	50	51.1	ug/L	102				52	139	
	Benzo(b)fluoranthene	50	46.9	ug/L	94				77	113	
	Benzo(k)fluoranthene	50	47.6	ug/L	95				77	105	
	Benzo(a)pyrene	50	49.7	ug/L	99				72	131	
	Indeno(1,2,3-cd)pyrene	50	46.9	ug/L	94				72	105	
	Dibenz(a,h)anthracene	50	47.3	ug/L	95				78	115	
	Benzo(g,h,i)perylene	50	48.2	ug/L	96				75	118	
	1,2,4,5-Tetrachlorobenzene	50	43.3	ug/L	87				72	101	
	1,4-Dioxane	50	38.9	ug/L	78				38	125	
	2,3,4,6-Tetrachlorophenol	50	51.8	ug/L	104				63	116	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2900

Analytical Method: 8270E

Client: AECOM

DataFile: BF143489.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		
								Qual	Low	High	RPD
PB169313BSD	Benzaldehyde	50	26.7	ug/L	53	5			10	162	20
	Phenol	50	43.9	ug/L	88	3			66	118	20
	bis(2-Chloroethyl)ether	50	46.1	ug/L	92	3			62	103	20
	2-Chlorophenol	50	46.9	ug/L	94	3			70	117	20
	2-Methylphenol	50	47.2	ug/L	94	3			69	109	20
	2,2-oxybis(1-Chloropropane)	50	46.8	ug/L	94	5			65	100	20
	Acetophenone	50	47.0	ug/L	94	2			60	104	20
	3+4-Methylphenols	50	46.3	ug/L	93	3			67	106	20
	N-Nitroso-di-n-propylamine	50	47.8	ug/L	96	2			57	107	20
	Hexachloroethane	50	46.8	ug/L	94	4			76	118	20
	Nitrobenzene	50	46.8	ug/L	94	4			58	106	20
	Isophorone	50	48.7	ug/L	97	5			61	102	20
	2-Nitrophenol	50	48.9	ug/L	98	4			70	115	20
	2,4-Dimethylphenol	50	49.4	ug/L	99	2			42	142	20
	bis(2-Chloroethoxy)methane	50	46.6	ug/L	93	2			58	109	20
	2,4-Dichlorophenol	50	45.3	ug/L	91	2			66	115	20
	Naphthalene	50	44.8	ug/L	90	2			64	107	20
	4-Chloroaniline	50	12.9	ug/L	26	78		*	10	85	20
	Hexachlorobutadiene	50	46.5	ug/L	93	4			69	101	20
	Caprolactam	50	52.1	ug/L	104	1			58	128	20
	4-Chloro-3-methylphenol	50	46.1	ug/L	92	2			65	114	20
	2-Methylnaphthalene	50	41.2	ug/L	82	2			64	107	20
	Hexachlorocyclopentadiene	100	120	ug/L	120	0			36	160	20
	2,4,6-Trichlorophenol	50	45.7	ug/L	91	1			61	110	20
	2,4,5-Trichlorophenol	50	44.0	ug/L	88	0			70	106	20
	1,1-Biphenyl	50	44.9	ug/L	90	0			72	98	20
	2-Chloronaphthalene	50	44.4	ug/L	89	1			59	106	20
	2-Nitroaniline	50	47.5	ug/L	95	3			73	114	20
	Dimethylphthalate	50	47.0	ug/L	94	1			64	103	20
	Acenaphthylene	50	50.0	ug/L	100	1			79	103	20
	2,6-Dinitrotoluene	50	51.2	ug/L	102	2			64	110	20
	3-Nitroaniline	50	24.3	ug/L	49	34		*	28	100	20
	Acenaphthene	50	49.3	ug/L	99	3			59	113	20
	2,4-Dinitrophenol	100	93.0	ug/L	93	3			36	166	20
	4-Nitrophenol	100	100	ug/L	100	1			45	147	20
	Dibenzofuran	50	44.9	ug/L	90	0			65	106	20
	2,4-Dinitrotoluene	50	52.3	ug/L	105	3			60	115	20
	Diethylphthalate	50	48.4	ug/L	97	1			63	105	20
	4-Chlorophenyl-phenylether	50	45.5	ug/L	91	1			61	104	20
	Fluorene	50	45.8	ug/L	92	1			64	107	20
	4-Nitroaniline	50	48.3	ug/L	97	0			55	125	20
	4,6-Dinitro-2-methylphenol	50	53.3	ug/L	107	1			62	132	20
	N-Nitrosodiphenylamine	50	48.3	ug/L	97	1			61	109	20
	4-Bromophenyl-phenylether	50	46.8	ug/L	94	2			73	103	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2900

Analytical Method: 8270E

Client: AECOM

DataFile: BF143489.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB169313BSD	Hexachlorobenzene	50	47.1	ug/L	94	3			73	106	20	
	Atrazine	50	52.6	ug/L	105	6			76	120	20	
	Pentachlorophenol	100	100	ug/L	100	2			47	114	20	
	Phenanthrene	50	46.8	ug/L	94	4			62	109	20	
	Anthracene	50	47.1	ug/L	94	3			65	110	20	
	Carbazole	50	47.8	ug/L	96	4			62	106	20	
	Di-n-butylphthalate	50	51.4	ug/L	103	3			64	106	20	
	Fluoranthene	50	46.9	ug/L	94	3			64	110	20	
	Pyrene	50	48.7	ug/L	97	4			71	103	20	
	Butylbenzylphthalate	50	53.8	ug/L	108	6	*		61	105	20	
	3,3-Dichlorobenzidine	50	20.5	ug/L	41	47	*	*	43	108	20	
	Benzo(a)anthracene	50	46.9	ug/L	94	1			62	107	20	
	Chrysene	50	49.3	ug/L	99	1			61	108	20	
	bis(2-Ethylhexyl)phthalate	50	50.7	ug/L	101	1			59	110	20	
	Di-n-octyl phthalate	50	51.3	ug/L	103	0			52	139	20	
	Benzo(b)fluoranthene	50	47.0	ug/L	94	0			77	113	20	
	Benzo(k)fluoranthene	50	49.1	ug/L	98	3			77	105	20	
	Benzo(a)pyrene	50	50.8	ug/L	102	2			72	131	20	
	Indeno(1,2,3-cd)pyrene	50	47.4	ug/L	95	1			72	105	20	
	Dibenz(a,h)anthracene	50	48.4	ug/L	97	2			78	115	20	
	Benzo(g,h,i)perylene	50	49.0	ug/L	98	2			75	118	20	
	1,2,4,5-Tetrachlorobenzene	50	43.6	ug/L	87	1			72	101	20	
	1,4-Dioxane	50	39.4	ug/L	79	1			38	125	20	
	2,3,4,6-Tetrachlorophenol	50	52.0	ug/L	104	0			63	116	20	

4B

SEMIVOLATILE METHOD BLANK SUMMARY

Client ID

PB169313BL

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG NO.: Q2900

Lab File ID: BF143499.D

Lab Sample ID: PB169313BL

Instrument ID: BNA_F

Date Extracted: 08/19/2025

Matrix: (soil/water) Water

Date Analyzed: 08/20/2025

Level: (low/med) LOW

Time Analyzed: 22:34

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB169313BS	PB169313BS	BF143488.D	08/20/2025
PB169313BSD	PB169313BSD	BF143489.D	08/20/2025
MN-9C-081825	Q2900-01	BF143503.D	08/21/2025
MN-12C-081825	Q2900-02	BF143511.D	08/21/2025
DUP-01-081825	Q2900-03	BF143512.D	08/21/2025
MW-17C-081825	Q2900-06	BF143494.D	08/20/2025
MW-11B-081825	Q2900-07	BF143495.D	08/20/2025
MW-17B-081825	Q2900-04	BF143496.D	08/20/2025
MW-11C-081825	Q2900-05	BF143513.D	08/21/2025

COMMENTS: _____

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BF143476.D
Instrument ID: BNA_F

Contract: AECO02
SDG NO.: Q2900
DFTPP Injection Date: 08/20/2025
DFTPP Injection Time: 10:26

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.5 (1.9) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.1 (0.5) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
365	Greater than 1% of mass 198	3.8
441	Present, but less than mass 443	79.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.3 (19.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF143477.D	08/20/2025	10:55
SSTDICC005	SSTDICC005	BF143478.D	08/20/2025	11:25
SSTDICC010	SSTDICC010	BF143479.D	08/20/2025	11:54
SSTDICC020	SSTDICC020	BF143480.D	08/20/2025	12:23
SSTDICCC040	SSTDICCC040	BF143481.D	08/20/2025	12:53
SSTDICC050	SSTDICC050	BF143482.D	08/20/2025	13:22
SSTDICC060	SSTDICC060	BF143483.D	08/20/2025	13:51
SSTDICC080	SSTDICC080	BF143484.D	08/20/2025	14:20
PB169313BS	PB169313BS	BF143488.D	08/20/2025	16:44
PB169313BSD	PB169313BSD	BF143489.D	08/20/2025	17:13
MW-17C-081825	Q2900-06	BF143494.D	08/20/2025	19:39
MW-11B-081825	Q2900-07	BF143495.D	08/20/2025	20:08
MW-17B-081825	Q2900-04	BF143496.D	08/20/2025	20:37

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BF143497.D
Instrument ID: BNA_F

Contract: AECO02
SDG NO.: Q2900
DFTPP Injection Date: 08/20/2025
DFTPP Injection Time: 21:36

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.6 (2) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.6) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
365	Greater than 1% of mass 198	3.6
441	Present, but less than mass 443	83.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19 (19) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF143498.D	08/20/2025	22:05
PB169313BL	PB169313BL	BF143499.D	08/20/2025	22:34
MN-9C-081825	Q2900-01	BF143503.D	08/21/2025	00:29
MN-12C-081825	Q2900-02	BF143511.D	08/21/2025	04:21
DUP-01-081825	Q2900-03	BF143512.D	08/21/2025	04:49
MW-11C-081825	Q2900-05	BF143513.D	08/21/2025	05:18

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BF143517.D
Instrument ID: BNA_F

Contract: AECO02
SDG NO.: Q2900
DFTPP Injection Date: 08/21/2025
DFTPP Injection Time: 08:20

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.5 (1.9) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.6) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
365	Greater than 1% of mass 198	3.8
441	Present, but less than mass 443	77.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.7 (19.7) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF143518.D	08/21/2025	09:21
MN-12C-081825DL	Q2900-02DL	BF143522.D	08/21/2025	11:18
DUP-01-081825DL	Q2900-03DL	BF143523.D	08/21/2025	11:47
MW-11C-081825DL	Q2900-05DL	BF143524.D	08/21/2025	12:17
MN-12C-081825DL2	Q2900-02DL2	BF143527.D	08/21/2025	13:44
DUP-01-081825DL2	Q2900-03DL2	BF143528.D	08/21/2025	14:13
MW-11C-081825DL2	Q2900-05DL2	BF143529.D	08/21/2025	14:42

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2900

Client ID : SSTDICCC040

Date Analyzed: 08/20/2025

Lab File ID: BF143481.D

Time Analyzed: 12:53

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	137770	6.928	521928	8.21	282852	9.97
UPPER LIMIT	275540	7.428	1043860	8.71	565704	10.469
LOWER LIMIT	68885	6.428	260964	7.71	141426	9.469
EPA SAMPLE NO.						
01 MW-17B-081825	131839	6.93	507504	8.20	273837	9.96
02 MW-17C-081825	136964	6.93	522451	8.20	285056	9.96
03 MW-11B-081825	135207	6.93	510064	8.20	268416	9.96
04 PB169313BS	135406	6.93	517174	8.21	285363	9.97
05 PB169313BSD	134095	6.93	510372	8.21	281230	9.97

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2900

Client ID: SSTDICCC040

Date Analyzed: 08/20/2025

Lab File ID: BF143481.D

Time Analyzed: 12:53

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	436106	11.457	249960	14.098	288139	15.598
UPPER LIMIT	872212	11.957	499920	14.598	576278	16.098
LOWER LIMIT	218053	10.957	124980	13.598	144070	15.098
EPA SAMPLE NO.						
01 MW-17B-081825	429631	11.45	242165	14.09	249186	15.59
02 MW-17C-081825	458576	11.45	262919	14.10	271441	15.59
03 MW-11B-081825	401142	11.45	238389	14.10	292027	15.59
04 PB169313BS	434773	11.46	242024	14.10	290471	15.60
05 PB169313BSD	429898	11.46	234937	14.10	275165	15.60

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2900

Client ID : SSTDCCC040

Date Analyzed: 08/20/2025

Lab File ID: BF143498.D

Time Analyzed: 22:05

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	138596	6.928	522619	8.21	286971	9.97
UPPER LIMIT	277192	7.428	1045240	8.71	573942	10.469
LOWER LIMIT	69298	6.428	261310	7.71	143486	9.469
EPA SAMPLE NO.						
01 PB169313BL	129120	6.93	502778	8.20	274867	9.96
02 MN-9C-081825	128002	6.93	499284	8.20	269714	9.96
03 MN-12C-081825	121876	6.93	165304 *	8.23	245299	9.96
04 DUP-01-081825	127848	6.94	161605 *	8.23	252028	9.96
05 MW-11C-081825	125223	6.93	212422 *	8.23	251603	9.96

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2900

Client ID: SSTDCCC040

Date Analyzed: 08/20/2025

Lab File ID: BF143498.D

Time Analyzed: 22:05

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	448105	11.451	253253	14.098	290056	15.592
UPPER LIMIT	896210	11.951	506506	14.598	580112	16.092
LOWER LIMIT	224053	10.951	126627	13.598	145028	15.092
EPA SAMPLE NO.						
01 PB169313BL	442472	11.45	240607	14.10	242779	15.59
02 MN-9C-081825	422264	11.45	236275	14.09	262288	15.59
03 MN-12C-081825	368458	11.45	219458	14.09	260219	15.59
04 DUP-01-081825	378389	11.45	228850	14.09	280609	15.59
05 MW-11C-081825	379796	11.45	223331	14.09	273794	15.59

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2900

Client ID : SSTDCCC040

Date Analyzed: 08/21/2025

Lab File ID: BF143518.D

Time Analyzed: 09:21

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	118001	6.928	448774	8.21	244595	9.96
UPPER LIMIT	236002	7.428	897548	8.71	489190	10.463
LOWER LIMIT	59000.5	6.428	224387	7.71	122298	9.463
EPA SAMPLE NO.						
01 MN-12C-081825DL	109819	6.93	413990	8.21	236192	9.96
02 MN-12C-081825DL2	101103	6.93	381650	8.21	199099	9.96
03 DUP-01-081825DL	120053	6.93	450034	8.21	257551	9.96
04 DUP-01-081825DL2	106391	6.93	406559	8.21	215536	9.96
05 MW-11C-081825DL	113395	6.92	431449	8.21	238091	9.96
06 MW-11C-081825DL2	101432	6.93	390858	8.20	202413	9.96

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2900

Client ID: SSTDCCC040

Date Analyzed: 08/21/2025

Lab File ID: BF143518.D

Time Analyzed: 09:21

Instrument ID: BNA_F

GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	369706	11.451	222503	14.098	262934	15.592
UPPER LIMIT	739412	11.951	445006	14.598	525868	16.092
LOWER LIMIT	184853	10.951	111252	13.598	131467	15.092
EPA SAMPLE NO.						
01 MN-12C-081825DL	347629	11.45	223408	14.09	252278	15.59
02 MN-12C-081825DL2	282280	11.45	191729	14.10	235106	15.60
03 DUP-01-081825DL	381973	11.45	254066	14.09	290998	15.59
04 DUP-01-081825DL2	311169	11.45	210624	14.09	252482	15.59
05 MW-11C-081825DL	357150	11.45	225989	14.09	263314	15.59
06 MW-11C-081825DL2	279024	11.45	183510	14.09	231339	15.59

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	AECOM		Date Collected:		
Project:	National Grid Equity - Brooklyn NY		Date Received:		
Client Sample ID:	PB169313BL		SDG No.:	Q2900	
Lab Sample ID:	PB169313BL		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143499.D	1	08/19/25 09:29	08/20/25 22:34	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
100-52-7	Benzaldehyde	10.0	U	3.90	10.0	ug/L
108-95-2	Phenol	5.00	U	0.91	5.00	ug/L
111-44-4	bis(2-Chloroethyl)ether	5.00	U	0.81	5.00	ug/L
95-57-8	2-Chlorophenol	5.00	U	0.58	5.00	ug/L
95-48-7	2-Methylphenol	5.00	U	1.10	5.00	ug/L
108-60-1	2,2-oxybis(1-Chloropropane)	5.00	U	1.30	5.00	ug/L
98-86-2	Acetophenone	5.00	U	0.74	5.00	ug/L
65794-96-9	3+4-Methylphenols	10.0	U	1.10	10.0	ug/L
621-64-7	n-Nitroso-di-n-propylamine	2.50	U	1.40	2.50	ug/L
67-72-1	Hexachloroethane	5.00	U	0.65	5.00	ug/L
98-95-3	Nitrobenzene	5.00	U	0.76	5.00	ug/L
78-59-1	Isophorone	5.00	U	0.75	5.00	ug/L
88-75-5	2-Nitrophenol	5.00	U	1.80	5.00	ug/L
105-67-9	2,4-Dimethylphenol	5.00	U	1.90	5.00	ug/L
111-91-1	bis(2-Chloroethoxy)methane	5.00	U	0.68	5.00	ug/L
120-83-2	2,4-Dichlorophenol	5.00	U	0.52	5.00	ug/L
91-20-3	Naphthalene	5.00	U	0.50	5.00	ug/L
106-47-8	4-Chloroaniline	5.00	U	0.84	5.00	ug/L
87-68-3	Hexachlorobutadiene	5.00	U	0.54	5.00	ug/L
105-60-2	Caprolactam	10.0	U	1.10	10.0	ug/L
59-50-7	4-Chloro-3-methylphenol	5.00	U	0.59	5.00	ug/L
91-57-6	2-Methylnaphthalene	5.00	U	0.56	5.00	ug/L
77-47-4	Hexachlorocyclopentadiene	10.0	U	3.60	10.0	ug/L
88-06-2	2,4,6-Trichlorophenol	5.00	U	0.51	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	5.00	U	0.62	5.00	ug/L
92-52-4	1,1-Biphenyl	5.00	U	0.53	5.00	ug/L
91-58-7	2-Chloronaphthalene	5.00	U	0.61	5.00	ug/L
88-74-4	2-Nitroaniline	5.00	U	1.30	5.00	ug/L
131-11-3	Dimethylphthalate	5.00	U	0.61	5.00	ug/L

Report of Analysis

Client:	AECOM		Date Collected:		
Project:	National Grid Equity - Brooklyn NY		Date Received:		
Client Sample ID:	PB169313BL		SDG No.:	Q2900	
Lab Sample ID:	PB169313BL		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143499.D	1	08/19/25 09:29	08/20/25 22:34	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
208-96-8	Acenaphthylene	5.00	U	0.75	5.00	ug/L
606-20-2	2,6-Dinitrotoluene	5.00	U	0.92	5.00	ug/L
99-09-2	3-Nitroaniline	5.00	U	1.10	5.00	ug/L
83-32-9	Acenaphthene	5.00	U	0.55	5.00	ug/L
51-28-5	2,4-Dinitrophenol	10.0	U	6.00	10.0	ug/L
100-02-7	4-Nitrophenol	10.0	U	2.40	10.0	ug/L
132-64-9	Dibenzofuran	5.00	U	0.61	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	5.00	U	1.20	5.00	ug/L
84-66-2	Diethylphthalate	5.00	U	0.69	5.00	ug/L
7005-72-3	4-Chlorophenyl-phenylether	5.00	U	0.68	5.00	ug/L
86-73-7	Fluorene	5.00	U	0.63	5.00	ug/L
100-01-6	4-Nitroaniline	5.00	U	1.50	5.00	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	10.0	U	2.90	10.0	ug/L
86-30-6	n-Nitrosodiphenylamine	5.00	U	0.58	5.00	ug/L
101-55-3	4-Bromophenyl-phenylether	5.00	U	0.40	5.00	ug/L
118-74-1	Hexachlorobenzene	5.00	U	0.52	5.00	ug/L
1912-24-9	Atrazine	5.00	U	1.00	5.00	ug/L
87-86-5	Pentachlorophenol	10.0	U	1.60	10.0	ug/L
85-01-8	Phenanthrene	5.00	U	0.50	5.00	ug/L
120-12-7	Anthracene	5.00	U	0.61	5.00	ug/L
86-74-8	Carbazole	5.00	U	0.72	5.00	ug/L
84-74-2	Di-n-butylphthalate	5.00	U	1.20	5.00	ug/L
206-44-0	Fluoranthene	5.00	U	0.82	5.00	ug/L
129-00-0	Pyrene	5.00	U	0.50	5.00	ug/L
85-68-7	Butylbenzylphthalate	5.00	U	1.90	5.00	ug/L
91-94-1	3,3-Dichlorobenzidine	10.0	U	0.93	10.0	ug/L
56-55-3	Benzo(a)anthracene	5.00	U	0.45	5.00	ug/L
218-01-9	Chrysene	5.00	U	0.44	5.00	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	5.00	U	1.60	5.00	ug/L
117-84-0	Di-n-octyl phthalate	10.0	U	2.30	10.0	ug/L
205-99-2	Benzo(b)fluoranthene	5.00	U	0.49	5.00	ug/L

Report of Analysis

Client:	AECOM		Date Collected:		
Project:	National Grid Equity - Brooklyn NY		Date Received:		
Client Sample ID:	PB169313BL		SDG No.:	Q2900	
Lab Sample ID:	PB169313BL		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143499.D	1	08/19/25 09:29	08/20/25 22:34	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
207-08-9	Benzo(k)fluoranthene	5.00	U	0.48	5.00	ug/L
50-32-8	Benzo(a)pyrene	5.00	U	0.55	5.00	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	5.00	U	0.59	5.00	ug/L
53-70-3	Dibenzo(a,h)anthracene	5.00	U	0.67	5.00	ug/L
191-24-2	Benzo(g,h,i)perylene	5.00	U	0.69	5.00	ug/L
95-94-3	1,2,4,5-Tetrachlorobenzene	5.00	U	0.52	5.00	ug/L
123-91-1	1,4-Dioxane	5.00	U	1.00	5.00	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	5.00	U	0.72	5.00	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	120		23 - 138	80%	SPK: 150
13127-88-3	Phenol-d6	121		10 - 134	81%	SPK: 150
4165-60-0	Nitrobenzene-d5	85.9		67 - 132	86%	SPK: 100
321-60-8	2-Fluorobiphenyl	83.4		52 - 132	83%	SPK: 100
118-79-6	2,4,6-Tribromophenol	121		44 - 137	81%	SPK: 150
1718-51-0	Terphenyl-d14	94.0		42 - 152	94%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	129000	6.928			
1146-65-2	Naphthalene-d8	503000	8.204			
15067-26-2	Acenaphthene-d10	275000	9.963			
1517-22-2	Phenanthrene-d10	442000	11.451			
1719-03-5	Chrysene-d12	241000	14.098			
1520-96-3	Perylene-d12	243000	15.592			
TENTATIVE IDENTIFIED COMPOUNDS						
000123-42-2	2-Pentanone, 4-hydroxy-4-methyl-	9.30	A		5.17	ug/L
006311-48-4	(1,1-Biphenyl)-4,4-diamine, N,N	6.80	J		17.4	ug/L

Report of Analysis

Client:	AECOM		Date Collected:		
Project:	National Grid Equity - Brooklyn NY		Date Received:		
Client Sample ID:	PB169313BL		SDG No.:	Q2900	
Lab Sample ID:	PB169313BL		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143499.D	1	08/19/25 09:29	08/20/25 22:34	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM		Date Collected:		
Project:	National Grid Equity - Brooklyn NY		Date Received:		
Client Sample ID:	PB169313BS		SDG No.:	Q2900	
Lab Sample ID:	PB169313BS		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143488.D	1	08/19/25 09:29	08/20/25 16:44	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
100-52-7	Benzaldehyde	25.5		3.90	10.0	ug/L
108-95-2	Phenol	42.6		0.91	5.00	ug/L
111-44-4	bis(2-Chloroethyl)ether	44.6		0.81	5.00	ug/L
95-57-8	2-Chlorophenol	45.3		0.58	5.00	ug/L
95-48-7	2-Methylphenol	45.7		1.10	5.00	ug/L
108-60-1	2,2-oxybis(1-Chloropropane)	44.5		1.30	5.00	ug/L
98-86-2	Acetophenone	45.9		0.74	5.00	ug/L
65794-96-9	3+4-Methylphenols	45.1		1.10	10.0	ug/L
621-64-7	n-Nitroso-di-n-propylamine	46.7		1.40	2.50	ug/L
67-72-1	Hexachloroethane	45.1		0.65	5.00	ug/L
98-95-3	Nitrobenzene	45.0		0.76	5.00	ug/L
78-59-1	Isophorone	46.3		0.75	5.00	ug/L
88-75-5	2-Nitrophenol	47.1		1.80	5.00	ug/L
105-67-9	2,4-Dimethylphenol	48.5		1.90	5.00	ug/L
111-91-1	bis(2-Chloroethoxy)methane	45.7		0.68	5.00	ug/L
120-83-2	2,4-Dichlorophenol	44.3		0.52	5.00	ug/L
91-20-3	Naphthalene	43.9		0.50	5.00	ug/L
106-47-8	4-Chloroaniline	29.3		0.84	5.00	ug/L
87-68-3	Hexachlorobutadiene	44.8		0.54	5.00	ug/L
105-60-2	Caprolactam	51.4		1.10	10.0	ug/L
59-50-7	4-Chloro-3-methylphenol	45.4		0.59	5.00	ug/L
91-57-6	2-Methylnaphthalene	40.4		0.56	5.00	ug/L
77-47-4	Hexachlorocyclopentadiene	120	E	3.60	10.0	ug/L
88-06-2	2,4,6-Trichlorophenol	45.3		0.51	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	44.2		0.62	5.00	ug/L
92-52-4	1,1-Biphenyl	44.7		0.53	5.00	ug/L
91-58-7	2-Chloronaphthalene	43.8		0.61	5.00	ug/L
88-74-4	2-Nitroaniline	46.2		1.30	5.00	ug/L
131-11-3	Dimethylphthalate	46.7		0.61	5.00	ug/L

Report of Analysis

Client:	AECOM		Date Collected:		
Project:	National Grid Equity - Brooklyn NY		Date Received:		
Client Sample ID:	PB169313BS		SDG No.:	Q2900	
Lab Sample ID:	PB169313BS		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143488.D	1	08/19/25 09:29	08/20/25 16:44	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
208-96-8	Acenaphthylene	49.5		0.75	5.00	ug/L
606-20-2	2,6-Dinitrotoluene	50.4		0.92	5.00	ug/L
99-09-2	3-Nitroaniline	34.3		1.10	5.00	ug/L
83-32-9	Acenaphthene	47.9		0.55	5.00	ug/L
51-28-5	2,4-Dinitrophenol	90.6	E	6.00	10.0	ug/L
100-02-7	4-Nitrophenol	99.2	E	2.40	10.0	ug/L
132-64-9	Dibenzofuran	44.9		0.61	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	50.9		1.20	5.00	ug/L
84-66-2	Diethylphthalate	47.7		0.69	5.00	ug/L
7005-72-3	4-Chlorophenyl-phenylether	45.2		0.68	5.00	ug/L
86-73-7	Fluorene	45.4		0.63	5.00	ug/L
100-01-6	4-Nitroaniline	48.5		1.50	5.00	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	52.6		2.90	10.0	ug/L
86-30-6	n-Nitrosodiphenylamine	47.9		0.58	5.00	ug/L
101-55-3	4-Bromophenyl-phenylether	46.0		0.40	5.00	ug/L
118-74-1	Hexachlorobenzene	45.7		0.52	5.00	ug/L
1912-24-9	Atrazine	49.7		1.00	5.00	ug/L
87-86-5	Pentachlorophenol	98.0	E	1.60	10.0	ug/L
85-01-8	Phenanthrene	45.1		0.50	5.00	ug/L
120-12-7	Anthracene	45.8		0.61	5.00	ug/L
86-74-8	Carbazole	46.1		0.72	5.00	ug/L
84-74-2	Di-n-butylphthalate	49.8		1.20	5.00	ug/L
206-44-0	Fluoranthene	45.7		0.82	5.00	ug/L
129-00-0	Pyrene	47.0		0.50	5.00	ug/L
85-68-7	Butylbenzylphthalate	50.7		1.90	5.00	ug/L
91-94-1	3,3-Dichlorobenzidine	33.0		0.93	10.0	ug/L
56-55-3	Benzo(a)anthracene	46.5		0.45	5.00	ug/L
218-01-9	Chrysene	48.8		0.44	5.00	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	51.0		1.60	5.00	ug/L
117-84-0	Di-n-octyl phthalate	51.1		2.30	10.0	ug/L
205-99-2	Benzo(b)fluoranthene	46.9		0.49	5.00	ug/L

Report of Analysis

Client:	AECOM		Date Collected:		
Project:	National Grid Equity - Brooklyn NY		Date Received:		
Client Sample ID:	PB169313BS		SDG No.:	Q2900	
Lab Sample ID:	PB169313BS		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143488.D	1	08/19/25 09:29	08/20/25 16:44	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
207-08-9	Benzo(k)fluoranthene	47.6		0.48	5.00	ug/L
50-32-8	Benzo(a)pyrene	49.7		0.55	5.00	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	46.9		0.59	5.00	ug/L
53-70-3	Dibenzo(a,h)anthracene	47.3		0.67	5.00	ug/L
191-24-2	Benzo(g,h,i)perylene	48.2		0.69	5.00	ug/L
95-94-3	1,2,4,5-Tetrachlorobenzene	43.3		0.52	5.00	ug/L
123-91-1	1,4-Dioxane	38.9		1.00	5.00	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	51.8		0.72	5.00	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	126		23 - 138	84%	SPK: 150
13127-88-3	Phenol-d6	127		10 - 134	84%	SPK: 150
4165-60-0	Nitrobenzene-d5	92.7		67 - 132	93%	SPK: 100
321-60-8	2-Fluorobiphenyl	86.6		52 - 132	87%	SPK: 100
118-79-6	2,4,6-Tribromophenol	140		44 - 137	93%	SPK: 150
1718-51-0	Terphenyl-d14	94.8		42 - 152	95%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	135000	6.928			
1146-65-2	Naphthalene-d8	517000	8.21			
15067-26-2	Acenaphthene-d10	285000	9.969			
1517-22-2	Phenanthrene-d10	435000	11.457			
1719-03-5	Chrysene-d12	242000	14.098			
1520-96-3	Perylene-d12	290000	15.598			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM		Date Collected:		
Project:	National Grid Equity - Brooklyn NY		Date Received:		
Client Sample ID:	PB169313BSD		SDG No.:	Q2900	
Lab Sample ID:	PB169313BSD		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143489.D	1	08/19/25 09:29	08/20/25 17:13	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
100-52-7	Benzaldehyde	26.7		3.90	10.0	ug/L
108-95-2	Phenol	43.9		0.91	5.00	ug/L
111-44-4	bis(2-Chloroethyl)ether	46.1		0.81	5.00	ug/L
95-57-8	2-Chlorophenol	46.9		0.58	5.00	ug/L
95-48-7	2-Methylphenol	47.2		1.10	5.00	ug/L
108-60-1	2,2-oxybis(1-Chloropropane)	46.8		1.30	5.00	ug/L
98-86-2	Acetophenone	47.0		0.74	5.00	ug/L
65794-96-9	3+4-Methylphenols	46.3		1.10	10.0	ug/L
621-64-7	n-Nitroso-di-n-propylamine	47.8		1.40	2.50	ug/L
67-72-1	Hexachloroethane	46.8		0.65	5.00	ug/L
98-95-3	Nitrobenzene	46.8		0.76	5.00	ug/L
78-59-1	Isophorone	48.7		0.75	5.00	ug/L
88-75-5	2-Nitrophenol	48.9		1.80	5.00	ug/L
105-67-9	2,4-Dimethylphenol	49.4		1.90	5.00	ug/L
111-91-1	bis(2-Chloroethoxy)methane	46.6		0.68	5.00	ug/L
120-83-2	2,4-Dichlorophenol	45.3		0.52	5.00	ug/L
91-20-3	Naphthalene	44.8		0.50	5.00	ug/L
106-47-8	4-Chloroaniline	12.9		0.84	5.00	ug/L
87-68-3	Hexachlorobutadiene	46.5		0.54	5.00	ug/L
105-60-2	Caprolactam	52.1		1.10	10.0	ug/L
59-50-7	4-Chloro-3-methylphenol	46.1		0.59	5.00	ug/L
91-57-6	2-Methylnaphthalene	41.2		0.56	5.00	ug/L
77-47-4	Hexachlorocyclopentadiene	120	E	3.60	10.0	ug/L
88-06-2	2,4,6-Trichlorophenol	45.7		0.51	5.00	ug/L
95-95-4	2,4,5-Trichlorophenol	44.0		0.62	5.00	ug/L
92-52-4	1,1-Biphenyl	44.9		0.53	5.00	ug/L
91-58-7	2-Chloronaphthalene	44.4		0.61	5.00	ug/L
88-74-4	2-Nitroaniline	47.5		1.30	5.00	ug/L
131-11-3	Dimethylphthalate	47.0		0.61	5.00	ug/L

Report of Analysis

Client:	AECOM		Date Collected:		
Project:	National Grid Equity - Brooklyn NY		Date Received:		
Client Sample ID:	PB169313BSD		SDG No.:	Q2900	
Lab Sample ID:	PB169313BSD		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143489.D	1	08/19/25 09:29	08/20/25 17:13	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
208-96-8	Acenaphthylene	50.0		0.75	5.00	ug/L
606-20-2	2,6-Dinitrotoluene	51.2		0.92	5.00	ug/L
99-09-2	3-Nitroaniline	24.3		1.10	5.00	ug/L
83-32-9	Acenaphthene	49.3		0.55	5.00	ug/L
51-28-5	2,4-Dinitrophenol	93.0	E	6.00	10.0	ug/L
100-02-7	4-Nitrophenol	100	E	2.40	10.0	ug/L
132-64-9	Dibenzofuran	44.9		0.61	5.00	ug/L
121-14-2	2,4-Dinitrotoluene	52.3		1.20	5.00	ug/L
84-66-2	Diethylphthalate	48.4		0.69	5.00	ug/L
7005-72-3	4-Chlorophenyl-phenylether	45.5		0.68	5.00	ug/L
86-73-7	Fluorene	45.8		0.63	5.00	ug/L
100-01-6	4-Nitroaniline	48.3		1.50	5.00	ug/L
534-52-1	4,6-Dinitro-2-methylphenol	53.3		2.90	10.0	ug/L
86-30-6	n-Nitrosodiphenylamine	48.3		0.58	5.00	ug/L
101-55-3	4-Bromophenyl-phenylether	46.8		0.40	5.00	ug/L
118-74-1	Hexachlorobenzene	47.1		0.52	5.00	ug/L
1912-24-9	Atrazine	52.6		1.00	5.00	ug/L
87-86-5	Pentachlorophenol	100	E	1.60	10.0	ug/L
85-01-8	Phenanthrene	46.8		0.50	5.00	ug/L
120-12-7	Anthracene	47.1		0.61	5.00	ug/L
86-74-8	Carbazole	47.8		0.72	5.00	ug/L
84-74-2	Di-n-butylphthalate	51.4		1.20	5.00	ug/L
206-44-0	Fluoranthene	46.9		0.82	5.00	ug/L
129-00-0	Pyrene	48.7		0.50	5.00	ug/L
85-68-7	Butylbenzylphthalate	53.8		1.90	5.00	ug/L
91-94-1	3,3-Dichlorobenzidine	20.5		0.93	10.0	ug/L
56-55-3	Benzo(a)anthracene	46.9		0.45	5.00	ug/L
218-01-9	Chrysene	49.3		0.44	5.00	ug/L
117-81-7	Bis(2-ethylhexyl)phthalate	50.7		1.60	5.00	ug/L
117-84-0	Di-n-octyl phthalate	51.3		2.30	10.0	ug/L
205-99-2	Benzo(b)fluoranthene	47.0		0.49	5.00	ug/L

Report of Analysis

Client:	AECOM		Date Collected:		
Project:	National Grid Equity - Brooklyn NY		Date Received:		
Client Sample ID:	PB169313BSD		SDG No.:	Q2900	
Lab Sample ID:	PB169313BSD		Matrix:	Water	
Analytical Method:	8270E		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-TCL BNA -20	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF143489.D	1	08/19/25 09:29	08/20/25 17:13	PB169313

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
207-08-9	Benzo(k)fluoranthene	49.1		0.48	5.00	ug/L
50-32-8	Benzo(a)pyrene	50.8		0.55	5.00	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	47.4		0.59	5.00	ug/L
53-70-3	Dibenzo(a,h)anthracene	48.4		0.67	5.00	ug/L
191-24-2	Benzo(g,h,i)perylene	49.0		0.69	5.00	ug/L
95-94-3	1,2,4,5-Tetrachlorobenzene	43.6		0.52	5.00	ug/L
123-91-1	1,4-Dioxane	39.4		1.00	5.00	ug/L
58-90-2	2,3,4,6-Tetrachlorophenol	52.0		0.72	5.00	ug/L
SURROGATES						
367-12-4	2-Fluorophenol	118		23 - 138	79%	SPK: 150
13127-88-3	Phenol-d6	120		10 - 134	80%	SPK: 150
4165-60-0	Nitrobenzene-d5	87.2		67 - 132	87%	SPK: 100
321-60-8	2-Fluorobiphenyl	81.4		52 - 132	81%	SPK: 100
118-79-6	2,4,6-Tribromophenol	127		44 - 137	84%	SPK: 150
1718-51-0	Terphenyl-d14	91.2		42 - 152	91%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	134000	6.928			
1146-65-2	Naphthalene-d8	510000	8.21			
15067-26-2	Acenaphthene-d10	281000	9.969			
1517-22-2	Phenanthrene-d10	430000	11.457			
1719-03-5	Chrysene-d12	235000	14.098			
1520-96-3	Perylene-d12	275000	15.598			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
 Method File : 8270-BF082025.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Aug 21 06:12:21 2025
 Response Via : Initial Calibration

Calibration Files

2.5 =BF143477.D 5 =BF143478.D 10 =BF143479.D 20 =BF143480.D 40 =BF143481.D 50 =BF143482.D 60 =BF143483.D 80 =BF143484.D

Compound		2.5	5	10	20	40	50	60	80	Avg	%RSD

1) I	1,4-Dichlorobenzen...	-----ISTD-----									
2)	1,4-Dioxane	0.528	0.529	0.539	0.536	0.519	0.542	0.522	0.531	1.62	
3)	Pyridine	1.339	1.392	1.435	1.438	1.389	1.477	1.421	1.413	3.13	
4)	n-Nitrosodimet...		0.604	0.619	0.641	0.633	0.666	0.648	0.635	3.44	
5) S	2-Fluorophenol	1.203	1.198	1.194	1.138	1.087	1.137	1.061	1.145	4.92	
6)	Aniline	2.034	1.969	1.959	1.890	1.800	1.864	1.714	1.890	5.78	
7) S	Phenol-d6	1.536	1.460	1.471	1.407	1.342	1.384	1.299	1.414	5.76	
8)	2-Chlorophenol	1.301	1.284	1.302	1.268	1.229	1.270	1.202	1.265	2.94	
9)	Benzaldehyde		1.008	0.981	1.200	1.133	1.402		1.145	14.80	
10) C	Phenol	1.741	1.695	1.725	1.646	1.554	1.593	1.449	1.629	6.43	
11)	bis(2-Chloroet...	1.296	1.230	1.235	1.213	1.170	1.221	1.156	1.217	3.78	
12)	1,3-Dichlorobe...	1.558	1.451	1.466	1.414	1.345	1.399	1.317	1.421	5.65	
13) C	1,4-Dichlorobe...	1.531	1.487	1.483	1.411	1.355	1.409	1.304	1.426	5.62	
14)	1,2-Dichlorobe...	1.448	1.413	1.386	1.348	1.289	1.333	1.233	1.350	5.45	
15)	Benzyl Alcohol		1.037	1.063	1.083	1.040	1.088	1.041	1.059	2.14	
16)	2,2'-oxybis(1-...	2.144	2.104	2.053	1.954	1.805	1.868	1.703	1.947	8.40	
17)	2-Methylphenol	1.052	1.042	1.056	1.011	0.975	1.011	0.960	1.015	3.67	
18)	Hexachloroethane	0.495	0.498	0.502	0.491	0.462	0.489	0.456	0.485	3.79	
19) P	n-Nitroso-di-n...	0.922	0.943	0.912	0.879	0.844	0.800	0.832	0.781	0.864	6.86
20)	3+4-Methylphenols		1.337	1.320	1.218	1.143	1.178	1.076	1.212	8.40	
21) I	Naphthalene-d8	-----ISTD-----									
22)	Acetophenone	0.489	0.464	0.459	0.418	0.402	0.410	0.382	0.432	8.99	
23) S	Nitrobenzene-d5	0.354	0.345	0.357	0.344	0.335	0.340	0.325	0.343	3.18	
24)	Nitrobenzene	0.353	0.351	0.366	0.356	0.348	0.358	0.337	0.353	2.55	
25)	Isophorone	0.666	0.630	0.644	0.621	0.608	0.625	0.603	0.628	3.41	
26) C	2-Nitrophenol	0.162	0.165	0.177	0.177	0.175	0.180	0.173	0.173	3.76	
27)	2,4-Dimethylph...	0.283	0.272	0.281	0.270	0.262	0.265	0.251	0.269	4.11	
28)	bis(2-Chloroet...	0.409	0.399	0.401	0.386	0.371	0.379	0.357	0.386	4.74	
29) C	2,4-Dichloroph...	0.293	0.299	0.303	0.288	0.281	0.283	0.266	0.288	4.30	
30)	1,2,4-Trichlor...	0.318	0.300	0.314	0.292	0.282	0.291	0.274	0.296	5.46	
31)	Naphthalene	1.069	1.009	1.014	0.943	0.907	0.921	0.859	0.960	7.58	
32)	Benzoic acid		0.086	0.116	0.145	0.153	0.163	0.160	0.137	21.97	
33)	4-Chloroaniline	0.359	0.345	0.362	0.343	0.331	0.330	0.315	0.341	4.87	
34) C	Hexachlorobuta...	0.207	0.199	0.201	0.192	0.186	0.190	0.177	0.193	5.21	
35)	Caprolactam		0.063	0.068	0.071	0.071	0.070	0.072	0.069	4.96	
36) C	4-Chloro-3-met...	0.303	0.292	0.303	0.287	0.280	0.284	0.269	0.288	4.25	
37)	2-Methylnaphth...	0.716	0.684	0.684	0.633	0.600	0.610	0.562	0.641	8.62	
38)	1-Methylnaphth...	0.692	0.651	0.650	0.603	0.577	0.581	0.540	0.613	8.62	

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
 Method File : 8270-BF082025.M

39) I	Acenaphthene-d10	-----ISTD-----									
40)	1,2,4,5-Tetrac...	0.629	0.584	0.613	0.565	0.544	0.552	0.517	0.572	6.88	
41) P	Hexachlorocycl...		0.098	0.153	0.198	0.205	0.226	0.222	0.184	26.99	
42) S	2,4,6-Tribromo...	0.191	0.188	0.194	0.188	0.181	0.184	0.177	0.186	3.12	
43) C	2,4,6-Trichlor...	0.338	0.347	0.364	0.364	0.351	0.360	0.339	0.352	3.18	
44)	2,4,5-Trichlor...	0.391	0.376	0.413	0.382	0.377	0.387	0.365	0.385	3.94	
45) S	2-Fluorobiphenyl	1.445	1.336	1.323	1.156	1.106	1.106	0.999	1.210	13.20	
46)	1,1'-Biphenyl	1.582	1.505	1.519	1.399	1.335	1.349	1.247	1.420	8.43	
47)	2-Chloronaphth...	1.222	1.165	1.172	1.105	1.056	1.068	1.008	1.114	6.77	
48)	2-Nitroaniline	0.314	0.313	0.338	0.337	0.329	0.331	0.320	0.326	3.19	
49)	Acenaphthylene	1.747	1.663	1.687	1.580	1.516	1.533	1.437	1.595	6.85	
50)	Dimethylphthalate	1.293	1.219	1.252	1.201	1.159	1.175	1.133	1.205	4.58	
51)	2,6-Dinitrotol...	0.224	0.237	0.259	0.259	0.260	0.257	0.251	0.250	5.63	
52) C	Acenaphthene	1.170	1.131	1.137	1.076	1.037	1.044	0.985	1.083	6.07	
53)	3-Nitroaniline	0.260	0.257	0.280	0.278	0.273	0.271	0.269	0.270	3.21	
54) P	2,4-Dinitrophenol		0.072	0.099	0.120	0.132	0.137	0.129	0.115	21.92	
55)	Dibenzofuran	1.717	1.625	1.647	1.523	1.461	1.453	1.381	1.544	7.91	
56) P	4-Nitrophenol		0.115	0.147	0.171	0.177	0.172	0.181	0.161	15.78	
57)	2,4-Dinitrotol...	0.292	0.318	0.349	0.351	0.344	0.344	0.336	0.333	6.38	
58)	Fluorene	1.367	1.286	1.270	1.158	1.111	1.102	1.025	1.188	10.25	
59)	2,3,4,6-Tetrac...	0.261	0.264	0.299	0.297	0.294	0.297	0.288	0.286	5.59	
60)	Diethylphthalate	1.172	1.097	1.110	1.084	1.053	1.036	1.008	1.080	5.01	
61)	4-Chlorophenyl...	0.671	0.624	0.628	0.585	0.558	0.567	0.527	0.594	8.30	
62)	4-Nitroaniline	0.230	0.237	0.258	0.257	0.255	0.248	0.250	0.248	4.28	
63)	Azobenzene	1.221	1.154	1.177	1.122	1.080	1.067	1.018	1.120	6.26	
64) I	Phenanthrene-d10	-----ISTD-----									
65)	4,6-Dinitro-2-...		0.085	0.103	0.119	0.120	0.124	0.126	0.113	13.97	
66) c	n-Nitrosodiphe...	0.684	0.661	0.679	0.641	0.615	0.636	0.590	0.644	5.29	
67)	4-Bromophenyl-...	0.243	0.240	0.250	0.241	0.230	0.240	0.227	0.239	3.35	
68)	Hexachlorobenzene	0.266	0.262	0.267	0.257	0.248	0.256	0.243	0.257	3.53	
69)	Atrazine	0.154	0.156	0.164	0.168	0.171	0.174	0.176	0.166	5.09	
70) C	Pentachlorophenol		0.088	0.106	0.124	0.127	0.131	0.131	0.118	14.47	
71)	Phenanthrene	1.188	1.123	1.130	1.039	1.015	1.033	0.969	1.071	7.23	
72)	Anthracene	1.191	1.144	1.148	1.074	1.022	1.036	0.980	1.085	7.16	
73)	Carbazole	0.949	0.925	0.932	0.882	0.849	0.861	0.825	0.889	5.30	
74)	Di-n-butylphth...	0.782	0.800	0.830	0.850	0.833	0.857	0.855	0.829	3.47	
75) C	Fluoranthene	1.061	1.020	1.002	0.946	0.914	0.923	0.862	0.961	7.21	
76) I	Chrysene-d12	-----ISTD-----									
77)	Benzidine		0.611	0.665	0.516	0.391	0.368	0.372	0.487	26.70	
78)	Pyrene	1.891	1.768	1.793	1.635	1.590	1.512	1.327	1.645	11.63	
79) S	Terphenyl-d14	1.341	1.255	1.248	1.130	1.094	1.052	0.904	1.146	12.86	
80)	Butylbenzylpht...	0.380	0.402	0.443	0.467	0.461	0.509	0.485	0.449	10.15	
81)	Benzo(a)anthra...	1.302	1.274	1.300	1.318	1.266	1.316	1.237	1.288	2.30	
82)	3,3'-Dichlorob...		0.353	0.386	0.373	0.354	0.377	0.370	0.369	3.60	
83)	Chrysene	1.251	1.195	1.213	1.113	1.089	1.125	1.114	1.157	5.33	
84)	Bis(2-ethylhex...	0.600	0.626	0.699	0.710	0.690	0.767	0.700	0.685	8.12	
85) c	Di-n-octyl pht...		1.129	1.289	1.298	1.251	1.413	1.228	1.268	7.36	

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
Method File : 8270-BF082025.M

86) I	Perylene-d12	-----ISTD-----								
87)	Indeno(1,2,3-c...	1.425	1.433	1.565	1.449	1.427	1.466	1.295	1.437	5.51
88)	Benzo(b)fluora...	1.135	1.147	1.182	1.074	1.058	1.068	1.119	1.112	4.20
89)	Benzo(k)fluora...	1.157	1.014	1.056	1.081	1.034	1.104	1.005	1.064	5.10
90) C	Benzo(a)pyrene	1.033	1.021	1.074	1.032	1.010	1.043	1.011	1.032	2.15
91)	Dibenzo(a,h)an...	1.139	1.164	1.260	1.162	1.142	1.168	1.023	1.151	6.04
92)	Benzo(g,h,i)pe...	1.128	1.129	1.242	1.156	1.145	1.172	1.039	1.145	5.30

(#) = Out of Range

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AEC002</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2900</u>
Instrument ID:	<u>BNA_F</u>	Calibration Date/Time:	<u>08/20/2025 22:05</u>
Lab File ID:	<u>BF143498.D</u>	Init. Calib. Date(s):	<u>08/20/2025 08/20/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>10:55 14:20</u>
GC Column:	<u>DB-UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.145	1.150		0.4	
Benzaldehyde	1.145	1.081		-5.6	
Phenol-d6	1.414	1.395		-1.3	
Phenol	1.629	1.649		1.2	20.0
bis(2-Chloroethyl)ether	1.217	1.211		-0.5	
2-Chlorophenol	1.265	1.270		0.4	
2-Methylphenol	1.015	1.021		0.6	
2,2-oxybis(1-Chloropropane)	1.947	1.949		0.1	
Acetophenone	0.432	0.424		-1.9	
3+4-Methylphenols	1.212	1.222		0.8	
n-Nitroso-di-n-propylamine	0.864	0.852	0.050	-1.4	
Nitrobenzene-d5	0.343	0.349		1.7	
Hexachloroethane	0.485	0.481		-0.8	
Nitrobenzene	0.353	0.362		2.5	
Isophorone	0.628	0.635		1.1	
2-Nitrophenol	0.173	0.181		4.6	20.0
2,4-Dimethylphenol	0.269	0.270		0.4	
bis(2-Chloroethoxy)methane	0.386	0.385		-0.3	
2,4-Dichlorophenol	0.288	0.290		0.7	20.0
Naphthalene	0.960	0.933		-2.8	
4-Chloroaniline	0.341	0.343		0.6	
Hexachlorobutadiene	0.193	0.192		-0.5	20.0
Caprolactam	0.069	0.078		13.0	
4-Chloro-3-methylphenol	0.288	0.293		1.7	20.0
2-Methylnaphthalene	0.641	0.630		-1.7	
Hexachlorocyclopentadiene	0.184	0.163	0.050	-11.4	
2,4,6-Trichlorophenol	0.352	0.356		1.1	20.0
2-Fluorobiphenyl	1.210	1.135		-6.2	
2,4,5-Trichlorophenol	0.385	0.377		-2.1	
1,1-Biphenyl	1.420	1.373		-3.3	
2-Chloronaphthalene	1.114	1.101		-1.2	
2-Nitroaniline	0.326	0.341		4.6	
Dimethylphthalate	1.205	1.228		1.9	
Acenaphthylene	1.595	1.574		-1.3	
2,6-Dinitrotoluene	0.250	0.262		4.8	
3-Nitroaniline	0.270	0.286		5.9	
Acenaphthene	1.083	1.061		-2.0	20.0
2,4-Dinitrophenol	0.115	0.099	0.050	-13.9	
4-Nitrophenol	0.161	0.174	0.050	8.1	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2900</u>
Instrument ID:	<u>BNA_F</u>	Calibration Date/Time:	<u>08/20/2025 22:05</u>
Lab File ID:	<u>BF143498.D</u>	Init. Calib. Date(s):	<u>08/20/2025 08/20/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>10:55 14:20</u>
GC Column:	<u>DB-UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.544	1.518		-1.7	
2,4-Dinitrotoluene	0.333	0.354		6.3	
Diethylphthalate	1.080	1.104		2.2	
4-Chlorophenyl-phenylether	0.594	0.578		-2.7	
Fluorene	1.188	1.156		-2.7	
4-Nitroaniline	0.248	0.268		8.1	
4,6-Dinitro-2-methylphenol	0.113	0.110		-2.7	
n-Nitrosodiphenylamine	0.644	0.627		-2.6	20.0
2,4,6-Tribromophenol	0.186	0.184		-1.1	
4-Bromophenyl-phenylether	0.239	0.237		-0.8	
Hexachlorobenzene	0.257	0.247		-3.9	
Atrazine	0.166	0.185		11.4	
Pentachlorophenol	0.118	0.119		0.8	20.0
Phenanthrene	1.071	1.039		-3.0	
Anthracene	1.085	1.059		-2.4	
Carbazole	0.889	0.887		-0.2	
Di-n-butylphthalate	0.829	0.911		9.9	
Fluoranthene	0.961	0.962		0.1	20.0
Pyrene	1.645	1.665		1.2	
Terphenyl-d14	1.146	1.151		0.4	
Butylbenzylphthalate	0.449	0.480		6.9	
3,3-Dichlorobenzidine	0.369	0.369		0.0	
Benzo (a) anthracene	1.288	1.254		-2.6	
Chrysene	1.157	1.176		1.6	
Bis(2-ethylhexyl)phthalate	0.685	0.667		-2.6	
Di-n-octyl phthalate	1.268	1.215		-4.2	20.0
Benzo (b) fluoranthene	1.112	1.107		-0.4	
Benzo (k) fluoranthene	1.064	1.049		-1.4	
Benzo (a) pyrene	1.032	1.033		0.1	20.0
Indeno (1,2,3-cd) pyrene	1.437	1.484		3.3	
Dibenzo (a,h) anthracene	1.151	1.194		3.7	
Benzo (g,h,i) perylene	1.145	1.178		2.9	
1,2,4,5-Tetrachlorobenzene	0.572	0.553		-3.3	
1,4-Dioxane	0.531	0.507		-4.5	20.0
2,3,4,6-Tetrachlorophenol	0.286	0.296		3.5	

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2900</u>
Instrument ID:	<u>BNA_F</u>	Calibration Date/Time:	<u>08/21/2025 09:21</u>
Lab File ID:	<u>BF143518.D</u>	Init. Calib. Date(s):	<u>08/20/2025 08/20/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>10:55 14:20</u>
GC Column:	<u>DB-UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.145	1.132		-1.1	
Benzaldehyde	1.145	1.256		9.7	
Phenol-d6	1.414	1.413		-0.1	
Phenol	1.629	1.644		0.9	20.0
bis(2-Chloroethyl)ether	1.217	1.227		0.8	
2-Chlorophenol	1.265	1.272		0.6	
2-Methylphenol	1.015	1.028		1.3	
2,2-oxybis(1-Chloropropane)	1.947	1.939		-0.4	
Acetophenone	0.432	0.427		-1.2	
3+4-Methylphenols	1.212	1.227		1.2	
n-Nitroso-di-n-propylamine	0.864	0.860	0.050	-0.5	
Nitrobenzene-d5	0.343	0.348		1.5	
Hexachloroethane	0.485	0.490		1.0	
Nitrobenzene	0.353	0.362		2.5	
Isophorone	0.628	0.630		0.3	
2-Nitrophenol	0.173	0.180		4.0	20.0
2,4-Dimethylphenol	0.269	0.270		0.4	
bis(2-Chloroethoxy)methane	0.386	0.384		-0.5	
2,4-Dichlorophenol	0.288	0.290		0.7	20.0
Naphthalene	0.960	0.950		-1.0	
4-Chloroaniline	0.341	0.339		-0.6	
Hexachlorobutadiene	0.193	0.194		0.5	20.0
Caprolactam	0.069	0.078		13.0	
4-Chloro-3-methylphenol	0.288	0.289		0.3	20.0
2-Methylnaphthalene	0.641	0.627		-2.2	
Hexachlorocyclopentadiene	0.184	0.169	0.050	-8.2	
2,4,6-Trichlorophenol	0.352	0.355		0.9	20.0
2-Fluorobiphenyl	1.210	1.177		-2.7	
2,4,5-Trichlorophenol	0.385	0.380		-1.3	
1,1-Biphenyl	1.420	1.391		-2.0	
2-Chloronaphthalene	1.114	1.096		-1.6	
2-Nitroaniline	0.326	0.335		2.8	
Dimethylphthalate	1.205	1.250		3.7	
Acenaphthylene	1.595	1.578		-1.1	
2,6-Dinitrotoluene	0.250	0.265		6.0	
3-Nitroaniline	0.270	0.278		3.0	
Acenaphthene	1.083	1.067		-1.5	20.0
2,4-Dinitrophenol	0.115	0.109	0.050	-5.2	
4-Nitrophenol	0.161	0.155	0.050	-3.7	

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2900</u>
Instrument ID:	<u>BNA_F</u>	Calibration Date/Time:	<u>08/21/2025 09:21</u>
Lab File ID:	<u>BF143518.D</u>	Init. Calib. Date(s):	<u>08/20/2025 08/20/2025</u>
EPA Sample No.:	<u>SSTDCCC040</u>	Init. Calib. Time(s):	<u>10:55 14:20</u>
GC Column:	<u>DB-UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
Dibenzofuran	1.544	1.525		-1.2	
2,4-Dinitrotoluene	0.333	0.357		7.2	
Diethylphthalate	1.080	1.140		5.6	
4-Chlorophenyl-phenylether	0.594	0.589		-0.8	
Fluorene	1.188	1.164		-2.0	
4-Nitroaniline	0.248	0.258		4.0	
4,6-Dinitro-2-methylphenol	0.113	0.107		-5.3	
n-Nitrosodiphenylamine	0.644	0.645		0.2	20.0
2,4,6-Tribromophenol	0.186	0.186		0.0	
4-Bromophenyl-phenylether	0.239	0.242		1.3	
Hexachlorobenzene	0.257	0.257		0.0	
Atrazine	0.166	0.195		17.5	
Pentachlorophenol	0.118	0.106		-10.2	20.0
Phenanthrene	1.071	1.052		-1.8	
Anthracene	1.085	1.065		-1.8	
Carbazole	0.889	0.875		-1.6	
Di-n-butylphthalate	0.829	0.956		15.3	
Fluoranthene	0.961	0.922		-4.1	20.0
Pyrene	1.645	1.514		-8.0	
Terphenyl-d14	1.146	1.094		-4.5	
Butylbenzylphthalate	0.449	0.485		8.0	
3,3-Dichlorobenzidine	0.369	0.377		2.2	
Benzo (a) anthracene	1.288	1.239		-3.8	
Chrysene	1.157	1.147		-0.9	
Bis(2-ethylhexyl)phthalate	0.685	0.692		1.0	
Di-n-octyl phthalate	1.268	1.238		-2.4	20.0
Benzo (b) fluoranthene	1.112	1.175		5.7	
Benzo (k) fluoranthene	1.064	0.980		-7.9	
Benzo (a) pyrene	1.032	1.024		-0.8	20.0
Indeno (1,2,3-cd) pyrene	1.437	1.375		-4.3	
Dibenzo (a,h) anthracene	1.151	1.117		-3.0	
Benzo (g,h,i) perylene	1.145	1.055		-7.9	
1,2,4,5-Tetrachlorobenzene	0.572	0.568		-0.7	
1,4-Dioxane	0.531	0.541		1.9	20.0
2,3,4,6-Tetrachlorophenol	0.286	0.289		1.0	

All other compounds must meet a minimum RRF of 0.010.