

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2900</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date(s):	<u>08/15/2025</u> <u>08/15/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Calibration Time(s):	<u>09:40</u> <u>12:30</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

LAB FILE ID:		RRF005 = VX047349.D		RRF020 = VX047350.D		RRF050 = VX047351.D		
		RRF100 = VX047352.D		RRF150 = VX047353.D		RRF001 = VX047356.D		
COMPOUND	RRF005	RRF020	RRF050	RRF100	RRF150	RRF001	RRF	% RSD
Dichlorodifluoromethane	0.558	0.638	0.731	0.671	0.680	0.546	0.637	11.4
Chloromethane	0.539	0.568	0.624	0.567	0.594	0.549	0.573	5.4
Vinyl Chloride	0.651	0.751	0.800	0.729	0.757	0.630	0.720	9.2
Bromomethane	0.251	0.296	0.336	0.309	0.261		0.291	11.9
Chloroethane	0.392	0.440	0.463	0.410	0.422	0.515	0.440	10
Trichlorofluoromethane	1.012	1.121	1.188	1.060	1.069	0.938	1.065	8.1
1,1,2-Trichlorotrifluoroethane	0.589	0.664	0.697	0.624	0.637	0.494	0.618	11.4
1,1-Dichloroethene	0.575	0.664	0.697	0.635	0.651	0.555	0.630	8.6
Acetone	0.259	0.310	0.280	0.240	0.240	0.249	0.263	10.5
Carbon Disulfide	1.747	1.916	1.993	1.819	1.854	2.069	1.900	6.2
Methyl tert-butyl Ether	1.768	2.019	2.175	1.962	2.014	1.554	1.915	11.5
Methyl Acetate	0.620	0.661	0.769	0.689	0.728	0.621	0.682	8.7
Methylene Chloride	0.645	0.732	0.757	0.669	0.684	0.691	0.697	5.9
trans-1,2-Dichloroethene	0.638	0.705	0.725	0.655	0.664	0.621	0.668	6
1,1-Dichloroethane	1.180	1.294	1.356	1.210	1.238	1.052	1.222	8.5
Cyclohexane	1.024	1.202	1.210	1.082	1.113		1.126	7
2-Butanone	0.328	0.378	0.389	0.342	0.343	0.287	0.344	10.6
Carbon Tetrachloride	0.536	0.596	0.607	0.549	0.554	0.515	0.560	6.3
cis-1,2-Dichloroethene	0.758	0.826	0.846	0.769	0.779	0.692	0.778	7
Bromochloromethane	0.543	0.402	0.540	0.560	0.525	0.481	0.508	11.6
Chloroform	1.236	1.328	1.368	1.214	1.243	1.091	1.247	7.8
1,1,1-Trichloroethane	1.043	1.151	1.195	1.081	1.106	0.863	1.073	10.8
Methylcyclohexane	0.589	0.673	0.695	0.640	0.642	0.595	0.639	6.5
Benzene	1.502	1.665	1.701	1.516	1.504	1.312	1.533	9.1
1,2-Dichloroethane	0.521	0.594	0.600	0.532	0.528	0.484	0.543	8.3
Trichloroethene	0.383	0.423	0.432	0.390	0.391	0.336	0.393	8.6
1,2-Dichloropropane	0.349	0.415	0.421	0.381	0.382	0.357	0.384	7.6
Bromodichloromethane	0.548	0.621	0.635	0.578	0.573	0.513	0.578	7.8
4-Methyl-2-Pentanone	0.408	0.479	0.509	0.451	0.444	0.346	0.440	13
Toluene	0.920	1.043	1.057	0.936	0.927	0.802	0.947	9.9

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: <u>Alliance</u>	Contract: <u>AECO02</u>
Lab Code: <u>ACE</u>	SDG No.: <u>Q2900</u>
Instrument ID: <u>MSVOA_X</u>	Calibration Date(s): <u>08/15/2025</u> <u>08/15/2025</u>
Heated Purge: (Y/N) <u>N</u>	Calibration Time(s): <u>09:40</u> <u>12:30</u>
GC Column: <u>DB-624UI</u> ID: <u>0.18</u> (mm)	

LAB FILE ID:		RRF005 = VX047349.D		RRF020 = VX047350.D		RRF050 = VX047351.D		
		RRF100 = VX047352.D		RRF150 = VX047353.D		RRF001 = VX047356.D		
COMPOUND	RRF005	RRF020	RRF050	RRF100	RRF150	RRF001	RRF	% RSD
t-1,3-Dichloropropene	0.444	0.537	0.584	0.541	0.550	0.371	0.505	15.9
cis-1,3-Dichloropropene	0.531	0.608	0.642	0.592	0.602	0.451	0.571	12
1,1,2-Trichloroethane	0.358	0.393	0.396	0.352	0.347	0.313	0.360	8.6
2-Hexanone	0.284	0.338	0.347	0.310	0.306	0.239	0.304	12.9
Dibromochloromethane	0.412	0.461	0.464	0.426	0.420	0.382	0.428	7.3
1,2-Dibromoethane	0.358	0.397	0.409	0.365	0.361	0.336	0.371	7.2
Tetrachloroethene	0.351	0.384	0.398	0.360	0.360	0.316	0.362	7.8
Chlorobenzene	1.163	1.268	1.286	1.163	1.175	1.073	1.188	6.6
Ethyl Benzene	1.941	2.216	2.277	2.062	2.056	1.718	2.045	9.8
m/p-Xylenes	0.722	0.831	0.858	0.765	0.759	0.642	0.763	10.2
o-Xylene	0.679	0.793	0.813	0.735	0.732	0.651	0.734	8.6
Styrene	1.186	1.397	1.411	1.271	1.266	0.909	1.240	14.8
Bromoform	0.294	0.322	0.335	0.306	0.313	0.241	0.302	10.9
Isopropylbenzene	3.547	4.192	4.399	4.048	4.080	3.211	3.913	11.4
1,1,2,2-Tetrachloroethane	1.077	1.208	1.269	1.142	1.152	1.044	1.149	7.2
1,3-Dichlorobenzene	1.719	1.907	1.983	1.811	1.823	1.710	1.825	5.8
1,4-Dichlorobenzene	1.788	1.952	1.986	1.788	1.814	1.936	1.877	4.8
1,2-Dichlorobenzene	1.617	1.851	1.862	1.705	1.729	1.631	1.733	6.1
1,2-Dibromo-3-Chloropropane	0.182	0.225	0.247	0.238	0.249	0.191	0.222	13
1,2,4-Trichlorobenzene	1.030	1.168	1.234	1.191	1.204	1.144	1.161	6.2
1,2,3-Trichlorobenzene	0.919	1.109	1.174	1.132	1.160	1.075	1.095	8.5
1,2-Dichloroethane-d4	0.709	0.581	0.700	0.747	0.762		0.700	10.2
Dibromofluoromethane	0.310	0.278	0.331	0.354	0.349		0.325	9.6
Toluene-d8	1.157	0.994	1.171	1.246	1.232		1.160	8.7
4-Bromofluorobenzene	0.441	0.371	0.428	0.452	0.448		0.428	7.8

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.