

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2900</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>AECOM</u>	Datafile :	<u>VX047430.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0820WBS01	Dichlorodifluoromethane	20	17.2	ug/L	86			69	116	
	Chloromethane	20	16.6	ug/L	83			65	116	
	Vinyl chloride	20	16.1	ug/L	81			65	117	
	Bromomethane	20	16.7	ug/L	84			58	125	
	Chloroethane	20	15.8	ug/L	79			56	128	
	Trichlorofluoromethane	20	17.1	ug/L	86			73	115	
	1,1,2-Trichlorotrifluoroethane	20	17.8	ug/L	89			80	112	
	1,1-Dichloroethene	20	16.8	ug/L	84			74	110	
	Acetone	100	89.6	ug/L	90			60	125	
	Carbon disulfide	20	15.5	ug/L	78			64	112	
	Methyl tert-butyl Ether	20	17.0	ug/L	85			78	114	
	Methyl Acetate	20	18.7	ug/L	94			67	125	
	Methylene Chloride	20	17.2	ug/L	86			72	114	
	trans-1,2-Dichloroethene	20	16.7	ug/L	84			75	108	
	1,1-Dichloroethane	20	17.2	ug/L	86			78	112	
	Cyclohexane	20	17.0	ug/L	85			75	110	
	2-Butanone	100	93.1	ug/L	93			65	122	
	Carbon Tetrachloride	20	17.5	ug/L	88			77	113	
	cis-1,2-Dichloroethene	20	17.1	ug/L	86			77	110	
	Bromochloromethane	20	19.9	ug/L	100			70	124	
	Chloroform	20	17.4	ug/L	87			79	113	
	1,1,1-Trichloroethane	20	17.2	ug/L	86			80	108	
	Methylcyclohexane	20	17.1	ug/L	86			72	115	
	Benzene	20	17.6	ug/L	88			82	109	
	1,2-Dichloroethane	20	17.6	ug/L	88			80	115	
	Trichloroethene	20	17.2	ug/L	86			77	113	
	1,2-Dichloropropane	20	17.3	ug/L	86			83	111	
	Bromodichloromethane	20	17.5	ug/L	88			83	110	
	4-Methyl-2-Pentanone	100	93.5	ug/L	94			74	118	
	Toluene	20	18.0	ug/L	90			82	110	
	t-1,3-Dichloropropene	20	17.6	ug/L	88			79	110	
	cis-1,3-Dichloropropene	20	17.4	ug/L	87			82	110	
	1,1,2-Trichloroethane	20	18.1	ug/L	91			83	112	
	2-Hexanone	100	91.6	ug/L	92			73	117	
	Dibromochloromethane	20	17.7	ug/L	89			82	110	
	1,2-Dibromoethane	20	17.4	ug/L	87			81	110	
	Tetrachloroethene	20	17.2	ug/L	86			67	123	
	Chlorobenzene	20	17.3	ug/L	86			82	109	
	Ethyl Benzene	20	17.6	ug/L	88			83	109	
	m/p-Xylenes	40	35.8	ug/L	90			82	110	
	o-Xylene	20	17.9	ug/L	90			83	109	
	Styrene	20	18.0	ug/L	90			80	111	
	Bromoform	20	17.6	ug/L	88			79	109	
	Isopropylbenzene	20	17.3	ug/L	86			83	112	
	1,1,2,2-Tetrachloroethane	20	17.6	ug/L	88			76	118	
	1,3-Dichlorobenzene	20	16.9	ug/L	85			82	108	
	1,4-Dichlorobenzene	20	17.0	ug/L	85			82	107	
	1,2-Dichlorobenzene	20	17.2	ug/L	86			82	109	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2900 Analytical Method: SW8260-Low
Client: AECOM Datafile : VX047430.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0820WBS01	1,2-Dibromo-3-Chloropropane	20	17.5	ug/L	88			68	112	
	1,2,4-Trichlorobenzene	20	16.3	ug/L	81			75	113	
	1,2,3-Trichlorobenzene	20	16.9	ug/L	85			76	114	