

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2900</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>AECOM</u>	Datafile :	<u>VX047431.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0820WBSD01	Dichlorodifluoromethane	20	16.4	ug/L	82	5		69	116	19
	Chloromethane	20	16.9	ug/L	85	2		65	116	21
	Vinyl chloride	20	16.3	ug/L	81	0		65	117	19
	Bromomethane	20	17.3	ug/L	86	2		58	125	20
	Chloroethane	20	16.5	ug/L	83	5		56	128	20
	Trichlorofluoromethane	20	16.6	ug/L	83	4		73	115	16
	1,1,2-Trichlorotrifluoroethane	20	17.5	ug/L	88	1		80	112	15
	1,1-Dichloroethene	20	16.3	ug/L	81	4		74	110	20
	Acetone	100	88.6	ug/L	89	1		60	125	20
	Carbon disulfide	20	15.5	ug/L	78	0		64	112	20
	Methyl tert-butyl Ether	20	17.9	ug/L	90	6		78	114	20
	Methyl Acetate	20	19.9	ug/L	100	6		67	125	20
	Methylene Chloride	20	17.6	ug/L	88	2		72	114	20
	trans-1,2-Dichloroethene	20	16.5	ug/L	83	1		75	108	16
	1,1-Dichloroethane	20	17.6	ug/L	88	2		78	112	20
	Cyclohexane	20	16.5	ug/L	83	2		75	110	20
	2-Butanone	100	99.5	ug/L	100	7		65	122	26
	Carbon Tetrachloride	20	17.1	ug/L	86	2		77	113	15
	cis-1,2-Dichloroethene	20	17.2	ug/L	86	0		77	110	20
	Bromochloromethane	20	20.5	ug/L	103	3		70	124	20
	Chloroform	20	18.0	ug/L	90	3		79	113	20
	1,1,1-Trichloroethane	20	17.3	ug/L	86	0		80	108	20
	Methylcyclohexane	20	16.6	ug/L	83	4		72	115	20
	Benzene	20	17.9	ug/L	90	2		82	109	15
	1,2-Dichloroethane	20	18.2	ug/L	91	3		80	115	20
	Trichloroethene	20	17.1	ug/L	86	0		77	113	15
	1,2-Dichloropropane	20	17.6	ug/L	88	2		83	111	16
	Bromodichloromethane	20	17.9	ug/L	90	2		83	110	16
	4-Methyl-2-Pentanone	100	98.0	ug/L	98	4		74	118	25
	Toluene	20	18.1	ug/L	91	1		82	110	16
	t-1,3-Dichloropropene	20	17.8	ug/L	89	1		79	110	20
	cis-1,3-Dichloropropene	20	17.8	ug/L	89	2		82	110	16
	1,1,2-Trichloroethane	20	18.9	ug/L	95	4		83	112	20
	2-Hexanone	100	95.5	ug/L	96	4		73	117	25
	Dibromochloromethane	20	18.3	ug/L	92	3		82	110	20
	1,2-Dibromoethane	20	19.0	ug/L	95	9		81	110	20
	Tetrachloroethene	20	17.0	ug/L	85	1		67	123	15
	Chlorobenzene	20	17.4	ug/L	87	1		82	109	15
	Ethyl Benzene	20	17.3	ug/L	86	2		83	109	16
	m/p-Xylenes	40	35.6	ug/L	89	1		82	110	15
	o-Xylene	20	17.3	ug/L	86	5		83	109	20
	Styrene	20	18.0	ug/L	90	0		80	111	17
	Bromoform	20	18.1	ug/L	91	3		79	109	20
	Isopropylbenzene	20	16.9	ug/L	85	1		83	112	29
	1,1,2,2-Tetrachloroethane	20	18.0	ug/L	90	2		76	118	20
	1,3-Dichlorobenzene	20	17.2	ug/L	86	1		82	108	20
	1,4-Dichlorobenzene	20	16.6	ug/L	83	2		82	107	15
	1,2-Dichlorobenzene	20	17.5	ug/L	88	2		82	109	20



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2900 Analytical Method: SW8260-Low
Client: AECOM Datafile : VX047431.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0820WBSD01	1,2-Dibromo-3-Chloropropane	20	17.5	ug/L	88	0		68	112	20
	1,2,4-Trichlorobenzene	20	16.6	ug/L	83	2		75	113	29
	1,2,3-Trichlorobenzene	20	17.4	ug/L	87	2		76	114	29