

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO02  
 Lab Code: ACE SDG NO.: Q2900  
 Lab File ID: VX047427.D Date Analyzed: 08/20/2025  
 Instrument ID: MSVOA\_X Time Analyzed: 09:25  
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #   |
|----------------|---------------|-------|---------------|-------|---------------|--------|
| 12 HOUR STD    | 222102        | 5.56  | 369648        | 6.76  | 338607        | 10.05  |
| UPPER LIMIT    | 444204        | 6.056 | 739296        | 7.263 | 677214        | 10.549 |
| LOWER LIMIT    | 111051        | 5.056 | 184824        | 6.263 | 169304        | 9.549  |
| EPA SAMPLE NO. |               |       |               |       |               |        |
| MN-9C-081825   | 203680        | 5.57  | 413174        | 6.78  | 412185        | 10.06  |
| MN-12C-081825  | 210277        | 5.57  | 348314        | 6.78  | 344337        | 10.06  |
| DUP-01-081825  | 210048        | 5.57  | 368410        | 6.78  | 367281        | 10.06  |
| MW-11C-081825  | 207711        | 5.57  | 360319        | 6.78  | 375441        | 10.06  |
| MW-11B-081825  | 269768        | 5.57  | 559138        | 6.78  | 560769        | 10.06  |
| VX0820WBL01    | 280589        | 5.57  | 574542        | 6.77  | 560929        | 10.06  |
| VX0820WBS01    | 244648        | 5.56  | 407608        | 6.76  | 368354        | 10.06  |
| VX0820WBSD01   | 227988        | 5.56  | 384011        | 6.76  | 353527        | 10.06  |

IS1 = Pentafluorobenzene  
 IS2 = 1,4-Difluorobenzene  
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.