

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO02
 Lab Code: ACE SDG NO.: Q2900
 Lab File ID: VX047427.D Date Analyzed: 08/20/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:25
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	165668	12.018				
UPPER LIMIT	331336	12.518				
LOWER LIMIT	82834	11.518				
EPA SAMPLE NO.						
MN-9C-081825	214243	12.02				
MN-12C-081825	172936	12.02				
DUP-01-081825	176836	12.02				
MW-11C-081825	174957	12.02				
MW-11B-081825	289774	12.02				
VX0820WBL01	279679	12.02				
VX0820WBS01	188940	12.02				
VX0820WBSD01	182846	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.