

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2900</u>
Lab File ID:	<u>VX047455.D</u>	Date Analyzed:	<u>08/21/2025</u>
Instrument ID:	<u>MSVOA_X</u>	Time Analyzed:	<u>09:35</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	170891	12.018				
UPPER LIMIT	341782	12.518				
LOWER LIMIT	85445.5	11.518				
EPA SAMPLE NO.						
MN-12C-081825DL	240978	12.02				
DUP-01-081825DL	233270	12.02				
MW-17B-081825	216444	12.02				
MW-11C-081825DL	229381	12.02				
MW-17C-081825	210796	12.02				
MW-11B-081825RE	247196	12.02				
TB-081825	239770	12.02				
VX0821WBL01	231897	12.02				
VX0821WBS02	193487	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.