

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATCE02

Lab Code: ACE

SDG NO.: Q2901

Continuing Calib Date: 08/20/2025

Initial Calibration Date(s): 08/18/2025 08/18/2025

Continuing Calib Time: 16:14

Initial Calibration Time(s): 09:53 18:24

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1248-1 (1)   | 4.75       | 4.75      | 4.65      | 4.85 | 0.00       |
| Aroclor-1248-2 (2)   | 4.98       | 4.98      | 4.88      | 5.08 | 0.00       |
| Aroclor-1248-3 (3)   | 5.20       | 5.20      | 5.10      | 5.30 | 0.00       |
| Aroclor-1248-4 (4)   | 5.55       | 5.55      | 5.45      | 5.65 | 0.00       |
| Aroclor-1248-5 (5)   | 5.59       | 5.59      | 5.49      | 5.69 | 0.00       |
| Tetrachloro-m-xylene | 3.66       | 3.66      | 3.56      | 3.76 | 0.00       |
| Decachlorobiphenyl   | 8.68       | 8.68      | 8.58      | 8.78 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATCE02

Lab Code: ACE

SDG NO.: Q2901

Continuing Calib Date: 08/20/2025

Initial Calibration Date(s): 08/18/2025 08/18/2025

Continuing Calib Time: 16:14

Initial Calibration Time(s): 09:53 18:24

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1248-1 (1)   | 4.73       | 4.73      | 4.63      | 4.83 | 0.00       |
| Aroclor-1248-2 (2)   | 4.96       | 4.96      | 4.86      | 5.06 | 0.00       |
| Aroclor-1248-3 (3)   | 5.01       | 5.01      | 4.91      | 5.11 | 0.00       |
| Aroclor-1248-4 (4)   | 5.18       | 5.18      | 5.08      | 5.28 | 0.00       |
| Aroclor-1248-5 (5)   | 5.57       | 5.57      | 5.47      | 5.67 | 0.00       |
| Tetrachloro-m-xylene | 3.66       | 3.66      | 3.56      | 3.76 | 0.00       |
| Decachlorobiphenyl   | 8.63       | 8.63      | 8.53      | 8.73 | 0.00       |