

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2901
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL04 **Date Analyzed:** 08/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113111.D **Time Analyzed:** 00:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.747	4.648	4.848	545.420	500.000	9.1
Aroclor-1016-2	4.765	4.666	4.866	549.890	500.000	10.0
Aroclor-1016-3	4.822	4.724	4.924	556.100	500.000	11.2
Aroclor-1016-4	4.942	4.843	5.043	506.220	500.000	1.2
Aroclor-1016-5	5.198	5.099	5.299	538.050	500.000	7.6
Aroclor-1260-1	6.235	6.136	6.336	581.070	500.000	16.2
Aroclor-1260-2	6.425	6.326	6.526	555.720	500.000	11.1
Aroclor-1260-3	6.790	6.691	6.891	538.020	500.000	7.6
Aroclor-1260-4	7.049	6.951	7.151	537.520	500.000	7.5
Aroclor-1260-5	7.294	7.194	7.394	546.250	500.000	9.3
Decachlorobiphenyl	8.681	8.582	8.782	51.220	50.000	2.4
Tetrachloro-m-xylene	3.662	3.563	3.763	54.360	50.000	8.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2901
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL04 **Date Analyzed:** 08/21/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113111.D **Time Analyzed:** 00:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.727	4.629	4.829	522.170	500.000	4.4
Aroclor-1016-2	4.745	4.647	4.847	520.480	500.000	4.1
Aroclor-1016-3	4.920	4.822	5.022	513.790	500.000	2.8
Aroclor-1016-4	4.963	4.864	5.064	493.740	500.000	-1.3
Aroclor-1016-5	5.175	5.076	5.276	581.450	500.000	16.3
Aroclor-1260-1	6.202	6.104	6.304	510.900	500.000	2.2
Aroclor-1260-2	6.390	6.292	6.492	510.860	500.000	2.2
Aroclor-1260-3	6.542	6.443	6.643	472.200	500.000	-5.6
Aroclor-1260-4	7.011	6.912	7.112	493.830	500.000	-1.2
Aroclor-1260-5	7.253	7.155	7.355	488.540	500.000	-2.3
Decachlorobiphenyl	8.626	8.528	8.728	52.950	50.000	5.9
Tetrachloro-m-xylene	3.654	3.555	3.755	52.620	50.000	5.2