

Report of Analysis

Client:	ATC Group Services LLC		Date Collected:	08/13/25	
Project:	Hazmat survey-Bellevue hospital CPEP		Date Received:	08/19/25	
Client Sample ID:	EXTERIOR WINDOW CAULKDL		SDG No.:	Q2901	
Lab Sample ID:	Q2901-02DL		Matrix:	SOIL	
Analytical Method:	8082A		% Solid:	100	Decanted:
Sample Wt/Vol:	0.68	Units: g	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PCB Group1	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	SW3541B				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO113110.D	5	08/20/25 08:40	08/20/25 23:14	PB169322

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	871	UD	871	3800	ug/kg
11104-28-2	Aroclor-1221	889	UD	889	3800	ug/kg
11141-16-5	Aroclor-1232	821	UD	821	3800	ug/kg
53469-21-9	Aroclor-1242	885	UD	885	3800	ug/kg
12672-29-6	Aroclor-1248	1300	UD	1300	3800	ug/kg
11097-69-1	Aroclor-1254	708	UD	708	3800	ug/kg
37324-23-5	Aroclor-1262	1100	UD	1100	3800	ug/kg
11100-14-4	Aroclor-1268	794	UD	794	3800	ug/kg
11096-82-5	Aroclor-1260	16000	D	713	3800	ug/kg
Total PCBs	Total PCBs	16000	D	713	3800	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.6		32 - 144	93%	SPK: 20
2051-24-3	Decachlorobiphenyl	21.4		32 - 175	107%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit