

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tris Pharma, Inc.

SDG No.: Q2902

Contract: TRIS02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	1000	1000	100	95 - 105	P	08/21/2025	14:24	LB136917
	Zinc	1980	2000	99	95 - 105	P	08/21/2025	14:24	LB136917

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tris Pharma, Inc.

SDG No.: Q2902

Contract: TRIS02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	21.4	20.0	107	80 - 120	P	08/21/2025	14:33	LB136917
	Zinc	41.4	40.0	104	80 - 120	P	08/21/2025	14:33	LB136917

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tris Pharma, Inc.

SDG No.: Q2902

Contract: TRIS02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1250	1250	100	90 - 110	P	08/21/2025	15:03	LB136917
	Zinc	2500	2500	100	90 - 110	P	08/21/2025	15:03	LB136917
CCV02	Copper	1260	1250	101	90 - 110	P	08/21/2025	16:23	LB136917
	Zinc	2530	2500	101	90 - 110	P	08/21/2025	16:23	LB136917
CCV03	Copper	1270	1250	101	90 - 110	P	08/21/2025	17:14	LB136917
	Zinc	2520	2500	101	90 - 110	P	08/21/2025	17:14	LB136917
CCV04	Copper	1280	1250	102	90 - 110	P	08/21/2025	18:04	LB136917
	Zinc	2550	2500	102	90 - 110	P	08/21/2025	18:04	LB136917
CCV05	Copper	1270	1250	101	90 - 110	P	08/21/2025	18:52	LB136917
	Zinc	2520	2500	101	90 - 110	P	08/21/2025	18:52	LB136917

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tris Pharma, Inc.

SDG No.: Q2902

Contract: TRIS02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	997	1000	100	95 - 105	P	08/22/2025	11:37	LB136940
	Zinc	1960	2000	98	95 - 105	P	08/22/2025	11:37	LB136940

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tris Pharma, Inc.

SDG No.: Q2902

Contract: TRIS02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	21.2	20.0	106	80 - 120	P	08/22/2025	11:45	LB136940
	Zinc	41.9	40.0	105	80 - 120	P	08/22/2025	11:45	LB136940

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tris Pharma, Inc.

SDG No.: Q2902

Contract: TRIS02

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1270	1250	101	90 - 110	P	08/22/2025	12:18	LB136940
	Zinc	2560	2500	102	90 - 110	P	08/22/2025	12:18	LB136940
CCV02	Copper	1270	1250	102	90 - 110	P	08/22/2025	13:36	LB136940
	Zinc	2570	2500	103	90 - 110	P	08/22/2025	13:36	LB136940
CCV03	Copper	1290	1250	103	90 - 110	P	08/22/2025	14:25	LB136940
	Zinc	2610	2500	104	90 - 110	P	08/22/2025	14:25	LB136940
CCV04	Copper	1240	1250	100	90 - 110	P	08/22/2025	15:44	LB136940
	Zinc	2540	2500	102	90 - 110	P	08/22/2025	15:44	LB136940
CCV05	Copper	1210	1250	97	90 - 110	P	08/22/2025	16:30	LB136940
	Zinc	2470	2500	99	90 - 110	P	08/22/2025	16:30	LB136940