



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tris Pharma, Inc.

SDG No.: Q2902

Contract: TRIS02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Copper	4.60	+/-10	U	20.0	P	08/21/2025	14:37	LB136917
	Zinc	16.7	+/-20	U	40.0	P	08/21/2025	14:37	LB136917

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tris Pharma, Inc.

SDG No.: Q2902

Contract: TRIS02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Copper	4.60	+/-10	U	20.0	P	08/21/2025	15:07	LB136917
	Zinc	16.7	+/-20	U	40.0	P	08/21/2025	15:07	LB136917
CCB02	Copper	4.60	+/-10	U	20.0	P	08/21/2025	16:27	LB136917
	Zinc	16.7	+/-20	U	40.0	P	08/21/2025	16:27	LB136917
CCB03	Copper	4.60	+/-10	U	20.0	P	08/21/2025	17:18	LB136917
	Zinc	16.7	+/-20	U	40.0	P	08/21/2025	17:18	LB136917
CCB04	Copper	4.60	+/-10	U	20.0	P	08/21/2025	18:08	LB136917
	Zinc	16.7	+/-20	U	40.0	P	08/21/2025	18:08	LB136917
CCB05	Copper	4.60	+/-10	U	20.0	P	08/21/2025	18:57	LB136917
	Zinc	16.7	+/-20	U	40.0	P	08/21/2025	18:57	LB136917

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tris Pharma, Inc.

SDG No.: Q2902

Contract: TRIS02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Copper	4.60	+/-10	U	20.0	P	08/22/2025	11:50	LB136940
	Zinc	16.7	+/-20	U	40.0	P	08/22/2025	11:50	LB136940

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tris Pharma, Inc.

SDG No.: Q2902

Contract: TRIS02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Copper	4.60	+/-10	U	20.0	P	08/22/2025	12:22	LB136940
	Zinc	16.7	+/-20	U	40.0	P	08/22/2025	12:22	LB136940
CCB02	Copper	4.60	+/-10	U	20.0	P	08/22/2025	13:40	LB136940
	Zinc	16.7	+/-20	U	40.0	P	08/22/2025	13:40	LB136940
CCB03	Copper	4.60	+/-10	U	20.0	P	08/22/2025	14:30	LB136940
	Zinc	16.7	+/-20	U	40.0	P	08/22/2025	14:30	LB136940
CCB04	Copper	4.60	+/-10	U	20.0	P	08/22/2025	15:48	LB136940
	Zinc	16.7	+/-20	U	40.0	P	08/22/2025	15:48	LB136940
CCB05	Copper	4.60	+/-10	U	20.0	P	08/22/2025	16:34	LB136940
	Zinc	16.7	+/-20	U	40.0	P	08/22/2025	16:34	LB136940