

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Tris Pharma, Inc.</u>	SDG No.: <u>Q2902</u>
Contract: <u>TRIS02</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	10.1	2.0	502	-18	22	08/21/2025	14:45	LB136917
	Zinc	1.65			-40	40	08/21/2025	14:45	LB136917
ICSAB01	Copper	478	511	94	434	588	08/21/2025	14:50	LB136917
	Zinc	1010	952	106	809	1095	08/21/2025	14:50	LB136917
ICSA	Copper	14.3	2.0	715	-18	22	08/21/2025	14:54	LB136917
	Zinc	-16.0			-40	40	08/21/2025	14:54	LB136917
ICSAB	Copper	498	511	98	434	588	08/21/2025	14:59	LB136917
	Zinc	975	952	102	809	1095	08/21/2025	14:59	LB136917
ICSA01	Copper	8.67	2.0	434	-18	22	08/22/2025	12:00	LB136940
	Zinc	2.57			-40	40	08/22/2025	12:00	LB136940
ICSAB01	Copper	491	511	96	434	588	08/22/2025	12:04	LB136940
	Zinc	1040	952	109	809	1095	08/22/2025	12:04	LB136940
ICSA	Copper	3.62	2.0	181	-18	22	08/22/2025	12:09	LB136940
	Zinc	-9.86			-40	40	08/22/2025	12:09	LB136940
ICSAB	Copper	507	511	99	434	588	08/22/2025	12:13	LB136940
	Zinc	1020	952	107	809	1095	08/22/2025	12:13	LB136940