

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Tris Pharma, Inc.</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q2902</u>
Contract:	<u>TRIS02</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q2902-04</u>	Client ID:	<u>OUTFALL-DSN-002DUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q2902-04DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Copper	ug/L	20	42.1		44.7		6		P
Zinc	ug/L	20	122		16.9	J	151	*	P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Tris Pharma, Inc.</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q2902</u>
Contract:	<u>TRIS02</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q2902-04MS</u>	Client ID:	<u>OUTFALL-DSN-002MSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q2902-04MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Copper	ug/L	20	192		199		4		P
Zinc	ug/L	20	120		125		4		P