

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TRIS02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q2902</u>
Instrument ID:	<u>MSVOA_N</u>	Calibration Date/Time:	<u>08/20/2025</u> <u>08:58</u>
Lab File ID:	<u>VN087603.D</u>	Init. Calib. Date(s):	<u>08/05/2025</u> <u>08/05/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>13:37</u> <u>15:01</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Ethyl Acetate	0.697	0.596		-14.49	
Isopropyl Acetate	1.181	0.984		-16.68	
Acetone	0.378	0.445		17.73	
Methylene Chloride	2.115	1.929		-8.79	
1,2-Dichloroethane-d4	2.767	2.683	0.01	-3.04	
Toluene-d8	1.380	1.403	0.01	1.67	
4-Bromofluorobenzene	0.516	0.503	0.2	-2.52	
n-amyl Acetate	0.652	0.673		3.22	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.