

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2902</u>	Analytical Method:	<u>SW624.1</u>
Client:	<u>Tris Pharma, Inc.</u>	Datafile :	<u>VN087604.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VN0820WBS01	Ethyl Acetate	20	17.9	ug/L	90			68	140	
	Isopropyl Acetate	20	17.0	ug/L	85			70	130	
	Acetone	100	100	ug/L	100			41	148	
	Methylene Chloride	20	17.5	ug/L	88			60	140	
	n-amyl Acetate	20	18.7	ug/L	94			70	130	