

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TRIS02
 Lab Code: ACE SDG NO.: Q2902
 Lab File ID: VN087603.D Date Analyzed: 08/20/2025
 Instrument ID: MSVOA_N Time Analyzed: 08:58
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	57702	7.80	317393	9.08	285286	11.85
UPPER LIMIT	115404	8.3	634786	9.582	570572	12.347
LOWER LIMIT	28851	7.3	158697	8.582	142643	11.347
EPA SAMPLE NO.						
OUTFALL-DSN-001	49956	7.81	273753	9.08	243163	11.85
OUTFALL-DSN-001MS	50255	7.80	274038	9.09	252169	11.85
OUTFALL-DSN-001MSD	50973	7.81	275779	9.08	252017	11.85
OUTFALL-DSN-002	53651	7.81	289860	9.08	254089	11.85
VN0820WBL01	50156	7.80	275670	9.08	241360	11.85
VN0820WBS01	57311	7.80	304862	9.08	275151	11.85

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.