

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG No.: Q2903

Instrument ID: BNA_P

Calibration Date(s): 08/14/2025 08/14/2025

Calibration Time(s): 12:33 17:24

LAB FILE ID:		RRF2.5 = BP025392.D			RRF005 = BP025393.D		RRF010 = BP025394.D	
		RRF020 = BP025395.D			RRF040 = BP025396.D		RRF050 = BP025397.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.232	1.177	1.194	1.162	1.240	1.204	2.4
Benzaldehyde			0.987	0.956	1.366	1.262	1.082	17.1
Phenol-d6		1.483	1.467	1.533	1.513	1.641	1.544	4.2
Phenol		1.594	1.529	1.613	1.575	1.707	1.620	3.9
bis(2-Chloroethyl) ether		1.283	1.183	1.273	1.222	1.321	1.263	3.8
2-Chlorophenol		1.351	1.283	1.347	1.313	1.428	1.359	3.8
2-Methylphenol		1.075	1.030	1.123	1.107	1.198	1.125	5.4
2,2-oxybis(1-Chloropropane)		1.762	1.592	1.712	1.620	1.751	1.692	4.0
Acetophenone		0.515	0.487	0.506	0.485	0.523	0.502	3.0
3+4-Methylphenols			1.442	1.533	1.525	1.639	1.560	4.9
n-Nitroso-di-n-propylamine	1.014	1.066	0.959	1.056	0.992	1.053	1.023	3.7
Nitrobenzene-d5		0.395	0.376	0.392	0.383	0.419	0.395	3.7
Hexachloroethane		0.577	0.544	0.552	0.545	0.588	0.563	3.1
Nitrobenzene		0.350	0.342	0.348	0.344	0.375	0.353	3.2
Isophorone		0.734	0.672	0.700	0.674	0.730	0.700	3.7
2-Nitrophenol		0.167	0.163	0.177	0.181	0.197	0.181	7.4
2,4-Dimethylphenol		0.303	0.303	0.296	0.308	0.336	0.312	4.5
bis(2-Chloroethoxy) methane		0.443	0.407	0.425	0.407	0.439	0.423	3.6
2,4-Dichlorophenol		0.293	0.293	0.308	0.305	0.331	0.310	4.7
Naphthalene		1.069	1.020	1.037	1.000	1.087	1.040	3.1
4-Chloroaniline		0.453	0.435	0.447	0.452	0.492	0.459	4.2
Hexachlorobutadiene		0.222	0.205	0.211	0.205	0.218	0.211	3.1
Caprolactam			0.114	0.110	0.109	0.120	0.114	3.7
4-Chloro-3-methylphenol		0.366	0.335	0.347	0.346	0.373	0.355	3.9
2-Methylnaphthalene		0.716	0.653	0.678	0.652	0.700	0.676	3.7
Hexachlorocyclopentadiene			0.274	0.313	0.348	0.387	0.349	13.5
2,4,6-Trichlorophenol		0.368	0.364	0.381	0.382	0.421	0.390	5.6
2-Fluorobiphenyl		1.558	1.483	1.507	1.392	1.517	1.463	5.0
2,4,5-Trichlorophenol		0.417	0.385	0.424	0.424	0.457	0.427	5.5
1,1-Biphenyl		1.495	1.418	1.457	1.408	1.537	1.453	3.4
2-Chloronaphthalene		1.152	1.093	1.144	1.101	1.184	1.131	3.0
2-Nitroaniline		0.301	0.306	0.317	0.330	0.360	0.329	6.9
Dimethylphthalate		1.569	1.449	1.438	1.403	1.512	1.465	4.1
Acenaphthylene		1.907	1.806	1.848	1.806	1.992	1.871	3.6

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: AECO02

Lab Code: ACE

SDG No.: Q2903

Instrument ID: BNA_P

Calibration Date(s): 08/14/2025 08/14/2025

Calibration Time(s): 12:33 17:24

LAB FILE ID:		RRF2.5 = BP025392.D			RRF005 = BP025393.D		RRF010 = BP025394.D	
		RRF020 = BP025395.D			RRF040 = BP025396.D		RRF050 = BP025397.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2,6-Dinitrotoluene		0.297	0.287	0.302	0.309	0.331	0.309	4.9
3-Nitroaniline		0.318	0.329	0.337	0.342	0.384	0.347	6.6
Acenaphthene		1.149	1.082	1.100	1.041	1.155	1.098	3.9
2,4-Dinitrophenol			0.114	0.140	0.163	0.186	0.165	19.8
4-Nitrophenol			0.251	0.265	0.279	0.315	0.285	8.6
Dibenzofuran		1.820	1.711	1.737	1.677	1.811	1.736	3.4
2,4-Dinitrotoluene		0.404	0.407	0.431	0.438	0.479	0.440	6.6
Diethylphthalate		1.615	1.475	1.465	1.434	1.540	1.488	4.5
4-Chlorophenyl-phenylether		0.761	0.684	0.679	0.651	0.690	0.681	5.9
Fluorene		1.506	1.391	1.390	1.306	1.415	1.370	6.0
4-Nitroaniline		0.341	0.336	0.349	0.351	0.387	0.356	4.9
4,6-Dinitro-2-methylphenol			0.096	0.113	0.123	0.139	0.125	14.2
n-Nitrosodiphenylamine		0.636	0.603	0.607	0.591	0.634	0.610	3.0
2,4,6-Tribromophenol		0.266	0.259	0.267	0.263	0.286	0.269	3.4
4-Bromophenyl-phenylether		0.233	0.206	0.217	0.213	0.230	0.220	4.3
Hexachlorobenzene		0.264	0.253	0.254	0.242	0.264	0.255	3.1
Atrazine		0.229	0.226	0.228	0.218	0.240	0.228	2.7
Pentachlorophenol			0.135	0.154	0.157	0.177	0.161	9.8
Phenanthrene		1.162	1.101	1.111	1.040	1.135	1.099	3.9
Anthracene		1.161	1.124	1.133	1.079	1.175	1.122	3.4
Carbazole		1.079	1.052	1.074	1.007	1.111	1.057	3.3
Di-n-butylphthalate		1.345	1.255	1.332	1.244	1.374	1.300	3.9
Fluoranthene		1.370	1.323	1.348	1.228	1.361	1.311	4.1
Pyrene		1.398	1.261	1.287	1.274	1.320	1.300	4.0
Terphenyl-d14		1.249	1.100	1.097	1.045	1.100	1.093	7.2
Butylbenzylphthalate		0.589	0.557	0.592	0.579	0.614	0.589	3.6
3,3-Dichlorobenzidine			0.482	0.507	0.488	0.521	0.497	3.3
Benzo (a) anthracene		1.384	1.302	1.313	1.268	1.359	1.319	3.2
Chrysene		1.292	1.214	1.249	1.182	1.280	1.235	3.5
Bis (2-ethylhexyl) phthalate		0.907	0.826	0.874	0.831	0.901	0.867	3.7
Di-n-octyl phthalate			1.381	1.498	1.444	1.572	1.492	4.8
Benzo (b) fluoranthene		1.203	1.126	1.171	1.134	1.250	1.179	3.6
Benzo (k) fluoranthene		1.202	1.157	1.200	1.130	1.226	1.180	3.2
Benzo (a) pyrene		1.139	1.093	1.150	1.108	1.219	1.148	3.8
Indeno (1,2,3-cd) pyrene		1.454	1.385	1.440	1.411	1.569	1.467	4.4

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance

Contract: AEC002

Lab Code: ACE

SDG No.: Q2903

Instrument ID: BNA_P

Calibration Date(s): 08/14/2025 08/14/2025

Calibration Time(s): 12:33 17:24

LAB FILE ID:		RRF2.5 = BP025392.D			RRF005 = BP025393.D		RRF010 = BP025394.D	
		RRF020 = BP025395.D			RRF040 = BP025396.D		RRF050 = BP025397.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
Dibenzo (a,h) anthracene		1.180	1.125	1.179	1.156	1.290	1.199	4.6
Benzo (g,h,i) perylene		1.162	1.118	1.169	1.138	1.248	1.178	3.9
1,2,4,5-Tetrachlorobenzene		0.578	0.559	0.585	0.560	0.610	0.578	3.0
1,4-Dioxane		0.520	0.464	0.475	0.445	0.479	0.472	5.1
2,3,4,6-Tetrachlorophenol		0.375	0.360	0.366	0.368	0.402	0.377	4.0