

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081525\
 Data File : VX047353.D
 Acq On : 15 Aug 2025 11:05
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC150

Quant Time: Aug 18 02:55:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 18 02:51:35 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 08/18/2025
 Supervised By : Mahesh Dadoda 08/19/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	218039	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	378315	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	330090	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	153878	50.000	ug/l	# 0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	5.964	65	498398	163.328	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 326.660%#	
35) Dibromofluoromethane	5.397	113	396260	161.390	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 322.780%#	
50) Toluene-d8	8.653	98	1397755	159.272	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 318.540%#	
62) 4-Bromofluorobenzene	11.079	95	508844	157.125	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 314.240%#	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.185	85	445045	160.118	ug/l	98
3) Chloromethane	1.313	50	388526	155.385	ug/l	97
4) Vinyl Chloride	1.392	62	494933	157.702	ug/l	97
5) Bromomethane	1.618	94	170864	134.788	ug/l	100
6) Chloroethane	1.703	64	276042	143.758	ug/l	97
7) Trichlorofluoromethane	1.904	101	699331	150.647	ug/l	98
8) Diethyl Ether	2.148	74	251983	154.158	ug/l	98
9) 1,1,2-Trichlorotrifluo...	2.349	101	416485	154.664	ug/l	99
10) Methyl Iodide	2.471	142	794834	172.239	ug/l	98
11) Tert butyl alcohol	2.977	59	206679	778.553	ug/l	100
12) 1,1-Dichloroethene	2.337	96	425622	155.037	ug/l	99
13) Acrolein	2.252	56	429798	766.968	ug/l	100
14) Allyl chloride	2.678	41	735481	154.247	ug/l	99
15) Acrylonitrile	3.087	53	991254	772.742	ug/l	100
16) Acetone	2.392	43	784600	684.039	ug/l	99
17) Carbon Disulfide	2.532	76	1212975	146.420	ug/l	100
18) Methyl Acetate	2.715	43	476497	160.309	ug/l	99
19) Methyl tert-butyl Ether	3.130	73	1317168	157.713	ug/l	100
20) Methylene Chloride	2.806	84	447216	147.242	ug/l	98
21) trans-1,2-Dichloroethene	3.105	96	434634	149.166	ug/l	98
22) Diisopropyl ether	3.776	45	1429793	151.699	ug/l	99
23) Vinyl Acetate	3.733	43	6020474	779.316	ug/l	100
24) 1,1-Dichloroethane	3.623	63	809793	152.005	ug/l	99
25) 2-Butanone	4.568	43	1122803	747.433	ug/l	98
26) 2,2-Dichloropropane	4.495	77	627662	156.100	ug/l	100
27) cis-1,2-Dichloroethene	4.501	96	509580	150.154	ug/l	99
28) Bromochloromethane	4.916	49	343257	154.839	ug/l	100
29) Tetrahydrofuran	5.019	42	717541	739.017	ug/l	97
30) Chloroform	5.105	83	812939	149.521	ug/l	99
31) Cyclohexane	5.483	56	727704	148.180	ug/l	98
32) 1,1,1-Trichloroethane	5.397	97	723635	154.630	ug/l	99
36) 1,1-Dichloropropene	5.702	75	570157	147.104	ug/l	99
37) Ethyl Acetate	4.727	43	593362	147.465	ug/l	97
38) Carbon Tetrachloride	5.690	117	628733	148.444	ug/l	98
39) Methylcyclohexane	7.385	83	729123	150.814	ug/l	97
40) Benzene	6.050	78	1706889	147.130	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081525\
 Data File : VX047353.D
 Acq On : 15 Aug 2025 11:05
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC150

Quant Time: Aug 18 02:55:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 18 02:51:35 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :John Carlone 08/18/2025
 Supervised By :Mahesh Dadoda 08/19/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.940	41	262323	154.600	ug/l	98
42) 1,2-Dichloroethane	6.098	62	599255	145.829	ug/l	100
43) Isopropyl Acetate	6.348	43	910348	157.486	ug/l	100
44) Trichloroethene	7.129	130	443361	149.252	ug/l	98
45) 1,2-Dichloropropane	7.434	63	433512	149.161	ug/l	99
46) Dibromomethane	7.586	93	312886	148.288	ug/l	99
47) Bromodichloromethane	7.824	83	650767	148.773	ug/l	99
48) Methyl methacrylate	7.696	41	473316	157.749	ug/l	99
49) 1,4-Dioxane	7.677	88	88481	3105.263	ug/l	97
51) 4-Methyl-2-Pentanone	8.580	43	2521407	757.992	ug/l	100
52) Toluene	8.720	92	1052008	146.750	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	624499	163.583	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	682860	158.029	ug/l	100
55) 1,1,2-Trichloroethane	9.153	97	394302	144.868	ug/l	97
56) Ethyl methacrylate	9.116	69	646154	161.020	ug/l	99
57) 1,3-Dichloropropane	9.311	76	672087	144.911	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	1468000	748.853	ug/l	99
59) 2-Hexanone	9.433	43	1737616	755.168	ug/l	100
60) Dibromochloromethane	9.525	129	476650	147.358	ug/l	99
61) 1,2-Dibromoethane	9.610	107	410057	146.099	ug/l	99
64) Tetrachloroethene	9.275	164	356989	149.518	ug/l	99
65) Chlorobenzene	10.079	112	1163848	148.370	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.165	131	405123	152.780	ug/l	100
67) Ethyl Benzene	10.195	91	2036272	150.823	ug/l	99
68) m/p-Xylenes	10.299	106	1502542	298.466	ug/l	99
69) o-Xylene	10.640	106	725154	149.683	ug/l	100
70) Styrene	10.653	104	1253652	153.150	ug/l	99
71) Bromoform	10.799	173	309626	155.427	ug/l #	99
73) Isopropylbenzene	10.963	105	1883521	156.417	ug/l	99
74) N-amyl acetate	10.842	43	790124	172.791	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	531651	150.375	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	425332m	150.704	ug/l	
77) Bromobenzene	11.195	156	451000	153.378	ug/l	99
78) n-propylbenzene	11.305	91	2235938	155.473	ug/l	99
79) 2-Chlorotoluene	11.366	91	1306418	150.272	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	1536791	155.110	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	113338	154.967	ug/l	97
82) 4-Chlorotoluene	11.451	91	1541070	150.376	ug/l	100
83) tert-Butylbenzene	11.713	119	1542621	156.062	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	1535261	154.952	ug/l	99
85) sec-Butylbenzene	11.890	105	1912224	156.078	ug/l	100
86) p-Isopropyltoluene	12.006	119	1601345	154.459	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	841394	149.783	ug/l	100
88) 1,4-Dichlorobenzene	12.036	146	837395	144.939	ug/l	100
89) n-Butylbenzene	12.329	91	1489473	154.462	ug/l	100
90) Hexachloroethane	12.536	117	303295	166.757	ug/l	98
91) 1,2-Dichlorobenzene	12.335	146	798192	149.701	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.939	75	115101	168.342	ug/l	97
93) 1,2,4-Trichlorobenzene	13.585	180	555658	155.453	ug/l	100
94) Hexachlorobutadiene	13.725	225	176870	139.013	ug/l	98
95) Naphthalene	13.774	128	1715681	167.959	ug/l	99
96) 1,2,3-Trichlorobenzene	13.957	180	535287	158.872	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081525\
Data File : VX047353.D
Acq On : 15 Aug 2025 11:05
Operator : JC/MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED

Reviewed By :John Carlone 08/18/2025
Supervised By :Mahesh Dadoda 08/19/2025

Quant Time: Aug 18 02:55:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
Quant Title : SW846 8260
QLast Update : Mon Aug 18 02:51:35 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

