



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082225\
Data File : VX047483.D
Acq On : 22 Aug 2025 08:25
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081525W.M
Title : SW846 8260
Last Update : Mon Aug 18 04:18:28 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.7	19453	PASS
75	95	30	60	51.1	50448	PASS
95	95	100	100	100.0	98781	PASS
96	95	5	9	6.6	6551	PASS
173	174	0.00	2	1.2	839	PASS
174	95	50	100	71.3	70411	PASS
175	174	5	9	7.6	5361	PASS
176	174	95	101	97.7	68784	PASS
177	176	5	9	6.5	4464	PASS