

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO02
 Lab Code: ACE SDG NO.: Q2903
 Lab File ID: VX047455.D Date Analyzed: 08/21/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:35
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	229258	5.56	398048	6.76	349454	10.05
UPPER LIMIT	458516	6.056	796096	7.263	698908	10.549
LOWER LIMIT	114629	5.056	199024	6.263	174727	9.549
EPA SAMPLE NO.						
MW-6A-081925	212230	5.56	433989	6.77	433463	10.06
MW-1C-081925	227781	5.56	475332	6.77	479784	10.06
MW-6B-081925	225941	5.57	469925	6.77	476037	10.06
MW-1B-081925	202322	5.57	410659	6.77	411448	10.06
MW-12B-081925	263943	5.57	512897	6.77	520562	10.06
MW-18B-081925	256554	5.56	524902	6.77	528456	10.06
VX0821WBL01	216518	5.56	445927	6.77	448054	10.06
VX0821WBS02	219757	5.56	395822	6.77	370697	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.