

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO02
 Lab Code: ACE SDG NO.: Q2903
 Lab File ID: VX047484.D Date Analyzed: 08/22/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:00
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	202377	5.56	354638	6.76	324101	10.05
UPPER LIMIT	404754	6.056	709276	7.263	648202	10.549
LOWER LIMIT	101189	5.056	177319	6.263	162051	9.549
EPA SAMPLE NO.						
MW-1C-081925RE	215208	5.57	438650	6.77	454253	10.06
MW-6B-081925MS	173158	5.57	331997	6.77	317077	10.06
MW-6B-081925MSD	156491	5.56	296002	6.77	277028	10.06
MW-12B-081925DL	199358	5.56	400754	6.77	390061	10.06
DUP-02-081925	245699	5.56	503621	6.77	501212	10.06
MW-17A-081925	267608	5.56	550186	6.77	553949	10.06
MW-18A-081925	199275	5.56	403908	6.77	398616	10.06
MW-16B-081925	181722	5.56	373576	6.77	382152	10.06
VX0822WBL01	210866	5.56	425323	6.77	433100	10.06
VX0822WBS01	232449	5.56	415073	6.76	390461	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.