

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: AECO02
 Lab Code: ACE SDG NO.: Q2903
 Lab File ID: VX047511.D Date Analyzed: 08/25/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:14
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	225340	5.56	383910	6.76	348301	10.05
UPPER LIMIT	450680	6.056	767820	7.263	696602	10.549
LOWER LIMIT	112670	5.056	191955	6.263	174151	9.549
EPA SAMPLE NO.						
TB	298330	5.56	594534	6.77	589303	10.06
VX0825WBL01	288899	5.56	581002	6.77	569460	10.06
VX0825WBS01	227627	5.56	393751	6.76	357988	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.