

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>AECO02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2903</u>
Lab File ID:	<u>VX047484.D</u>	Date Analyzed:	<u>08/22/2025</u>
Instrument ID:	<u>MSVOA_X</u>	Time Analyzed:	<u>10:00</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	162878	12.018				
UPPER LIMIT	325756	12.518				
LOWER LIMIT	81439	11.518				
EPA SAMPLE NO.						
MW-1C-081925RE	240673	12.02				
MW-6B-081925MS	162431	12.02				
MW-6B-081925MSD	139791	12.02				
MW-12B-081925DL	192795	12.02				
DUP-02-081925	255941	12.02				
MW-17A-081925	286642	12.02				
MW-18A-081925	210868	12.02				
MW-16B-081925	203331	12.02				
VX0822WBL01	221983	12.02				
VX0822WBS01	202080	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.