

## DATA REPORTING QUALIFIERS- INORGANIC

For reporting results, the following “ Results Qualifiers” are used:

<b>J</b>	Indicates the reported value was obtained from a reading that was less than the Contract Required Detection Limit (CRDL), but greater than or equal to the Instrument Detection Limit (IDL).
<b>U</b>	Indicates the analyte was analyzed for, but not detected.
<b>ND</b>	Indicates the analyte was analyzed for, but not detected
<b>E</b>	Indicates the reported value is estimated because of the presence of interference
<b>M</b>	Indicates Duplicate injection precision not met.
<b>N</b>	Indicates the spiked sample recovery is not within control limits.
<b>S</b>	Indicates the reported value was determined by the Method of Standard Addition (MSA).
<b>*</b>	Indicates that the duplicate analysis is not within control limits.
<b>+</b>	Indicates the correlation coefficient for the MSA is less than 0.995.
<b>D</b>	Indicates the reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.
<b>M</b>	Method qualifiers “P” for ICP instrument “PM” for ICP when Microwave Digestion is used “CV” for Manual Cold Vapor AA “AV” for automated Cold Vapor AA “CA” for MIDI-Distillation Spectrophotometric “AS” for Semi -Automated Spectrophotometric “C” for Manual Spectrophotometric “T” for Titrimetric “NR” for analyte not required to be analyzed
<b>OR</b>	Indicates the analyte’s concentration exceeds the calibrated range of the instrument for that specific analysis.
<b>Q</b>	Indicates the LCS did not meet the control limits requirements
<b>H</b>	Sample Analysis Out Of Hold Time

## LAB CHRONICLE

<b>OrderID:</b>	Q2904	<b>OrderDate:</b>	8/20/2025 11:51:00 AM
<b>Client:</b>	Tully Environmental, Inc	<b>Project:</b>	Transfer Station-SPDES
<b>Contact:</b>	Dean Devoe	<b>Location:</b>	J33,VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
<b>Q2904-01</b>	<b>001 willets Pt Blvd(july)</b>	<b>WATER</b>			<b>08/19/25 11:30</b>			<b>08/20/25</b>
			Oil and Grease	1664A			08/21/25 13:00	
<b>Q2904-02</b>	<b>002 35th Ave (JULY)</b>	<b>WATER</b>			<b>08/19/25 11:30</b>			<b>08/20/25</b>
			Oil and Grease	1664A			08/21/25 13:00	



# SAMPLE DATA

## Report of Analysis

Client:	Tully Environmental, Inc	Date Collected:	08/19/25 11:30
Project:	Transfer Station-SPDES	Date Received:	08/20/25
Client Sample ID:	001 willets Pt Blvd(july)	SDG No.:	Q2904
Lab Sample ID:	Q2904-01	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Oil and Grease	9.30		1	0.29	5.00	mg/L		08/21/25 13:00	1664A

Comments:

U = Not Detected  
LOQ = Limit of Quantitation  
MDL = Method Detection Limit  
LOD = Limit of Detection  
D = Dilution  
Q = indicates LCS control criteria did not meet requirements  
H = Sample Analysis Out Of Hold Time

J = Estimated Value  
B = Analyte Found in Associated Method Blank  
\* = indicates the duplicate analysis is not within control limits.  
E = Indicates the reported value is estimated because of the presence of interference.  
OR = Over Range  
N = Spiked sample recovery not within control limits

## Report of Analysis

Client:	Tully Environmental, Inc	Date Collected:	08/19/25 11:30
Project:	Transfer Station-SPDES	Date Received:	08/20/25
Client Sample ID:	002 35th Ave (JULY)	SDG No.:	Q2904
Lab Sample ID:	Q2904-02	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Oil and Grease	8.10		1	0.29	5.00	mg/L		08/21/25 13:00	1664A

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits



# QC RESULT SUMMARY

### Preparation Blank Summary

**Client:** Tully Environmental, Inc

**SDG No.:** Q2904

**Project:** Transfer Station-SPDES

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID:	<b>LB136909BL</b>						
Oil and Grease	mg/L	< 2.5000	2.5000	U	0.29	5.0	08/21/2025

## Matrix Spike Summary

<b>Client:</b>	Tully Environmental, Inc	<b>SDG No.:</b>	Q2904
<b>Project:</b>	Transfer Station-SPDES	<b>Sample ID:</b>	Q2925-01
<b>Client ID:</b>	OUTFALL-2MS	<b>Percent Solids for Spike Sample:</b>	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	21.4		1.30	J	20.0	1	101		08/21/2025



### Matrix Spike Summary

<b>Client:</b>	Tully Environmental, Inc	<b>SDG No.:</b>	Q2904
<b>Project:</b>	Transfer Station-SPDES	<b>Sample ID:</b>	Q2925-01
<b>Client ID:</b>	OUTFALL-2MSD	<b>Percent Solids for Spike Sample:</b>	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/L	78-114	21.1		1.30	J	20.0	1	99		08/21/2025

### Duplicate Sample Summary

<b>Client:</b>	Tully Environmental, Inc	<b>SDG No.:</b>	Q2904
<b>Project:</b>	Transfer Station-SPDES	<b>Sample ID:</b>	Q2925-01
<b>Client ID:</b>	OUTFALL-2MSD	<b>Percent Solids for Spike Sample:</b>	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	21.4		21.1		1	1.41		08/21/2025

### Laboratory Control Sample Summary

**Client:** Tully Environmental, Inc

**SDG No.:** Q2904

**Project:** Transfer Station-SPDES

**Run No.:** LB136909

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136909BS							
Oil and Grease	mg/L	20.0	16.8		84	1	78-114	08/21/2025



# RAW DATA

## Extraction and Analytical Summary Report

**Analysis Method:** 1664A  
**Test:** Oil and Grease  
**Run Number:** LB136909  
**Analysis Date:** 08/21/2025  
**BalanceID:** WC-SC-6  
**OvenID:** EXT OVEN-3

**ANALYST:** jignesh  
**REVIEWED BY:** rubina  
**Extraction Date:** 08/21/2025  
**Extraction IN Time:** 11:50  
**Extraction OUT Time:** 12:25  
**Thermometer ID:** EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB136909BL	LB136909BL	WATER	1.3	1000	100	3.0411	3.0411	0	3.0411	3.0411	0.0000	0
2	LB136909BS	LB136909BS	WATER	1.3	1000	100	2.8753	2.8753	0	2.8921	2.8921	0.0168	16.8
3	Q2904-01	001 willets Pt Blvd(ju	WATER	1.3	1000	100	3.0523	3.0523	0	3.0616	3.0616	0.0093	9.3
4	Q2904-02	002 35th Ave (JULY)	WATER	1.3	1000	100	3.0864	3.0864	0	3.0945	3.0945	0.0081	8.1
5	Q2925-01	OUTFALL-2	WATER	1.6	1000	100	3.0540	3.0540	0	3.0553	3.0553	0.0013	1.3
6	Q2925-02	Q2925-01MS	WATER	1.6	1000	100	3.1623	3.1623	0	3.1837	3.1837	0.0214	21.4
7	Q2925-03	Q2925-01MSD	WATER	1.6	1000	100	2.9354	2.9354	0	2.9565	2.9565	0.0211	21.1

QC Batch# LB136909

**Test:** Oil and Grease

**Analysis Date:** 08/21/2025

### Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W3204
pH Paper 0-14	M6069
Sodium Sulfate	EP2629
1:1 HCL	WP112782
Silica Gel	NA
Sand	NA

### Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	2.5 ML	WP112783
LCSWD	NA	NA
MS/MSD	2.5 ML	WP112784

### BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-6

## Before Analysis

0.0020 gram Balance: 0.0018 (0.0018-0.0022) In OVEN TEMP1 : 70 °C Dessicator Time In1 : 13:41

1.0000 gram Balance: 1.0004 (0.9950-1.0050) In Time1: 13:00

Bal Check Time: 12:10 Out OVEN TEMP1: 70 °C Dessicator Time Out1: 14:20

Out Time1: 13:40

## After Analysis

0.0020 gram Balance: 0.0019 (0.0018-0.0022) In OVEN TEMP2 : 71 °C Dessicator Time In2 : 15:31

1.0000 gram Balance: 1.0005 (0.9950-1.0050) In Time2: 15:00

Bal Check Time: 16:02 Out OVEN TEMP2: 71 °C Dessicator Time Out2: 16:00

Out Time2: 15:30

# WORKLIST(Hardcopy Internal Chain)

B136909

WorkList Name : OIL & GREASE Q2925      WorkList ID : 191398      Department : Wet-Chemistry      Date : 08-21-2025 11:31:13

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2904-01	C 001 willets Pt Blvd(july)	Water	Oil and Grease	Conc H2SO4 to pH < 2	TULL01	J33	08/19/2025	1664A
Q2904-02	D 002 35th Ave (JULY)	Water	Oil and Grease	Conc H2SO4 to pH < 2	TULL01	J33	08/19/2025	1664A
Q2925-01	D OUTFALL-2	Water	Oil and Grease	Conc H2SO4 to pH < 2	ATGG01	D31	08/20/2025	1664A
Q2925-02	Q2925-01MS	Water	Oil and Grease	Conc H2SO4 to pH < 2	ATGG01	D31	08/20/2025	1664A
Q2925-03	Q2925-01MSD	Water	Oil and Grease	Conc H2SO4 to pH < 2	ATGG01	D31	08/20/2025	1664A

Date/Time 08/21/25 11:46  
 Raw Sample Received by: SR (202)  
 Raw Sample Relinquished by: CR (SM)

Date/Time 08/21/25 16:00  
 Raw Sample Received by: CR (SM)  
 Raw Sample Relinquished by: SR (202)

**Instrument ID:** WC SC-3

**Daily Analysis Runlog For Sequence/QC Batch ID # LB136909**

Review By	jignesh	Review On	8/21/2025 4:25:54 PM
Supervise By	rubina	Supervise On	8/22/2025 8:38:34 AM
SubDirectory	LB136909	Test	Oil and Grease
<b>STD. NAME</b>	<b>STD REF.#</b>		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	W3204,M6069,EP2629,WP112782,NA,NA,WP112783,NA,WP112784		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	LB136909BL	LB136909BL	MB	08/21/25 13:00		jignesh	OK
2	LB136909BS	LB136909BS	LCS	08/21/25 13:00		jignesh	OK
3	Q2904-01	001 willets Pt Blvd(july	SAM	08/21/25 13:00		jignesh	OK
4	Q2904-02	002 35th Ave (JULY)	SAM	08/21/25 13:00		jignesh	OK
5	Q2925-01	OUTFALL-2	SAM	08/21/25 13:00		jignesh	OK
6	Q2925-02	Q2925-01MS	MS	08/21/25 13:00		jignesh	OK
7	Q2925-03	Q2925-01MSD	MSD	08/21/25 13:00		jignesh	OK



## Prep Standard - Chemical Standard Summary

**Order ID :** Q2904

**Test :** Oil and Grease

**Prepbatch ID :**

**Sequence ID/Qc Batch ID:** LB136909,

**Standard ID :**

EP2629,WP112782,WP112783,WP112784,

**Chemical ID :**

E3875,E3917,M6069,M6151,W2817,W2871,W3009,W3082,W3112,W3204,



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3923	Baked Sodium Sulfate	<a href="#">EP2629</a>	07/28/2025	01/28/2026	RUPESHKUMAR SHAH	Extraction_SCALE_2 (EX-SC-2)	None	Riteshkumar Patel 07/28/2025
<b><u>FROM</u></b> 4000.00000gram of E3875 = Final Quantity: 4000.000 gram								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
229	1:1 HCL	<a href="#">WP112782</a>	04/22/2025	08/18/2025	Jignesh Parikh	None	None	Iwona Zarych 04/22/2025
<b><u>FROM</u></b> 500.00000ml of M6151 + 500.00000ml of W3112 = Final Quantity: 1.000 L								



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
2470	1664A SPIKING SOLN	<a href="#">WP112783</a>	04/22/2025	10/03/2025	Jignesh Parikh	WETCHEM_SCALE_8 (WCS-7)	None	Iwona Zarych 04/22/2025
<u>FROM</u>	1000.00000ml of E3917 + 4.00000gram of W2817 + 4.00000gram of W2871 = Final Quantity: 1000.000 ml							

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3374	1664A QCS spiking solution-SS	<a href="#">WP112784</a>	04/22/2025	10/03/2025	Jignesh Parikh	WETCHEM_SCALE_8 (WC SC-7)	None	Iwona Zarych 04/22/2025
<u>FROM</u>	1000.00000ml of E3917 + 4.00000gram of W3009 + 4.00000gram of W3082 = Final Quantity: 1000.000 ml							

## CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	417203	01/28/2026	07/28/2025 / RUPESH	01/29/2025 / Rajesh	E3875

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H2762008	10/03/2025	04/03/2025 / Rajesh	03/31/2025 / Rajesh	E3917

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	140440 / TEST PAPERS,PH,0-2.5,.2SENSI, 100PK	80A0441	02/29/2028	09/03/2024 / jignesh	08/19/2024 / Jaswal	M6069

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9530-33 / Hydrochloric Acid, Instra-Analyzed (cs/6x2.5L)	22G2862015	02/17/2026	02/18/2025 / Sagar	01/15/2025 / Sagar	M6151

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	A12244 / Stearic acid, 98%, 100 g	U20E006	04/02/2026	04/02/2021 / apatel	04/02/2021 / apatel	W2817

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	H223-57 / Hexadecane, 99.0%	0000266903	05/04/2027	09/07/2021 / apatel	08/26/2021 / apatel	W2871

## CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	H223-57 / Hexadecane, 99.0%	SHBP8192	02/27/2028	02/27/2023 / lwona	02/27/2023 / lwona	W3009

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	A12244 / Stearic acid, 98%, 100 g	U23E020	02/26/2029	02/26/2024 / lwona	02/26/2024 / lwona	W3082

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / lwona	07/03/2024 / lwona	W3112

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25c0362005	04/30/2026	04/22/2025 / jignesh	04/18/2025 / jignesh	W3204

Hexadecane, 99.0%



Material No.: H223-57  
Batch No.: 0000266903  
Manufactured Date: 2020/05/05  
Retest Date: 2027/05/04  
Revision No: 1

## Certificate of Analysis

Test	Specification	Result
Assay ( $\text{CH}_3(\text{CH}_2)_{14}\text{CH}_3$ ) (by GC)	$\geq 99.0 \%$	99.3
Infrared Spectrum	Passes Test	PT

For Laboratory, Research or Manufacturing Use

Country of Origin: US  
Packaging Site: Paris Mfg Ctr & DC

  
Jamie Ethier  
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700  
Avantor Performance Materials, LLC  
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

**Product Name:** Stearic acid, 98%, Thermo Scientific Chemicals  
**Catalog Number:** A12244.14

---

**CAS Number:** 57-11-4  
**Molecular Formula:** C<sub>18</sub>H<sub>36</sub>O<sub>2</sub>  
**Molecular Weight:** 284.48  
**InChI Key:** QIQXTHQIDYTRH-UHFFFAOYSA-N  
**SMILES:** CCCCCCCCCCCCCCCC(O)=O  
**Synonym:** stearic acid acide stearique hydrofol acid 1855 hydrofol acid 1655 industrène 5016  
stearic acid, ion(1-) (8Cl) glycon TP glycon DP acidum stearinicum hydrofol acid 150

### Product Specification

**Appearance (Color):** White  
**Form:** Crystals or powder or crystalline powder or flakes or waxy solid  
**Assay (Silylated GC):** ≥97.5%  
**Melting Point (clear melt):** 67.0-74.0°C

---

**Date Of Print:** 11/30/2023

*Product Specifications are subject to amendment and may change over time. Data contained is accurate as of the date printed.*

W3009  
rec. 2/27/2023 12

Product Name:

Hexadecane - ReagentPlus®, 99%

## Certificate of Analysis

Product Number:

H6703

Batch Number:

SHBP8192

 $\text{CH}_3(\text{CH}_2)_{14}\text{CH}_3$ 

Brand:

SIAL

CAS Number:

544-76-3

MDL Number:

MFCD00008998

Formula:

C16H34

Formula Weight:

226.44 g/mol

Quality Release Date:

04 AUG 2022

Test	Specification	Result
Appearance (Color)	Colorless or White	Colorless
Appearance (Form)	Liquid or Solid	Liquid
Infrared Spectrum	Conforms to Structure	Conforms
Refractive index at 20 ° C	1.432 - 1.436	1.435
Purity (GC)	≥ 98.5 %	99.3 %
Color Test	≤ 20 APHA	< 5 APHA



Larry Coers, Director

Quality Control

Sheboygan Falls, WI US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at [Sigma-Aldrich.com](http://Sigma-Aldrich.com). For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.







**PRODUCTOS  
QUÍMICOS  
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR  
MONTERREY, N.L. MÉXICO  
CP 64070  
TEL +52 81 13 52 67 67  
www.pqm.com.mx

# CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na <sub>2</sub> SO <sub>4</sub>
SPECIFICATION NUMBER:	6399	RELEASE DATE:	MAY/23/2024
LOT NUMBER :	417203		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na <sub>2</sub> SO <sub>4</sub> )	Min. 99.0%	99.8 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.2
Insoluble matter	Max. 0.01%	0.001 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO <sub>4</sub> )	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.001 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.001 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreign matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.2 %
Retained on US Standard No. 60 sieve	Min. 94%	96.2 %
Through US Standard No. 60 sieve	Max. 5%	3.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

## COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

RE-02-01, Ed. 3

E 3875

Acetone

BAKER RESI-ANALYZED® Reagent  
For Organic Residue Analysis

avantor™



Material No.: 9254-03

Batch No.: 24H2762008

Manufactured Date: 2024-04-18

Expiration Date: 2027-04-18

Revision No.: 0

## Certificate of Analysis

Test	Specification	Result
Assay ((CH <sub>3</sub> ) <sub>2</sub> CO) (by GC, corrected for water)	>= 99.4 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.0 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (μeq/g)	<= 0.3	0.2
Titration Base (μeq/g)	<= 0.6	<0.1
Water (H <sub>2</sub> O)	<= 0.5 %	<0.1 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use

MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

Recd by RP on 03/31/25

E3917

Jamie Croak  
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

R: 8/19/24

M6069

**MACHEREY-NAGEL**

## Certificate of Analysis

### Product information

Product	pH-Fix 0.3-2.3
REF	92180
LOT	80A0441
Expiration date:	29.02.2028
Date of examination:	23.01.2024
Gradation:	pH 0.3-0.7-1.0-1.3-1.6-1.9-2.3

### Confirmation

Hereby we confirm, that the above mentioned product has successfully passed our quality control system in accordance with ISO 9001 and meets the specific quality criteria.

This document has been produced electronically and is valid without a signature.



MACHEREY-NAGEL GmbH & Co. KG  
Valenciennner Str. 11  
52355 Düren · Germany  
[www.mn-net.com](http://www.mn-net.com)

DE	Tel.: +49 24 21 969-0	<a href="mailto:info@mn-net.com">info@mn-net.com</a>
CH	Tel.: +41 62 388 55 00	<a href="mailto:sales-ch@mn-net.com">sales-ch@mn-net.com</a>
FR	Tel.: +33 388 68 22 68	<a href="mailto:sales-fr@mn-net.com">sales-fr@mn-net.com</a>
US	Tel.: +1 888 321 62 24	<a href="mailto:sales-us@mn-net.com">sales-us@mn-net.com</a>

Hydrochloric Acid, 36.5–38.0%  
BAKER INSTRA-ANALYZED® Reagent  
For Trace Metal Analysis

 **avantor™**



M6151

R → 11/15/25

Material No.: 9530-33  
Batch No.: 22G2862015  
Manufactured Date: 2022-06-15  
Retest Date: 2027-06-14  
Revision No.: 0

## Certificate of Analysis

Test	Specification	Result
ACS – Assay (as HCl) (by acid–base titrn)	36.5 – 38.0 %	37.9 %
ACS – Color (APHA)	≤ 10	5
ACS – Residue after Ignition	≤ 3 ppm	< 1 ppm
ACS – Specific Gravity at 60°/60°F	1.185 – 1.192	1.191
ACS – Bromide (Br)	≤ 0.005 %	< 0.005 %
ACS – Extractable Organic Substances	≤ 5 ppm	< 1 ppm
ACS – Free Chlorine (as Cl <sub>2</sub> )	≤ 0.5 ppm	< 0.5 ppm
Phosphate (PO <sub>4</sub> )	≤ 0.05 ppm	< 0.03 ppm
Sulfate (SO <sub>4</sub> )	≤ 0.5 ppm	< 0.3 ppm
Sulfite (SO <sub>3</sub> )	≤ 0.8 ppm	0.3 ppm
Ammonium (NH <sub>4</sub> )	≤ 3 ppm	< 1 ppm
Trace Impurities – Arsenic (As)	≤ 0.010 ppm	< 0.003 ppm
Trace Impurities – Aluminum (Al)	≤ 10.0 ppb	1.3 ppb
Arsenic and Antimony (as As)	≤ 5.0 ppb	< 3.0 ppb
Trace Impurities – Barium (Ba)	≤ 1.0 ppb	0.2 ppb
Trace Impurities – Beryllium (Be)	≤ 1.0 ppb	< 0.2 ppb
Trace Impurities – Bismuth (Bi)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Boron (B)	≤ 20.0 ppb	< 5.0 ppb
Trace Impurities – Cadmium (Cd)	≤ 1.0 ppb	< 0.3 ppb
Trace Impurities – Calcium (Ca)	≤ 50.0 ppb	163.0 ppb
Trace Impurities – Chromium (Cr)	≤ 1.0 ppb	0.7 ppb
Trace Impurities – Cobalt (Co)	≤ 1.0 ppb	< 0.3 ppb
Trace Impurities – Copper (Cu)	≤ 1.0 ppb	< 0.1 ppb
Trace Impurities – Gallium (Ga)	≤ 1.0 ppb	< 0.2 ppb
Trace Impurities – Germanium (Ge)	≤ 3.0 ppb	< 2.0 ppb
Trace Impurities – Gold (Au)	≤ 4.0 ppb	0.6 ppb
Heavy Metals (as Pb)	≤ 100 ppb	< 50 ppb
Trace Impurities – Iron (Fe)	≤ 15 ppb	6 ppb

>>> Continued on page 2 >>>

Hydrochloric Acid, 36.5–38.0%  
BAKER INSTRA-ANALYZED® Reagent  
For Trace Metal Analysis

 **avantor™**



Material No.: 9530-33  
Batch No.: 22G2862015

Test	Specification	Result
Trace Impurities – Lead (Pb)	≤ 1.0 ppb	< 0.5 ppb
Trace Impurities – Lithium (Li)	≤ 1.0 ppb	< 0.2 ppb
Trace Impurities – Magnesium (Mg)	≤ 10.0 ppb	2.9 ppb
Trace Impurities – Manganese (Mn)	≤ 1.0 ppb	< 0.4 ppb
Trace Impurities – Mercury (Hg)	≤ 0.5 ppb	0.1 ppb
Trace Impurities – Molybdenum (Mo)	≤ 10.0 ppb	< 3.0 ppb
Trace Impurities – Nickel (Ni)	≤ 4.0 ppb	< 0.3 ppb
Trace Impurities – Niobium (Nb)	≤ 1.0 ppb	0.8 ppb
Trace Impurities – Potassium (K)	≤ 9.0 ppb	< 2.0 ppb
Trace Impurities – Selenium (Se), For Information Only		< 1.0 ppb
Trace Impurities – Silicon (Si)	≤ 100.0 ppb	< 10.0 ppb
Trace Impurities – Silver (Ag)	≤ 1.0 ppb	0.5 ppb
Trace Impurities – Sodium (Na)	≤ 100.0 ppb	2.3 ppb
Trace Impurities – Strontium (Sr)	≤ 1.0 ppb	< 0.2 ppb
Trace Impurities – Tantalum (Ta)	≤ 1.0 ppb	1.6 ppb
Trace Impurities – Thallium (Tl)	≤ 5.0 ppb	< 2.0 ppb
Trace Impurities – Tin (Sn)	≤ 5.0 ppb	4.0 ppb
Trace Impurities – Titanium (Ti)	≤ 1.0 ppb	1.5 ppb
Trace Impurities – Vanadium (V)	≤ 1.0 ppb	< 0.2 ppb
Trace Impurities – Zinc (Zn)	≤ 5.0 ppb	0.8 ppb
Trace Impurities – Zirconium (Zr)	≤ 1.0 ppb	0.3 ppb

>>> Continued on page 3 >>>

Hydrochloric Acid, 36.5–38.0%  
BAKER INSTRA–ANALYZED® Reagent  
For Trace Metal Analysis



Material No.: 9530-33  
Batch No.: 22G2862015

Test	Specification	Result
------	---------------	--------

For Laboratory, Research, or Manufacturing Use  
Product Information (not specifications):  
Appearance (clear, fuming liquid)  
Meets ACS Specifications  
Storage Condition: Store below 25 °C.

Country of Origin: USA  
Packaging Site: Phillipsburg Mfg Ctr & DC

A handwritten signature in cursive script that reads 'Jamie Ethier'.  
Jamie Ethier  
Vice President Global Quality

# Certificate of analysis

W3082 Received on 2/26/2026 by IZ

Product No.: A12244  
Product: Stearic acid, 98%  
Lot No.: U23E020

Appearance White flakes  
Assay 98.7 %

This document has been electronically generated and does not require a signature.

Order our products online [www.alfa.com](http://www.alfa.com)

**ThermoFisher**  
S C I E N T I F I C

n-Hexane 95%  
ULTRA RESI-ANALYZED  
For Organic Residue Analysis

avantor™



W3204  
084K: 09/22/2025  
38

Material No.: 9262-03  
Batch No.: 25C0362005  
Manufactured Date: 2025-01-29  
Expiration Date: 2026-04-30  
Revision No.: 0

## Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	<= 5	5
Assay (Total Saturated C <sub>6</sub> Isomers) (by GC, corrected for water)	>= 99.5 %	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	>= 95 %	100 %
Color (APHA)	<= 10	10
Residue after Evaporation	<= 1.0 ppm	0.1 ppm
Substances Darkened by H <sub>2</sub> SO <sub>4</sub>	Passes Test	Passes Test
Water (by KF, coulometric)	<= 0.05 %	<0.01 %

For Laboratory, Research, or Manufacturing Use  
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States  
Packaging Site: Phillipsburg Mfg Ctr & DC

*J. Croak*

Jamie Croak  
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700





# SHIPPING DOCUMENTS



284 Sheffield Street, Mountainside, NJ 07092  
(908) 789-8900 Fax: (908) 788-9222  
www.chemtech.net

CHAIN OF CUSTODY RECORD

Alliance Project Number:

Q 2904

COC Number:

CLIENT INFORMATION

COMPANY: Tully Environmental Inc.  
ADDRESS: 57 Seaview Blvd  
CITY: Pt Washington STATE: NY ZIP: 11050  
ATTENTION: Dean Devoe  
PHONE: 718 446 7000 FAX:

PROJECT INFORMATION

PROJECT NAME: Transfer Station SPDES  
PROJECT #: 252113 LOCATION:  
PROJECT MANAGER:  
E-MAIL:  
PHONE: FAX:

BILLING INFORMATION

BILL TO: Same PO#  
ADDRESS:  
CITY: STATE: ZIP:  
ATTENTION: PHONE:

DATA TURNAROUND INFORMATION

FAX: \_\_\_\_\_ DAYS\*  
HARD COPY: \_\_\_\_\_ DAYS\*  
EDD \_\_\_\_\_ DAYS\*  
\* TO BE APPROVED BY ALLIANCE  
STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS

DATA DELIVERABLE INFORMATION

\* RESULTS ONLY ☐ USEPA CLP  
☐ RESULTS + QC ☐ New York State ASP "B"  
☐ New Jersey REDUCED ☐ New York State ASP "A"  
☐ New Jersey CLP ☐ Other \_\_\_\_\_  
☐ EDD Format \_\_\_\_\_

ANALYSIS

BTEX	O&G	Cu, Fe, PB							
1	2	3	4	5	6	7	8	9	

PRESERVATIVES

COMMENTS

<-- Specify Preservatives  
A-HCl B-HNO3  
C-H2SO4 D-NaOH  
E-ICE F-Other

CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# of Bottles										
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	
1.	001 Willets Pt Blvd (July)	W		X	8/19/25	11:30		X	X	X							
2.	002 35th Ave (July)	W		X	8/19/25	11:30		X	X	X							
3.																	
4.																	
5.																	
6.																	
7.																	
8.																	
9.																	
10.																	

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER	DATE/TIME Aug 19, 2025	RECEIVED BY	<p>Conditions of bottles or coolers at receipt: <input type="checkbox"/> Compliant <input type="checkbox"/> Non Compliant <input type="checkbox"/> Cooler Temp <u>43°C</u> MeOH extraction requires an additional 4oz. Jar for percent solid Ice in Cooler?: <u>y</u></p> <p>Comments: 001 Willets Pt. Blvd (July) Metals sample RECEIVED OPEN (empty)</p>
1. D Devoe		1. _____	
RELINQUISHED BY	DATE/TIME 8-20-25 1125	RECEIVED BY	
2. FedEx		2. _____	
RELINQUISHED BY	DATE/TIME	RECEIVED FOR LAB BY	
3. _____		3. _____	

Page \_\_\_\_\_ of \_\_\_\_\_

SHIPPED VIA: CLIENT: ☐ Hand Delivered ☐ Overnight  
ALLIANCE: ☐ Picked Up ☐ Overnight

Shipment Complete  
☐ YES ☐ NO

WHITE - ALLIANCE COPY FOR RETURN TO CLIENT YELLOW - ALLIANCE COPY PINK - SAMPLER COPY



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

### Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

## LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2904 TULL01  
Client Name : Tully Environmental, Inc  
Client Contact : Dean Devoe  
Invoice Name : Tully Environmental, Inc  
Invoice Contact : Dean Devoe

Order Date : 8/20/2025 11:51:00 AM  
Project Name : Transfer Station-SPDES  
Receive DateTime : 8/20/2025 11:25:00 AM  
Purchase Order :

Project Mgr :  
Report Type : Results Only  
EDD Type : EXCEL NOCLEANUP  
Hard Copy Date :  
Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q2904-01	001 willets Pt Blvd(july)	Water	08/19/2025	11:30	VOC-BTEX		624.1	5 Bus. Days	
Q2904-02	002 35th Ave (JULY)	Water	08/19/2025	11:30	VOC-BTEX		624.1	5 Bus. Days	

Relinquished By :

Date / Time :

DN  
8/20/25 1225

Received By :

Date / Time :

Storage Area : VOA Refridgerator Room

AK  
8/20/25 1025