

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q2904

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	997	1000	100	95 - 105	P	08/22/2025	11:37	LB136940
	Iron	3890	4000	97	95 - 105	P	08/22/2025	11:37	LB136940
	Lead	3920	4000	98	95 - 105	P	08/22/2025	11:37	LB136940

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q2904

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	21.2	20.0	106	80 - 120	P	08/22/2025	11:45	LB136940
	Iron	95.1	100	95	80 - 120	P	08/22/2025	11:45	LB136940
	Lead	13.0	12.0	108	80 - 120	P	08/22/2025	11:45	LB136940

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q2904

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1270	1250	101	90 - 110	P	08/22/2025	12:18	LB136940
	Iron	5060	5000	101	90 - 110	P	08/22/2025	12:18	LB136940
	Lead	4980	5000	100	90 - 110	P	08/22/2025	12:18	LB136940
CCV02	Copper	1270	1250	102	90 - 110	P	08/22/2025	13:36	LB136940
	Iron	5090	5000	102	90 - 110	P	08/22/2025	13:36	LB136940
	Lead	4980	5000	100	90 - 110	P	08/22/2025	13:36	LB136940
CCV03	Copper	1290	1250	103	90 - 110	P	08/22/2025	14:25	LB136940
	Iron	5200	5000	104	90 - 110	P	08/22/2025	14:25	LB136940
	Lead	5040	5000	101	90 - 110	P	08/22/2025	14:25	LB136940
CCV04	Copper	1240	1250	100	90 - 110	P	08/22/2025	15:44	LB136940
	Iron	4960	5000	99	90 - 110	P	08/22/2025	15:44	LB136940
	Lead	4920	5000	98	90 - 110	P	08/22/2025	15:44	LB136940
CCV05	Copper	1210	1250	97	90 - 110	P	08/22/2025	16:30	LB136940
	Iron	4870	5000	97	90 - 110	P	08/22/2025	16:30	LB136940
	Lead	4760	5000	95	90 - 110	P	08/22/2025	16:30	LB136940