

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tully Environmental, Inc

SDG No.: Q2904

Contract: TULL01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	8.67	2.0	434	-18	22	08/22/2025	12:00	LB136940
	Iron	98800	101000	98	85600	116500	08/22/2025	12:00	LB136940
	Lead	-7.23			-12	12	08/22/2025	12:00	LB136940
ICSAB01	Copper	491	511	96	434	588	08/22/2025	12:04	LB136940
	Iron	100000	99300	101	84400	114500	08/22/2025	12:04	LB136940
	Lead	39.7	49.0	81	37	61	08/22/2025	12:04	LB136940
ICSA	Copper	3.62	2.0	181	-18	22	08/22/2025	12:09	LB136940
	Iron	95800	101000	95	85600	116500	08/22/2025	12:09	LB136940
	Lead	-0.88			-12	12	08/22/2025	12:09	LB136940
ICSAB	Copper	507	511	99	434	588	08/22/2025	12:13	LB136940
	Iron	97300	99300	98	84400	114500	08/22/2025	12:13	LB136940
	Lead	82.1	49.0	168	37	61	08/22/2025	12:13	LB136940