

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>Q2904</u>
contract: <u>TULL01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q2902-04</u>	client id: <u>OUTFALL-DSN-002MS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2902-04MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	192		42.1		150	100		P
Iron	ug/L	75 - 125	1660		3100		1500	-96	N	P
Lead	ug/L	75 - 125	458		11.8		500	89		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>Q2904</u>
contract: <u>TULL01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q2902-04</u>	client id: <u>OUTFALL-DSN-002MSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2902-04MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	199		42.1		150	105		P
Iron	ug/L	75 - 125	1730		3100		1500	-91	N	P
Lead	ug/L	75 - 125	472		11.8		500	92		P