



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN082025\
Data File : VN087602.D
Acq On : 20 Aug 2025 08:23
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N080525W.M
Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
Last Update : Wed Aug 06 02:01:21 2025



AutoFind: Averaged scan 1842 to 1859; Bkg corrected with scan 1841

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	23.5	5585	PASS
75	95	30	60	56.9	13524	PASS
95	95	100	100	100.0	23770	PASS
96	95	5	9	6.9	1651	PASS
173	174	0.00	2	0.8	133	PASS
174	95	50	100	67.9	16132	PASS
175	174	5	9	8.2	1325	PASS
176	174	95	101	96.5	15575	PASS
177	176	5	9	5.2	814	PASS