

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TULL01
Lab Code: ACE SDG NO.: Q2904
Lab File ID: VN087603.D Date Analyzed: 08/20/2025
Instrument ID: MSVOA_N Time Analyzed: 08:58
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	57702	7.80	317393	9.08	285286	11.85
UPPER LIMIT	115404	8.3	634786	9.582	570572	12.347
LOWER LIMIT	28851	7.3	158697	8.582	142643	11.347
EPA SAMPLE NO.						
OUTFALL-DSN-001MS	50255	7.80	274038	9.09	252169	11.85
OUTFALL-DSN-001MSD	50973	7.81	275779	9.08	252017	11.85
001 willets Pt Blvd(july)	48872	7.80	262809	9.08	236696	11.85
002 35th Ave (JULY)	47710	7.80	260602	9.08	225830	11.85
VN0820WBL01	50156	7.80	275670	9.08	241360	11.85
VN0820WBS01	57311	7.80	304862	9.08	275151	11.85

IS1 = Bromochloromethane
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.