

SOP ID : MSM4500-NH3 B,G-Ammonia-18

SDG No : N/A

Start Digest Date: 08/22/2025 Time : 08:50 Temp : 150 °C

Matrix : SOIL

End Digest Date: 08/22/2025 Time : 09:50 Temp : 158 °C

Pipette ID : WC

Balance ID : WC SC-7

Hood ID : HOOD#2

Digestion tube ID : M5595

Block Thermometer ID : WC CYANIDE

Block ID : WC-DIST-BLOCK-1

Filter paper ID : N/A

Prep Technician Signature: R14

Weigh By : RM

pH Meter ID : N/A

Supervisor Signature: AS

Standard Name	MLS USED	STD REF. # FROM LOG
LCSS	1.0ML	WP114257
MS/MSD SPIKE SOL.	1.0ML	WP114256
PBS003	50.0ML	W3112
RL CHECK	0.1ML	WP114256
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
BORATE BUFFER	2.5ML	WP111325
NAOH 6N	0.5-2.0ML	WP111318
H2SO4 0.04N	5.0ML	WP112828
pH strip-Ammonia	N/A	W3133
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

## Extraction Conformance/Non-Conformance Comments:

ALL GLASSWEAR ARE STEAMED OUT AND THERE WERE NO TRACE OF AMMONIA USING NESLER REAGENT WP114104,

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
08/22/2025 10:00	R14 (WC)	RM (WC)
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	Initial Weight (g)	Final Vol (ml)	pH	Sulfide	Oxidizing	Nitrate/ Nitrite	Comment	Prep Pos
PB169350BL	PBS350	1.00	50	N/A	N/A	N/A	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
PB169350BS	LCS350	1.00	50	N/A	N/A	N/A	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q2905-01	MRS/NRS	1.03	50	N/A	N/A	N/A	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q2905-02	DELUMPER FEED	1.02	50	N/A	N/A	N/A	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q2905-03	MIX	1.04	50	N/A	N/A	N/A	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q2905-03DUP	MIXDUP	1.04	50	N/A	N/A	N/A	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q2905-03MS	MIXMS	1.03	50	N/A	N/A	N/A	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A
Q2905-03MSD	MIXMSD	1.04	50	N/A	N/A	N/A	N/A	AFTER ADDING 6N NAOH PH IS 9.5	N/A