

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATCE02

Lab Code: ACE

SDG NO.: Q2911

Continuing Calib Date: 08/22/2025

Initial Calibration Date(s): 08/19/2025

08/19/2025

Continuing Calib Time: 14:47

Initial Calibration Time(s): 10:02

17:36

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1248-1 (1)	5.81	5.81	5.71	5.91	0.00
Aroclor-1248-2 (2)	6.08	6.08	5.98	6.18	0.00
Aroclor-1248-3 (3)	6.29	6.28	6.18	6.38	0.00
Aroclor-1248-4 (4)	6.68	6.68	6.58	6.78	0.00
Aroclor-1248-5 (5)	6.72	6.72	6.62	6.82	0.00
Tetrachloro-m-xylene	4.66	4.66	4.56	4.76	0.00
Decachlorobiphenyl	10.44	10.44	10.34	10.54	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ATCE02

Lab Code: ACE

SDG NO.: Q2911

Continuing Calib Date: 08/22/2025

Initial Calibration Date(s): 08/19/2025

08/19/2025

Continuing Calib Time: 14:47

Initial Calibration Time(s): 10:02

17:36

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	4.90	4.90	4.80	5.00	0.00
Aroclor-1248-2 (2)	5.12	5.12	5.02	5.22	0.00
Aroclor-1248-3 (3)	5.16	5.16	5.06	5.26	0.00
Aroclor-1248-4 (4)	5.33	5.33	5.23	5.43	0.00
Aroclor-1248-5 (5)	5.72	5.72	5.62	5.82	0.00
Tetrachloro-m-xylene	3.80	3.80	3.70	3.90	0.00
Decachlorobiphenyl	8.81	8.82	8.72	8.92	0.01