

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2911
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113163.D **Time Analyzed:** 08:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.746	4.648	4.848	533.790	500.000	6.8
Aroclor-1016-2	4.764	4.666	4.866	546.600	500.000	9.3
Aroclor-1016-3	4.821	4.724	4.924	543.640	500.000	8.7
Aroclor-1016-4	4.941	4.843	5.043	487.710	500.000	-2.5
Aroclor-1016-5	5.197	5.099	5.299	522.990	500.000	4.6
Aroclor-1260-1	6.233	6.136	6.336	529.460	500.000	5.9
Aroclor-1260-2	6.423	6.326	6.526	527.320	500.000	5.5
Aroclor-1260-3	6.789	6.691	6.891	528.560	500.000	5.7
Aroclor-1260-4	7.048	6.951	7.151	522.740	500.000	4.5
Aroclor-1260-5	7.291	7.194	7.394	535.300	500.000	7.1
Decachlorobiphenyl	8.679	8.582	8.782	51.620	50.000	3.2
Tetrachloro-m-xylene	3.661	3.563	3.763	53.740	50.000	7.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2911
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL01 **Date Analyzed:** 08/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113163.D **Time Analyzed:** 08:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.726	4.629	4.829	509.250	500.000	1.9
Aroclor-1016-2	4.744	4.647	4.847	503.080	500.000	0.6
Aroclor-1016-3	4.919	4.822	5.022	490.360	500.000	-1.9
Aroclor-1016-4	4.962	4.864	5.064	478.370	500.000	-4.3
Aroclor-1016-5	5.174	5.076	5.276	506.050	500.000	1.2
Aroclor-1260-1	6.201	6.104	6.304	485.770	500.000	-2.8
Aroclor-1260-2	6.389	6.292	6.492	483.930	500.000	-3.2
Aroclor-1260-3	6.541	6.443	6.643	454.570	500.000	-9.1
Aroclor-1260-4	7.009	6.912	7.112	474.220	500.000	-5.2
Aroclor-1260-5	7.252	7.155	7.355	470.000	500.000	-6.0
Decachlorobiphenyl	8.624	8.528	8.728	50.990	50.000	2.0
Tetrachloro-m-xylene	3.653	3.555	3.755	50.840	50.000	1.7