

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2911
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113178.D **Time Analyzed:** 14:45

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.746	4.648	4.848	524.080	500.000	4.8
Aroclor-1016-2	4.765	4.666	4.866	530.030	500.000	6.0
Aroclor-1016-3	4.822	4.724	4.924	531.350	500.000	6.3
Aroclor-1016-4	4.941	4.843	5.043	475.730	500.000	-4.9
Aroclor-1016-5	5.197	5.099	5.299	520.370	500.000	4.1
Aroclor-1260-1	6.233	6.136	6.336	484.860	500.000	-3.0
Aroclor-1260-2	6.423	6.326	6.526	489.790	500.000	-2.0
Aroclor-1260-3	6.788	6.691	6.891	479.670	500.000	-4.1
Aroclor-1260-4	7.048	6.951	7.151	494.110	500.000	-1.2
Aroclor-1260-5	7.291	7.194	7.394	495.790	500.000	-0.8
Decachlorobiphenyl	8.679	8.582	8.782	47.160	50.000	-5.7
Tetrachloro-m-xylene	3.661	3.563	3.763	52.510	50.000	5.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2911
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 08/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113178.D **Time Analyzed:** 14:45

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.727	4.629	4.829	498.450	500.000	-0.3
Aroclor-1016-2	4.745	4.647	4.847	496.230	500.000	-0.8
Aroclor-1016-3	4.920	4.822	5.022	480.670	500.000	-3.9
Aroclor-1016-4	4.963	4.864	5.064	467.620	500.000	-6.5
Aroclor-1016-5	5.174	5.076	5.276	501.200	500.000	0.2
Aroclor-1260-1	6.201	6.104	6.304	458.390	500.000	-8.3
Aroclor-1260-2	6.390	6.292	6.492	459.570	500.000	-8.1
Aroclor-1260-3	6.541	6.443	6.643	430.140	500.000	-14.0
Aroclor-1260-4	7.010	6.912	7.112	445.540	500.000	-10.9
Aroclor-1260-5	7.253	7.155	7.355	439.410	500.000	-12.1
Decachlorobiphenyl	8.626	8.528	8.728	49.130	50.000	-1.7
Tetrachloro-m-xylene	3.653	3.555	3.755	49.960	50.000	-0.1