

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2911
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL04 **Date Analyzed:** 08/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074562.D **Time Analyzed:** 08:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.809	5.706	5.906	485.240	500.000	-3.0
Aroclor-1016-2	5.831	5.727	5.927	508.080	500.000	1.6
Aroclor-1016-3	5.894	5.790	5.990	471.790	500.000	-5.6
Aroclor-1016-4	5.991	5.888	6.088	492.820	500.000	-1.4
Aroclor-1016-5	6.283	6.180	6.380	511.900	500.000	2.4
Aroclor-1260-1	7.401	7.297	7.497	492.530	500.000	-1.5
Aroclor-1260-2	7.653	7.550	7.750	489.260	500.000	-2.1
Aroclor-1260-3	8.011	7.908	8.108	517.170	500.000	3.4
Aroclor-1260-4	8.239	8.136	8.336	500.020	500.000	0.0
Aroclor-1260-5	8.565	8.462	8.662	492.030	500.000	-1.6
Decachlorobiphenyl	10.436	10.333	10.533	50.780	50.000	1.6
Tetrachloro-m-xylene	4.657	4.554	4.754	48.970	50.000	-2.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2911
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL04 **Date Analyzed:** 08/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074562.D **Time Analyzed:** 08:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.899	4.799	4.999	496.520	500.000	-0.7
Aroclor-1016-2	4.957	4.855	5.055	506.360	500.000	1.3
Aroclor-1016-3	5.077	4.976	5.176	488.410	500.000	-2.3
Aroclor-1016-4	5.118	5.017	5.217	500.870	500.000	0.2
Aroclor-1016-5	5.331	5.231	5.431	485.760	500.000	-2.8
Aroclor-1260-1	6.362	6.261	6.461	483.650	500.000	-3.3
Aroclor-1260-2	6.548	6.448	6.648	483.570	500.000	-3.3
Aroclor-1260-3	6.702	6.602	6.802	439.540	500.000	-12.1
Aroclor-1260-4	7.172	7.072	7.272	479.380	500.000	-4.1
Aroclor-1260-5	7.411	7.311	7.511	464.100	500.000	-7.2
Decachlorobiphenyl	8.813	8.715	8.915	48.500	50.000	-3.0
Tetrachloro-m-xylene	3.800	3.699	3.899	51.200	50.000	2.4