

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** ATCE02  
**Lab Code:** ACE **SDG NO.:** Q2911  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

**Client Sample No.:** CCAL07 **Date Analyzed:** 08/23/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PP074607.D **Time Analyzed:** 01:22

| COMPOUND             | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|--------|-----------|--------|--------------------|-------------------|------|
|                      |        | FROM      | TO     |                    |                   |      |
| Aroclor-1016-1       | 5.810  | 5.706     | 5.906  | 517.150            | 500.000           | 3.4  |
| Aroclor-1016-2       | 5.832  | 5.727     | 5.927  | 535.080            | 500.000           | 7.0  |
| Aroclor-1016-3       | 5.894  | 5.790     | 5.990  | 490.880            | 500.000           | -1.8 |
| Aroclor-1016-4       | 5.992  | 5.888     | 6.088  | 514.370            | 500.000           | 2.9  |
| Aroclor-1016-5       | 6.284  | 6.180     | 6.380  | 532.650            | 500.000           | 6.5  |
| Aroclor-1260-1       | 7.401  | 7.297     | 7.497  | 502.910            | 500.000           | 0.6  |
| Aroclor-1260-2       | 7.653  | 7.550     | 7.750  | 497.210            | 500.000           | -0.6 |
| Aroclor-1260-3       | 8.011  | 7.908     | 8.108  | 526.350            | 500.000           | 5.3  |
| Aroclor-1260-4       | 8.240  | 8.136     | 8.336  | 506.800            | 500.000           | 1.4  |
| Aroclor-1260-5       | 8.565  | 8.462     | 8.662  | 489.770            | 500.000           | -2.0 |
| Decachlorobiphenyl   | 10.436 | 10.333    | 10.533 | 50.680             | 50.000            | 1.4  |
| Tetrachloro-m-xylene | 4.659  | 4.554     | 4.754  | 52.580             | 50.000            | 5.2  |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** ATCE02  
**Lab Code:** ACE **SDG NO.:** Q2911  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

**Client Sample No.:** CCAL07 **Date Analyzed:** 08/23/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PP074607.D **Time Analyzed:** 01:22

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| Aroclor-1016-1       | 4.900 | 4.799     | 4.999 | 528.170            | 500.000           | 5.6   |
| Aroclor-1016-2       | 4.958 | 4.855     | 5.055 | 550.600            | 500.000           | 10.1  |
| Aroclor-1016-3       | 5.079 | 4.976     | 5.176 | 521.820            | 500.000           | 4.4   |
| Aroclor-1016-4       | 5.119 | 5.017     | 5.217 | 501.960            | 500.000           | 0.4   |
| Aroclor-1016-5       | 5.334 | 5.231     | 5.431 | 472.280            | 500.000           | -5.5  |
| Aroclor-1260-1       | 6.363 | 6.261     | 6.461 | 480.650            | 500.000           | -3.9  |
| Aroclor-1260-2       | 6.549 | 6.448     | 6.648 | 458.750            | 500.000           | -8.3  |
| Aroclor-1260-3       | 6.703 | 6.602     | 6.802 | 420.480            | 500.000           | -15.9 |
| Aroclor-1260-4       | 7.173 | 7.072     | 7.272 | 436.150            | 500.000           | -12.8 |
| Aroclor-1260-5       | 7.412 | 7.311     | 7.511 | 431.770            | 500.000           | -13.6 |
| Decachlorobiphenyl   | 8.814 | 8.715     | 8.915 | 42.100             | 50.000            | -15.8 |
| Tetrachloro-m-xylene | 3.801 | 3.699     | 3.899 | 55.450             | 50.000            | 10.9  |