

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2911
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL08 **Date Analyzed:** 08/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074619.D **Time Analyzed:** 05:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.806	5.706	5.906	477.270	500.000	-4.5
Aroclor-1016-2	5.828	5.727	5.927	504.260	500.000	0.9
Aroclor-1016-3	5.891	5.790	5.990	467.730	500.000	-6.5
Aroclor-1016-4	5.987	5.888	6.088	482.490	500.000	-3.5
Aroclor-1016-5	6.280	6.180	6.380	500.490	500.000	0.1
Aroclor-1260-1	7.397	7.297	7.497	472.640	500.000	-5.5
Aroclor-1260-2	7.650	7.550	7.750	463.510	500.000	-7.3
Aroclor-1260-3	8.008	7.908	8.108	491.980	500.000	-1.6
Aroclor-1260-4	8.236	8.136	8.336	464.640	500.000	-7.1
Aroclor-1260-5	8.561	8.462	8.662	454.250	500.000	-9.2
Decachlorobiphenyl	10.430	10.333	10.533	47.100	50.000	-5.8
Tetrachloro-m-xylene	4.655	4.554	4.754	49.100	50.000	-1.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2911
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL08 **Date Analyzed:** 08/23/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074619.D **Time Analyzed:** 05:26

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.897	4.799	4.999	480.720	500.000	-3.9
Aroclor-1016-2	4.955	4.855	5.055	492.480	500.000	-1.5
Aroclor-1016-3	5.075	4.976	5.176	473.430	500.000	-5.3
Aroclor-1016-4	5.116	5.017	5.217	458.630	500.000	-8.3
Aroclor-1016-5	5.330	5.231	5.431	420.760	500.000	-15.8
Aroclor-1260-1	6.359	6.261	6.461	454.660	500.000	-9.1
Aroclor-1260-2	6.545	6.448	6.648	459.130	500.000	-8.2
Aroclor-1260-3	6.700	6.602	6.802	406.660	500.000	-18.7
Aroclor-1260-4	7.169	7.072	7.272	439.820	500.000	-12.0
Aroclor-1260-5	7.408	7.311	7.511	433.530	500.000	-13.3
Decachlorobiphenyl	8.810	8.715	8.915	40.970	50.000	-18.1
Tetrachloro-m-xylene	3.798	3.699	3.899	53.390	50.000	6.8