

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2911
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL09 **Date Analyzed:** 08/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074625.D **Time Analyzed:** 09:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.811	5.706	5.906	484.620	500.000	-3.1
Aroclor-1016-2	5.832	5.727	5.927	507.530	500.000	1.5
Aroclor-1016-3	5.895	5.790	5.990	466.780	500.000	-6.6
Aroclor-1016-4	5.992	5.888	6.088	488.820	500.000	-2.2
Aroclor-1016-5	6.285	6.180	6.380	510.640	500.000	2.1
Aroclor-1260-1	7.402	7.297	7.497	489.490	500.000	-2.1
Aroclor-1260-2	7.654	7.550	7.750	480.360	500.000	-3.9
Aroclor-1260-3	8.013	7.908	8.108	505.320	500.000	1.1
Aroclor-1260-4	8.241	8.136	8.336	484.400	500.000	-3.1
Aroclor-1260-5	8.567	8.462	8.662	476.020	500.000	-4.8
Decachlorobiphenyl	10.440	10.333	10.533	48.900	50.000	-2.2
Tetrachloro-m-xylene	4.659	4.554	4.754	48.930	50.000	-2.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2911
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL09 **Date Analyzed:** 08/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074625.D **Time Analyzed:** 09:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.899	4.799	4.999	483.550	500.000	-3.3
Aroclor-1016-2	4.958	4.855	5.055	498.030	500.000	-0.4
Aroclor-1016-3	5.077	4.976	5.176	472.590	500.000	-5.5
Aroclor-1016-4	5.118	5.017	5.217	495.260	500.000	-0.9
Aroclor-1016-5	5.332	5.231	5.431	450.530	500.000	-9.9
Aroclor-1260-1	6.362	6.261	6.461	472.760	500.000	-5.4
Aroclor-1260-2	6.548	6.448	6.648	458.710	500.000	-8.3
Aroclor-1260-3	6.703	6.602	6.802	418.350	500.000	-16.3
Aroclor-1260-4	7.172	7.072	7.272	454.340	500.000	-9.1
Aroclor-1260-5	7.411	7.311	7.511	464.570	500.000	-7.1
Decachlorobiphenyl	8.814	8.715	8.915	42.960	50.000	-14.1
Tetrachloro-m-xylene	3.801	3.699	3.899	50.430	50.000	0.9