

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2911
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL10 **Date Analyzed:** 08/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074639.D **Time Analyzed:** 15:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.817	5.706	5.906	460.580	500.000	-7.9
Aroclor-1016-2	5.837	5.727	5.927	442.830	500.000	-11.4
Aroclor-1016-3	5.902	5.790	5.990	448.570	500.000	-10.3
Aroclor-1016-4	5.998	5.888	6.088	466.940	500.000	-6.6
Aroclor-1016-5	6.291	6.180	6.380	486.020	500.000	-2.8
Aroclor-1260-1	7.409	7.297	7.497	462.980	500.000	-7.4
Aroclor-1260-2	7.662	7.550	7.750	466.840	500.000	-6.6
Aroclor-1260-3	8.020	7.908	8.108	498.480	500.000	-0.3
Aroclor-1260-4	8.249	8.136	8.336	481.810	500.000	-3.6
Aroclor-1260-5	8.575	8.462	8.662	472.860	500.000	-5.4
Decachlorobiphenyl	10.452	10.333	10.533	49.840	50.000	-0.3
Tetrachloro-m-xylene	4.664	4.554	4.754	46.200	50.000	-7.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2911
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/19/2025 08/19/2025

Client Sample No.: CCAL10 **Date Analyzed:** 08/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP074639.D **Time Analyzed:** 15:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.903	4.799	4.999	483.930	500.000	-3.2
Aroclor-1016-2	4.961	4.855	5.055	485.740	500.000	-2.9
Aroclor-1016-3	5.081	4.976	5.176	466.180	500.000	-6.8
Aroclor-1016-4	5.122	5.017	5.217	467.870	500.000	-6.4
Aroclor-1016-5	5.336	5.231	5.431	445.940	500.000	-10.8
Aroclor-1260-1	6.367	6.261	6.461	449.700	500.000	-10.1
Aroclor-1260-2	6.553	6.448	6.648	435.410	500.000	-12.9
Aroclor-1260-3	6.708	6.602	6.802	400.290	500.000	-19.9
Aroclor-1260-4	7.177	7.072	7.272	441.760	500.000	-11.6
Aroclor-1260-5	7.416	7.311	7.511	443.490	500.000	-11.3
Decachlorobiphenyl	8.821	8.715	8.915	43.320	50.000	-13.4
Tetrachloro-m-xylene	3.803	3.699	3.899	49.950	50.000	-0.1