

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2914
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL04 **Date Analyzed:** 08/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113157.D **Time Analyzed:** 02:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.747	4.648	4.848	541.930	500.000	8.4
Aroclor-1016-2	4.766	4.666	4.866	552.560	500.000	10.5
Aroclor-1016-3	4.822	4.724	4.924	554.740	500.000	10.9
Aroclor-1016-4	4.942	4.843	5.043	502.850	500.000	0.6
Aroclor-1016-5	5.198	5.099	5.299	531.170	500.000	6.2
Aroclor-1260-1	6.235	6.136	6.336	524.350	500.000	4.9
Aroclor-1260-2	6.424	6.326	6.526	538.030	500.000	7.6
Aroclor-1260-3	6.790	6.691	6.891	513.140	500.000	2.6
Aroclor-1260-4	7.049	6.951	7.151	510.700	500.000	2.1
Aroclor-1260-5	7.293	7.194	7.394	515.160	500.000	3.0
Decachlorobiphenyl	8.681	8.582	8.782	50.610	50.000	1.2
Tetrachloro-m-xylene	3.662	3.563	3.763	54.580	50.000	9.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ATCE02
Lab Code: ACE **SDG NO.:** Q2914
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 08/18/2025 08/18/2025

Client Sample No.: CCAL04 **Date Analyzed:** 08/22/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO113157.D **Time Analyzed:** 02:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.728	4.629	4.829	514.780	500.000	3.0
Aroclor-1016-2	4.746	4.647	4.847	515.900	500.000	3.2
Aroclor-1016-3	4.921	4.822	5.022	508.870	500.000	1.8
Aroclor-1016-4	4.964	4.864	5.064	480.000	500.000	-4.0
Aroclor-1016-5	5.175	5.076	5.276	593.850	500.000	18.8
Aroclor-1260-1	6.203	6.104	6.304	498.890	500.000	-0.2
Aroclor-1260-2	6.391	6.292	6.492	507.520	500.000	1.5
Aroclor-1260-3	6.543	6.443	6.643	461.260	500.000	-7.7
Aroclor-1260-4	7.012	6.912	7.112	490.500	500.000	-1.9
Aroclor-1260-5	7.255	7.155	7.355	494.250	500.000	-1.2
Decachlorobiphenyl	8.627	8.528	8.728	50.280	50.000	0.6
Tetrachloro-m-xylene	3.654	3.555	3.755	52.100	50.000	4.2