

Surrogate Summary

SDG No.: **Q2914**

Client: **ATC Group Services LLC**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO113002.D	PIBLK-PO113002.D	Tetrachloro-m-xyl	1	20	16.6	83		60	140
		Decachlorobiphen	1	20	17.2	86		60	140
		Tetrachloro-m-xyl	2	20	16.9	84		60	140
		Decachlorobiphen	2	20	18.1	91		60	140
I.BLK-PO113125.D	PIBLK-PO113125.D	Tetrachloro-m-xyl	1	20	18.3	92		60	140
		Decachlorobiphen	1	20	18.8	94		60	140
		Tetrachloro-m-xyl	2	20	17.8	89		60	140
		Decachlorobiphen	2	20	18.8	94		60	140
Q2914-01	K262-9A	Tetrachloro-m-xyl	1	20	23.4	117		32	144
		Decachlorobiphen	1	20	41.5	208	*	32	175
		Tetrachloro-m-xyl	2	20	22.2	111		32	144
		Decachlorobiphen	2	20	42.8	214	*	32	175
Q2914-04	K262-10A	Tetrachloro-m-xyl	1	20	19.9	99		32	144
		Decachlorobiphen	1	20	34.9	174		32	175
		Tetrachloro-m-xyl	2	20	19.0	95		32	144
		Decachlorobiphen	2	20	36.4	182	*	32	175
Q2914-07	K262-11A	Tetrachloro-m-xyl	1	20	23.7	118		32	144
		Decachlorobiphen	1	20	38.3	192	*	32	175
		Tetrachloro-m-xyl	2	20	22.5	112		32	144
		Decachlorobiphen	2	20	40.2	201	*	32	175
Q2914-10	K262-12A	Tetrachloro-m-xyl	1	20	23.6	118		32	144
		Decachlorobiphen	1	20	39.0	195	*	32	175
		Tetrachloro-m-xyl	2	20	21.7	108		32	144
		Decachlorobiphen	2	20	40.7	204	*	32	175
I.BLK-PO113140.D	PIBLK-PO113140.D	Tetrachloro-m-xyl	1	20	18.3	91		60	140
		Decachlorobiphen	1	20	18.1	91		60	140
		Tetrachloro-m-xyl	2	20	18.0	90		60	140
		Decachlorobiphen	2	20	18.6	93		60	140
PB169337BL	PB169337BL	Tetrachloro-m-xyl	1	20	20.4	102		32	144
		Decachlorobiphen	1	20	20.3	102		32	175
		Tetrachloro-m-xyl	2	20	19.8	99		32	144
		Decachlorobiphen	2	20	20.7	103		32	175
PB169337BS	PB169337BS	Tetrachloro-m-xyl	1	20	19.2	96		32	144
		Decachlorobiphen	1	20	20.2	101		32	175
		Tetrachloro-m-xyl	2	20	18.8	94		32	144
		Decachlorobiphen	2	20	20.6	103		32	175
I.BLK-PO113152.D	PIBLK-PO113152.D	Tetrachloro-m-xyl	1	20	18.4	92		60	140
		Decachlorobiphen	1	20	18.7	94		60	140
		Tetrachloro-m-xyl	2	20	18.1	90		60	140
		Decachlorobiphen	2	20	18.8	94		60	140
Q2914-01RE	K262-9ARE	Tetrachloro-m-xyl	1	20	24.1	120		32	144

Surrogate Summary

SDG No.: **Q2914**

Client: **ATC Group Services LLC**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
Q2914-01RE	K262-9ARE	Decachlorobiphen	1	20	49.1	246	*	32	175
		Tetrachloro-m-xyl	2	20	22.9	115		32	144
Q2914-07RE	K262-11ARE	Decachlorobiphen	2	20	47.9	240	*	32	175
		Tetrachloro-m-xyl	1	20	23.8	119		32	144
		Decachlorobiphen	1	20	41.5	208	*	32	175
		Tetrachloro-m-xyl	2	20	22.7	114		32	144
Q2914-10RE	K262-12ARE	Decachlorobiphen	2	20	41.8	209	*	32	175
		Tetrachloro-m-xyl	1	20	23.7	118		32	144
		Decachlorobiphen	1	20	41.3	207	*	32	175
		Tetrachloro-m-xyl	2	20	21.9	110		32	144
I.BLK-PO113161.D	PIBLK-PO113161.D	Decachlorobiphen	2	20	42.3	211	*	32	175
		Tetrachloro-m-xyl	1	20	18.4	92		60	140
		Decachlorobiphen	1	20	18.5	92		60	140
		Tetrachloro-m-xyl	2	20	17.7	89		60	140
I.BLK-PO113167.D	PIBLK-PO113167.D	Decachlorobiphen	2	20	18.6	93		60	140
		Tetrachloro-m-xyl	1	20	17.9	90		60	140
		Decachlorobiphen	1	20	18.6	93		60	140
		Tetrachloro-m-xyl	2	20	17.5	88		60	140
Q2914-13	K262-13A	Decachlorobiphen	2	20	18.9	95		60	140
		Tetrachloro-m-xyl	1	20	17.5	88		32	144
		Decachlorobiphen	1	20	22.3	112		32	175
		Tetrachloro-m-xyl	2	20	17.4	87		32	144
Q2914-13MS	K262-13AMS	Decachlorobiphen	2	20	26.3	132		32	175
		Tetrachloro-m-xyl	1	20	14.0	70		32	144
		Decachlorobiphen	1	20	26.5	133		32	175
		Tetrachloro-m-xyl	2	20	14.2	71		32	144
Q2914-13MSD	K262-13AMSD	Decachlorobiphen	2	20	24.9	125		32	175
		Tetrachloro-m-xyl	1	20	16.6	83		32	144
		Decachlorobiphen	1	20	28.6	143		32	175
		Tetrachloro-m-xyl	2	20	16.2	81		32	144
Q2914-16	K262-14A	Decachlorobiphen	2	20	28.5	143		32	175
		Tetrachloro-m-xyl	1	20	20.9	105		32	144
		Decachlorobiphen	1	20	33.2	166		32	175
		Tetrachloro-m-xyl	2	20	20.1	101		32	144
Q2914-19	K262-15A	Decachlorobiphen	2	20	32.7	164		32	175
		Tetrachloro-m-xyl	1	20	25.7	128		32	144
		Decachlorobiphen	1	20	33.7	169		32	175
		Tetrachloro-m-xyl	2	20	25.1	125		32	144
I.BLK-PO113182.D	PIBLK-PO113182.D	Decachlorobiphen	2	20	35.9	180	*	32	175
		Tetrachloro-m-xyl	1	20	17.9	89		60	140
		Decachlorobiphen	1	20	18.1	91		60	140

Surrogate Summary

SDG No.: Q2914

Client: ATC Group Services LLC

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO113182.D	PIBLK-PO113182.D	Tetrachloro-m-xyl	2	20	17.3	86		60	140
		Decachlorobiphen	2	20	18.5	92		60	140