

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2920</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>G Environmental</u>	Datafile :	<u>VN087650.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0825WBS01	Chloromethane	20	19.2	ug/L	96			65	116	
	Vinyl chloride	20	18.6	ug/L	93			65	117	
	Bromomethane	20	19.5	ug/L	98			58	125	
	Chloroethane	20	18.3	ug/L	92			56	128	
	Tert butyl alcohol	100	94.1	ug/L	94			48	142	
	1,1-Dichloroethene	20	17.9	ug/L	90			74	110	
	Acetone	100	92.7	ug/L	93			60	125	
	Carbon disulfide	20	18.3	ug/L	92			64	112	
	Methyl tert-butyl Ether	20	18.0	ug/L	90			78	114	
	Methylene Chloride	20	17.7	ug/L	89			72	114	
	trans-1,2-Dichloroethene	20	17.0	ug/L	85			75	108	
	1,1-Dichloroethane	20	17.7	ug/L	89			78	112	
	2-Butanone	100	90.9	ug/L	91			65	122	
	Carbon Tetrachloride	20	18.0	ug/L	90			77	113	
	cis-1,2-Dichloroethene	20	17.3	ug/L	86			77	110	
	Chloroform	20	17.8	ug/L	89			79	113	
	1,1,1-Trichloroethane	20	18.1	ug/L	91			80	108	
	Benzene	20	17.4	ug/L	87			82	109	
	1,2-Dichloroethane	20	17.3	ug/L	86			80	115	
	Trichloroethene	20	17.1	ug/L	86			77	113	
	1,2-Dichloropropane	20	16.6	ug/L	83			83	111	
	Bromodichloromethane	20	17.2	ug/L	86			83	110	
	4-Methyl-2-Pentanone	100	88.3	ug/L	88			74	118	
	Toluene	20	17.4	ug/L	87			82	110	
	t-1,3-Dichloropropene	20	17.3	ug/L	86			79	110	
	cis-1,3-Dichloropropene	20	17.5	ug/L	88			82	110	
	1,1,2-Trichloroethane	20	17.8	ug/L	89			83	112	
	2-Hexanone	100	92.2	ug/L	92			73	117	
	Dibromochloromethane	20	16.6	ug/L	83			82	110	
	Tetrachloroethene	20	17.0	ug/L	85			67	123	
	Chlorobenzene	20	17.4	ug/L	87			82	109	
	Ethyl Benzene	20	17.8	ug/L	89			83	109	
	m/p-Xylenes	40	36.1	ug/L	90			82	110	
	o-Xylene	20	17.2	ug/L	86			83	109	
	Styrene	20	17.9	ug/L	90			80	111	
	Bromoform	20	18.2	ug/L	91			79	109	
	1,1,2,2-Tetrachloroethane	20	18.2	ug/L	91			76	118	