

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>GENV01</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2920</u>
Lab File ID:	<u>VN087647.D</u>	Date Analyzed:	<u>08/25/2025</u>
Instrument ID:	<u>MSVOA_N</u>	Time Analyzed:	<u>09:37</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	238788	13.771				
UPPER LIMIT	477576	14.271				
LOWER LIMIT	119394	13.271				
EPA SAMPLE NO.						
MW1	178080	13.77				
FB	183679	13.77				
VN0825WBL01	182604	13.77				
VN0825WBS01	202288	13.77				
VN0825WBSD01	204092	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.